




Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN121125\  
Data File : BN038310.D  
Acq On : 11 Dec 2025 10:33  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 11 11:06:42 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Dec 11 11:06:13 2025  
Response via : Initial Calibration



TIC: BN038310.D\data.ms

**(70) Pentachlorophenol (C)**




16.810min ( 0.000) 32618.14 ng

response 947378

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 65.87  |
| 264.00 | 61.60  | 64.44  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN121125\  
Data File : BN038310.D  
Acq On : 11 Dec 2025 10:33  
Operator : RC/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 11 11:06:42 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Dec 11 11:06:13 2025  
Response via : Initial Calibration



TIC: BN038310.D\data.ms

**(77) Benzidine**

19.416min ( 0.000) 0.00 ng

response 4618026

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 14.14  |
| 183.00 | 13.20  | 12.12  |
| 0.00   | 0.00   | 0.00   |