

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN121218\
 Data File : BN003861.D
 Acq On : 12 Dec 2018 21:09
 Operator : JU/SJ
 Sample : J6228-14
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F9E49

Quant Time: Dec 13 04:16:18 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN121218MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 13 04:12:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.83	152	20618	20.00	ng/ul	0.00
18) Naphthalene-d8	10.63	136	102634	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.47	164	66499	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.21	188	166373	20.00	ng/ul	0.00
77) Chrysene-d12	21.39	240	222698	20.00	ng/ul	0.00
85) Perylene-d12	23.71	264	270433	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.29	96	1244	3.95	ng/uL	0.00
5) Phenol-d5	6.99	99	36148	17.83	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.16	67	22659	20.54	ng/ul	0.00
9) 2-Chlorophenol-d4	7.36	132	30463	19.18	ng/ul	0.00
13) 4-Methylphenol-d8	8.53	113	31310	18.43	ng/ul	0.00
19) Nitrobenzene-d5	9.00	128	16730	20.57	ng/ul	0.00
22) 2-Nitrophenol-d4	9.72	143	18803	20.42	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.24	165	32394	19.18	ng/ul	0.00
29) 4-Chloroaniline-d4	10.77	131	51622	32.66	ng/ul	0.00
43) Dimethylphthalate-d6	13.87	166	133316	22.46	ng/ul	0.00
46) Acenaphthylene-d8	14.16	160	156246	22.18	ng/ul	0.00
51) 4-Nitrophenol-d4	14.66	143	22867	18.63	ng/ul	0.00
57) Fluorene-d10	15.46	176	117806	22.95	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.58	200	19679	16.42	ng/ul	0.00
70) Anthracene-d10	17.31	188	197201	23.46	ng/ul	0.00
78) Pyrene-d10	19.60	212	240533	24.14	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.56	264	350653	23.90	ng/ul	0.00

Target Compounds


					Ovalue
6) Phenol	7.01	94	9293	4.542	ng/ul 99
44) Dimethylphthalate	13.92	163	38612	6.663	ng/ul 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121218\
 Data File : BN003861.D
 Acq On : 12 Dec 2018 21:09
 Operator : JU/SJ
 Sample : J6228-14
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampled :
 F9E49

Quant Time: Dec 13 04:16:18 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN121218MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 13 04:12:26 2018
 Response via : Initial Calibration









#6
 Phenol
 Concen: 4.542 ng/ul
 RT: 7.01 min Scan# 684
 Delta R.T. 0.00 min
 Lab File: BN003861.D
 Acq: 12 Dec 2018 21:09

Instrument :
 BNA_N
 ClientSampled :
 F9E49

Tgt Ion	Resp	Lower	Upper
94	100		
65	26.7	21.2	31.8
66	34.3	26.9	40.3



#44
 Dimethylphthalate
 Concen: 6.663 ng/ul
 RT: 13.92 min Scan# 1858
 Delta R.T. 0.00 min
 Lab File: BN003861.D
 Acq: 12 Dec 2018 21:09

Tgt Ion	Resp	Lower	Upper
163	100		
194	4.4	3.4	5.0
164	10.2	7.9	11.9

