

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121222\
 Data File : BN023132.D
 Acq On : 12 Dec 2022 13:05
 Operator : CG/JU
 Sample : N5944-04
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GW-BR-04-100-114-120722

Quant Time: Dec 12 15:45:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 12 05:18:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.021	152	7267	0.400	ng	0.00
7) Naphthalene-d8	10.840	136	21791	0.400	ng	0.00
13) Acenaphthene-d10	14.656	164	11941	0.400	ng	0.00
19) Phenanthrene-d10	17.402	188	26617	0.400	ng	# 0.00
29) Chrysene-d12	21.589	240	20848	0.400	ng	# 0.00
35) Perylene-d12	24.053	264	16659	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.558	112	1583	0.117	ng	0.00
5) Phenol-d6	7.168	99	1268	0.074	ng	0.00
8) Nitrobenzene-d5	9.185	82	3755	0.262	ng	0.00
11) 2-Methylnaphthalene-d10	12.424	152	7708	0.209	ng	0.00
14) 2,4,6-Tribromophenol	16.148	330	1424	0.329	ng	0.00
15) 2-Fluorobiphenyl	13.287	172	12546	0.263	ng	0.00
27) Fluoranthene-d10	19.435	212	20477	0.329	ng	0.00
31) Terphenyl-d14	20.031	244	11969	0.354	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.391	88	5913	0.824	ng	# 92
25) Phenanthrene	17.452	178	4646	0.059	ng	97
26) Anthracene	17.539	178	4367	0.069	ng	98
28) Fluoranthene	19.465	202	5763	0.068	ng	# 95
30) Pyrene	19.827	202	4305	0.056	ng	# 96
34) Bis(2-ethylhexyl)phtha...	21.500	149	6217	0.219	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121222\
Data File : BN023132.D
Acq On : 12 Dec 2022 13:05
Operator : CG/JU
Sample : N5944-04
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GW-BR-04-100-114-120722

Quant Time: Dec 12 15:45:22 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 12 05:18:26 2022
Response via : Initial Calibration

