

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121222\
 Data File : BN023139.D
 Acq On : 12 Dec 2022 17:24
 Operator : CG/JU
 Sample : N5972-02
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW9-PT-IN-02-12082022

Quant Time: Dec 13 05:45:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 12 05:18:26 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Christian Giraldo 12/13/2022
 Supervised By :mohammad ahmed 12/14/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.020	152	6637	0.400	ng	0.00
7) Naphthalene-d8	10.840	136	19203	0.400	ng	0.00
13) Acenaphthene-d10	14.655	164	9650	0.400	ng	0.00
19) Phenanthrene-d10	17.402	188	21047	0.400	ng	# 0.00
29) Chrysene-d12	21.589	240	15904	0.400	ng	0.00
35) Perylene-d12	24.047	264	12794	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.558	112	1687	0.136	ng	0.00
5) Phenol-d6	7.168	99	1243	0.079	ng	0.00
8) Nitrobenzene-d5	9.185	82	3522m	0.278	ng	0.00
11) 2-Methylnaphthalene-d10	12.423	152	6781	0.208	ng	0.00
14) 2,4,6-Tribromophenol	16.148	330	1122	0.320	ng	0.00
15) 2-Fluorobiphenyl	13.287	172	11180	0.290	ng	0.00
27) Fluoranthene-d10	19.435	212	15109	0.307	ng	0.00
31) Terphenyl-d14	20.022	244	9576	0.371	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.491	149	641	0.030	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121222\
Data File : BN023139.D
Acq On : 12 Dec 2022 17:24
Operator : CG/JU
Sample : N5972-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW9-PT-IN-02-12082022

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 12/13/2022
Supervised By :mohammad ahmed 12/14/2022

Quant Time: Dec 13 05:45:42 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 12 05:18:26 2022
Response via : Initial Calibration

