

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121222\
 Data File : BN023154.D
 Acq On : 13 Dec 2022 04:23
 Operator : CG/JU
 Sample : SP56087
 Misc : 8270 SIM SURROGATE
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP56087

Quant Time: Dec 13 05:48:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 12 05:18:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.013	152	8954	0.400	ng	0.00
7) Naphthalene-d8	10.829	136	26404	0.400	ng	-0.01
13) Acenaphthene-d10	14.656	164	13730	0.400	ng	0.00
19) Phenanthrene-d10	17.402	188	30902	0.400	ng	# 0.00
29) Chrysene-d12	21.589	240	21875	0.400	ng	0.00
35) Perylene-d12	24.044	264	19075	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.558	112	8032	0.482	ng	0.00
5) Phenol-d6	7.168	99	9967	0.470	ng	0.00
8) Nitrobenzene-d5	9.174	82	6349	0.365	ng	-0.01
11) 2-Methylnaphthalene-d10	12.420	152	12814	0.286	ng	0.00
14) 2,4,6-Tribromophenol	16.148	330	1970	0.395	ng	0.00
15) 2-Fluorobiphenyl	13.287	172	20265	0.369	ng	0.00
27) Fluoranthene-d10	19.431	212	24017	0.332	ng	0.00
31) Terphenyl-d14	20.022	244	14315	0.403	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121222\
Data File : BN023154.D
Acq On : 13 Dec 2022 04:23
Operator : CG/JU
Sample : SP56087
Misc : 8270 SIM SURROGATE
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP56087

Quant Time: Dec 13 05:48:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 12 05:18:26 2022
Response via : Initial Calibration

