

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121225\
 Data File : BN038358.D
 Acq On : 12 Dec 2025 19:03
 Operator : RC/JU
 Sample : PB170912BSD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB170912BSD

Quant Time: Dec 13 02:09:38 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.717	152	5504	0.400	ng	0.00
7) Naphthalene-d8	10.519	136	14243	0.400	ng	#-0.01
13) Acenaphthene-d10	14.388	164	6818	0.400	ng	-0.01
19) Phenanthrene-d10	17.149	188	11665	0.400	ng	0.00
29) Chrysene-d12	21.340	240	8305	0.400	ng	# 0.00
35) Perylene-d12	23.622	264	8801	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.290	112	4917	0.359	ng	0.00
5) Phenol-d6	6.887	99	5454	0.335	ng	0.00
8) Nitrobenzene-d5	8.875	82	4653	0.387	ng	0.00
11) 2-Methylnaphthalene-d10	12.121	152	9506	0.473	ng	0.00
14) 2,4,6-Tribromophenol	15.895	330	836	0.276	ng	0.00
15) 2-Fluorobiphenyl	13.008	172	13910	0.423	ng	0.00
27) Fluoranthene-d10	19.183	212	9671	0.335	ng	0.00
31) Terphenyl-d14	19.782	244	7442	0.402	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.196	88	1838	0.291	ng	# 65
3) n-Nitrosodimethylamine	3.507	42	2757	0.348	ng	# 97
6) bis(2-Chloroethyl)ether	7.139	93	5483	0.344	ng	96
9) Naphthalene	10.573	128	14349	0.341	ng	100
10) Hexachlorobutadiene	10.872	225	2638	0.352	ng	# 99
12) 2-Methylnaphthalene	12.197	142	8109	0.303	ng	99
16) Acenaphthylene	14.110	152	12558	0.375	ng	99
17) Acenaphthene	14.452	154	7570	0.325	ng	98
18) Fluorene	15.446	166	9826	0.327	ng	100
20) 4,6-Dinitro-2-methylph...	15.522	198	208	0.108	ng	# 1
21) 4-Bromophenyl-phenylether	16.342	248	2851	0.329	ng	97
22) Hexachlorobenzene	16.453	284	3612	0.347	ng	100
23) Atrazine	16.615	200	1948	0.330	ng	99
24) Pentachlorophenol	16.801	266	1446	0.356	ng	97
25) Phenanthrene	17.186	178	13452	0.331	ng	100
26) Anthracene	17.273	178	12005	0.343	ng	100
28) Fluoranthene	19.211	202	13300	0.308	ng	100
30) Pyrene	19.573	202	13499	0.330	ng	99
32) Benzo(a)anthracene	21.322	228	11195	0.349	ng	99
33) Chrysene	21.375	228	12514	0.354	ng	99
34) Bis(2-ethylhexyl)phtha...	21.259	149	5141	0.292	ng	# 99
36) Indeno(1,2,3-cd)pyrene	25.949	276	17687	0.379	ng	100
37) Benzo(b)fluoranthene	22.935	252	13459	0.333	ng	95
38) Benzo(k)fluoranthene	22.978	252	14124	0.350	ng	97
39) Benzo(a)pyrene	23.522	252	12179	0.368	ng	98
40) Dibenzo(a,h)anthracene	25.961	278	13531	0.376	ng	97
41) Benzo(g,h,i)perylene	26.653	276	15173	0.377	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121225\
Data File : BN038358.D
Acq On : 12 Dec 2025 19:03
Operator : RC/JU
Sample : PB170912BSD
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB170912BSD


Quant Time: Dec 13 02:09:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 10 13:56:50 2025
Response via : Initial Calibration







#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.717 min Scan# 64
 Delta R.T. -0.007 min
 Lab File: BN038358.D
 Acq: 12 Dec 2025 19:03


Instrument :
 BNA_N
 ClientSampleId :
 PB170912BSD





Tgt Ion:152 Resp: 5504
 Ion Ratio Lower Upper
 152 100
 150 156.1 122.4 183.6
 115 61.0 47.3 70.9




#2
 1,4-Dioxane
 Concen: 0.291 ng
 RT: 3.196 min Scan# 16
 Delta R.T. 0.000 min
 Lab File: BN038358.D
 Acq: 12 Dec 2025 19:03



Tgt Ion: 88 Resp: 1838
 Ion Ratio Lower Upper
 88 100
 43 78.1 24.9 37.3#
 58 89.4 61.4 92.0









#3
 n-Nitrosodimethylamine
 Concen: 0.348 ng
 RT: 3.507 min Scan# 59
 Delta R.T. 0.007 min
 Lab File: BN038358.D
 Acq: 12 Dec 2025 19:03




Instrument :
 BNA_N
 ClientSampleId :
 PB170912BSD


Tgt Ion: 42 Resp: 2757
 Ion Ratio Lower Upper
 42 100
 74 118.3 95.3 142.9
 44 0.0 7.9 11.9#



#4
 2-Fluorophenol
 Concen: 0.359 ng
 RT: 5.290 min Scan# 306
 Delta R.T. 0.000 min
 Lab File: BN038358.D
 Acq: 12 Dec 2025 19:03

Tgt Ion: 112 Resp: 4917
 Ion Ratio Lower Upper
 112 100
 64 56.8 44.4 66.6
 63 29.4 23.4 35.0









#5
Phenol-d6
Concen: 0.335 ng
RT: 6.887 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03




Instrument :
BNA_N
ClientSampleId :
PB170912BSD

Tgt Ion:	99	Resp:	5454
Ion	Ratio	Lower	Upper
99	100		
42	22.5	16.5	24.7
71	32.2	24.9	37.3



#6
bis(2-Chloroethyl)ether
Concen: 0.344 ng
RT: 7.139 min Scan# 562
Delta R.T. -0.007 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03

Tgt Ion:	93	Resp:	5483
Ion	Ratio	Lower	Upper
93	100		
63	77.9	59.2	88.8
95	32.6	25.2	37.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.519 min Scan# 931
Delta R.T. -0.011 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03




Instrument :
BNA_N
ClientSampleId :
PB170912BSD


Tgt Ion:	136	Resp:	14243
Ion Ratio	Lower	Upper	
136	100		
137	11.2	9.1	13.7
54	9.1	5.2	7.8#
68	6.9	4.1	6.1#



#8
Nitrobenzene-d5
Concen: 0.387 ng
RT: 8.875 min Scan# 777
Delta R.T. 0.000 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03

Tgt Ion:	82	Resp:	4653
Ion Ratio	Lower	Upper	
82	100		
128	47.7	34.0	51.0
54	53.2	44.4	66.6









#9
Naphthalene
Concen: 0.341 ng
RT: 10.573 min Scan# 936
Delta R.T. 0.000 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03




Instrument :
BNA_N
ClientSampleId :
PB170912BSD

Tgt Ion:	128	Resp:	14349
Ion	Ratio	Lower	Upper
128	100		
129	11.4	9.1	13.7
127	13.6	10.9	16.3



#10
Hexachlorobutadiene
Concen: 0.352 ng
RT: 10.872 min Scan# 964
Delta R.T. 0.000 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03

Tgt Ion:	225	Resp:	2638
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.5	51.2	76.8









#11
2-Methylnaphthalene-d10
Concen: 0.473 ng
RT: 12.121 min Scan# 1144
Delta R.T. -0.005 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03




Instrument :
BNA_N
ClientSampleId :
PB170912BSD


Tgt Ion:152 Resp: 9506
Ion Ratio Lower Upper
152 100
151 15.2 17.0 25.6#



#12
2-Methylnaphthalene
Concen: 0.303 ng
RT: 12.197 min Scan# 1159
Delta R.T. -0.005 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03

Tgt Ion:142 Resp: 8109
Ion Ratio Lower Upper
142 100
141 88.2 71.4 107.2
115 35.1 28.4 42.6





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.388 min Scan# 1448

Delta R.T. -0.011 min

Lab File: BN038358.D

Acq: 12 Dec 2025 19:03

Instrument :

BNA_N

ClientSampleId :

PB170912BSD





Tgt Ion:164 Resp: 6818

Ion Ratio Lower Upper

164 100

162 106.2 86.2 129.4

160 53.7 44.6 67.0



#14

2,4,6-Tribromophenol

Concen: 0.276 ng

RT: 15.895 min Scan# 1584

Delta R.T. 0.000 min

Lab File: BN038358.D

Acq: 12 Dec 2025 19:03




Tgt Ion:330 Resp: 836


Ion Ratio Lower Upper

330 100

332 97.4 75.8 113.6

141 42.3 31.8 47.8









#15
2-Fluorobiphenyl
Concen: 0.423 ng
RT: 13.008 min Scan# 1319
Delta R.T. 0.000 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03




Instrument :
BNA_N
ClientSampleId :
PB170912BSD


Tgt Ion:	172	Resp:	13910
Ion Ratio	Lower	Upper	
172	100		
171	35.0	28.6	42.8
170	23.4	19.3	28.9



#16
Acenaphthylene
Concen: 0.375 ng
RT: 14.110 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03

Tgt Ion:	152	Resp:	12558
Ion Ratio	Lower	Upper	
152	100		
151	20.2	15.4	23.2
153	13.0	10.2	15.4








#17
Acenaphthene
Concen: 0.325 ng
RT: 14.452 min Scan# 1454
Delta R.T. -0.011 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03




Instrument :
BNA_N
ClientSampleId :
PB170912BSD


Tgt Ion:154 Resp: 7570
Ion Ratio Lower Upper
154 100
153 114.9 89.8 134.8
152 60.5 47.5 71.3



#18
Fluorene
Concen: 0.327 ng
RT: 15.446 min Scan# 1547
Delta R.T. -0.000 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03

Tgt Ion:166 Resp: 9826
Ion Ratio Lower Upper
166 100
165 100.1 80.2 120.4
167 13.4 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.149 min Scan# 11665

Delta R.T. 0.000 min

Lab File: BN038358.D

Acq: 12 Dec 2025 19:03

Instrument :

BNA_N

ClientSampleId :

PB170912BSD





Tgt Ion:188 Resp: 11665

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.1 7.4 11.0



#20

4,6-Dinitro-2-methylphenol

Concen: 0.108 ng

RT: 15.522 min Scan# 1554

Delta R.T. 0.000 min

Lab File: BN038358.D

Acq: 12 Dec 2025 19:03




Tgt Ion:198 Resp: 208


Ion Ratio Lower Upper

198 100

51 247.7 86.3 129.5#

105 115.4 45.3 67.9#





#21

4-Bromophenyl-phenylether

Concen: 0.329 ng

RT: 16.342 min Scan# 1620

Delta R.T. -0.000 min

Lab File: BN038358.D

Acq: 12 Dec 2025 19:03





Instrument :

BNA_N

ClientSampleId :

PB170912BSD

Tgt Ion:	248	Resp:	2851
Ion Ratio	Lower	Upper	
248	100		
250	98.6	78.2	117.2
141	49.7	43.7	65.5



#22

Hexachlorobenzene

Concen: 0.347 ng




RT: 16.453 min Scan# 1629


Delta R.T. -0.012 min

Lab File: BN038358.D

Acq: 12 Dec 2025 19:03

Tgt Ion:	284	Resp:	3612
Ion Ratio	Lower	Upper	
284	100		
142	37.4	29.5	44.3
249	29.6	23.7	35.5









#23
 Atrazine
 Concen: 0.330 ng
 RT: 16.615 min Scan# 1642
 Delta R.T. 0.000 min
 Lab File: BN038358.D
 Acq: 12 Dec 2025 19:03




Instrument :
 BNA_N
 ClientSampleId :
 PB170912BSD


Tgt Ion	Ratio	Lower	Upper
200	100		
173	28.6	22.1	33.1
215	46.2	37.4	56.2



#24
 Pentachlorophenol
 Concen: 0.356 ng
 RT: 16.801 min Scan# 1657
 Delta R.T. 0.000 min
 Lab File: BN038358.D
 Acq: 12 Dec 2025 19:03

Tgt Ion	Ratio	Lower	Upper
266	100		
264	60.9	50.2	75.2
268	62.3	52.6	78.8









#25
Phenanthrene
Concen: 0.331 ng
RT: 17.186 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03




Instrument :
BNA_N
ClientSampleId :
PB170912BSD


Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.4	15.5	23.3
179	15.5	12.2	18.2



#26
Anthracene
Concen: 0.343 ng
RT: 17.273 min Scan# 1695
Delta R.T. -0.012 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.1	22.7
179	15.1	12.0	18.0









#27
Fluoranthene-d10
Concen: 0.335 ng
RT: 19.183 min Scan# 1939
Delta R.T. -0.005 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03




Instrument :
BNA_N
ClientSampleId :
PB170912BSD


Tgt Ion:212 Resp: 9671
Ion Ratio Lower Upper
212 100
106 17.5 13.7 20.5
104 9.8 7.8 11.8



#28
Fluoranthene
Concen: 0.308 ng
RT: 19.211 min Scan# 1945
Delta R.T. -0.009 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03

Tgt Ion:202 Resp: 13300
Ion Ratio Lower Upper
202 100
101 12.1 9.6 14.4
203 17.2 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.340 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN038358.D

Acq: 12 Dec 2025 19:03

Instrument :

BNA_N

ClientSampleId :

PB170912BSD





Tgt Ion:240 Resp: 8305

Ion Ratio Lower Upper

240 100

120 15.0 15.4 23.2#

236 29.1 23.9 35.9



#30

Pyrene

Concen: 0.330 ng

RT: 19.573 min Scan# 2023

Delta R.T. -0.009 min

Lab File: BN038358.D

Acq: 12 Dec 2025 19:03




Tgt Ion:202 Resp: 13499

Ion Ratio Lower Upper

202 100

200 21.2 16.9 25.3

203 17.2 14.2 21.2









#31
 Terphenyl-d14
 Concen: 0.402 ng
 RT: 19.782 min Scan# 2068
 Delta R.T. -0.009 min
 Lab File: BN038358.D
 Acq: 12 Dec 2025 19:03




Instrument :
 BNA_N
 ClientSampleId :
 PB170912BSD


Tgt Ion:244 Resp: 7442
 Ion Ratio Lower Upper
 244 100
 212 10.2 7.9 11.9
 122 20.1 15.0 22.6



#32
 Benzo(a)anthracene
 Concen: 0.349 ng
 RT: 21.322 min Scan# 2263
 Delta R.T. -0.009 min
 Lab File: BN038358.D
 Acq: 12 Dec 2025 19:03

Tgt Ion:228 Resp: 11195
 Ion Ratio Lower Upper
 228 100
 226 28.2 23.3 34.9
 229 20.1 15.7 23.5





#33

Chrysene

Concen: 0.354 ng

RT: 21.375 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN038358.D

Acq: 12 Dec 2025 19:03

Instrument :

BNA_N

ClientSampleId :

PB170912BSD





Tgt Ion:228 Resp: 12514

Ion Ratio Lower Upper

228 100

226 30.6 25.0 37.4

229 20.1 15.5 23.3



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.292 ng

RT: 21.259 min Scan# 2256

Delta R.T. -0.009 min

Lab File: BN038358.D

Acq: 12 Dec 2025 19:03




Tgt Ion:149 Resp: 5141


Ion Ratio Lower Upper

149 100

167 27.1 21.4 32.0

279 3.6 2.4 3.6#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.622 min Scan# 2800

Delta R.T. -0.012 min

Lab File: BN038358.D


Acq: 12 Dec 2025 19:03

Instrument :

BNA_N

ClientSampleId :

PB170912BSD






Tgt Ion:264 Resp: 8801

Ion Ratio Lower Upper

264 100

260 27.8 22.2 33.2

265 122.1 80.2 120.2#



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.379 ng

RT: 25.949 min Scan# 3596

Delta R.T. -0.015 min

Lab File: BN038358.D

Acq: 12 Dec 2025 19:03




Tgt Ion:276 Resp: 17687


Ion Ratio Lower Upper

276 100

138 29.5 23.6 35.4

277 24.6 19.7 29.5









#37
Benzo(b)fluoranthene
Concen: 0.333 ng
RT: 22.935 min Scan# 2565
Delta R.T. -0.012 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03




Instrument :
BNA_N
ClientSampleId :
PB170912BSD

Tgt Ion:252 Resp: 13459
Ion Ratio Lower Upper
252 100
253 28.5 21.1 31.7
125 23.7 16.4 24.6



#38
Benzo(k)fluoranthene
Concen: 0.350 ng
RT: 22.978 min Scan# 2580
Delta R.T. -0.012 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03


Tgt Ion:252 Resp: 14124
Ion Ratio Lower Upper
252 100
253 27.9 21.4 32.0
125 22.3 16.2 24.2








#39
Benzo(a)pyrene
Concen: 0.368 ng
RT: 23.522 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03

Instrument :
BNA_N
ClientSampleId :
PB170912BSD






Tgt Ion:252 Resp: 12179
Ion Ratio Lower Upper
252 100
253 31.0 23.3 34.9
125 28.5 22.4 33.6



#40
Dibenzo(a,h)anthracene
Concen: 0.376 ng
RT: 25.961 min Scan# 3600
Delta R.T. -0.020 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03


Tgt Ion:278 Resp: 13531
Ion Ratio Lower Upper
278 100
139 22.3 17.8 26.8
279 29.0 21.3 31.9





#41
Benzo(g,h,i)perylene
Concen: 0.377 ng
RT: 26.653 min Scan# 3842
Delta R.T. -0.023 min
Lab File: BN038358.D
Acq: 12 Dec 2025 19:03

Instrument :
BNA_N
ClientSampleId :
PB170912BSD



Tgt Ion:	276	Resp:	15173
Ion	Ratio	Lower	Upper
276	100		
277	24.9	20.0	30.0
138	28.1	21.8	32.6

