

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121225\
 Data File : BN038357.D
 Acq On : 12 Dec 2025 18:27
 Operator : RC/JU
 Sample : PB170912BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB170912BS

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 12/15/2025
 Supervised By :Jagrut Upadhyay 12/15/2025

Quant Time: Dec 13 02:09:13 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.717	152	6159	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	16009	0.400	ng	#-0.01
13) Acenaphthene-d10	14.388	164	7792	0.400	ng	-0.01
19) Phenanthrene-d10	17.149	188	13155	0.400	ng	0.00
29) Chrysene-d12	21.340	240	8480	0.400	ng	0.00
35) Perylene-d12	23.625	264	8947	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.291	112	5525	0.360	ng	0.00
5) Phenol-d6	6.887	99	6254	0.343	ng	0.00
8) Nitrobenzene-d5	8.875	82	5244	0.388	ng	0.00
11) 2-Methylnaphthalene-d10	12.121	152	8204m	0.363	ng	0.00
14) 2,4,6-Tribromophenol	15.883	330	944	0.273	ng	-0.01
15) 2-Fluorobiphenyl	13.009	172	16230	0.432	ng	0.00
27) Fluoranthene-d10	19.183	212	10641	0.327	ng	0.00
31) Terphenyl-d14	19.782	244	8141	0.430	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.196	88	2073	0.294	ng	# 57
3) n-Nitrosodimethylamine	3.507	42	2894	0.327	ng	# 95
6) bis(2-Chloroethyl)ether	7.139	93	5934	0.333	ng	98
9) Naphthalene	10.573	128	16000	0.338	ng	99
10) Hexachlorobutadiene	10.872	225	2793	0.332	ng	# 98
12) 2-Methylnaphthalene	12.197	142	9171	0.305	ng	99
16) Acenaphthylene	14.110	152	14230	0.372	ng	100
17) Acenaphthene	14.452	154	8645	0.324	ng	99
18) Fluorene	15.446	166	11335	0.330	ng	98
20) 4,6-Dinitro-2-methylph...	15.523	198	217	0.100	ng	# 13
21) 4-Bromophenyl-phenylether	16.342	248	3145	0.322	ng	93
22) Hexachlorobenzene	16.453	284	4007	0.342	ng	99
23) Atrazine	16.615	200	2150	0.323	ng	99
24) Pentachlorophenol	16.801	266	1260	0.275	ng	98
25) Phenanthrene	17.186	178	15862	0.346	ng	100
26) Anthracene	17.273	178	13583	0.344	ng	100
28) Fluoranthene	19.211	202	15139	0.311	ng	99
30) Pyrene	19.573	202	15342	0.367	ng	100
32) Benzo(a)anthracene	21.322	228	11503	0.351	ng	98
33) Chrysene	21.375	228	12406	0.344	ng	99
34) Bis(2-ethylhexyl)phtha...	21.259	149	5420	0.302	ng	100
36) Indeno(1,2,3-cd)pyrene	25.946	276	17589	0.370	ng	100
37) Benzo(b)fluoranthene	22.932	252	13100	0.319	ng	96
38) Benzo(k)fluoranthene	22.978	252	13592	0.332	ng	94
39) Benzo(a)pyrene	23.522	252	11710	0.348	ng	96
40) Dibenzo(a,h)anthracene	25.964	278	13361	0.365	ng	98
41) Benzo(g,h,i)perylene	26.651	276	15315	0.374	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121225\
Data File : BN038357.D
Acq On : 12 Dec 2025 18:27
Operator : RC/JU
Sample : PB170912BS
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB170912BS

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/15/2025
Supervised By :Jagrut Upadhyay 12/15/2025

Quant Time: Dec 13 02:09:13 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 10 13:56:50 2025
Response via : Initial Calibration

