

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121225\
 Data File : BN038389.D
 Acq On : 13 Dec 2025 16:28
 Operator : RC/JU
 Sample : Q3839-21
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0B09

Quant Time: Dec 14 22:24:50 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.717	152	5937	0.400	ng	0.00
7) Naphthalene-d8	10.520	136	17153	0.400	ng	-0.01
13) Acenaphthene-d10	14.388	164	9499	0.400	ng	-0.01
19) Phenanthrene-d10	17.136	188	17358	0.400	ng	#-0.01
29) Chrysene-d12	21.331	240	15003	0.400	ng	# 0.00
35) Perylene-d12	23.622	264	16496	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.283	112	864059	58.438	ng	0.00
5) Phenol-d6	6.879	99	664957	37.830	ng	0.00
8) Nitrobenzene-d5	8.865	82	1114899	77.058	ng	-0.01
11) 2-Methylnaphthalene-d10	12.091	152	24	0.001	ng	-0.04
14) 2,4,6-Tribromophenol	15.883	330	816232	193.684	ng	-0.01
15) 2-Fluorobiphenyl	13.004	172	3196252	69.845	ng	0.00
27) Fluoranthene-d10	19.174	212	78	0.002	ng	-0.01
31) Terphenyl-d14	19.782	244	2485297	74.228	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.189	88	795	0.117	ng	# 36
34) Bis(2-ethylhexyl)phtha...	21.259	149	9260	0.291	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121225\
Data File : BN038389.D
Acq On : 13 Dec 2025 16:28
Operator : RC/JU
Sample : Q3839-21
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0B09

Quant Time: Dec 14 22:24:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 10 13:56:50 2025
Response via : Initial Calibration

