


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN121318\
Data File : BN003886.D
Acq On : 13 Dec 2018 11:44
Operator : JU/SJ
Sample : DFTPP67
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP67

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN121218MA.M
Title : SVOA CALIBRATION
Last Update : Fri Dec 14 04:33:02 2018



AutoFind: Scans 2487, 2488, 2489; Background Corrected with Scan 2482

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.0	12285	PASS
68	69	0.00	2	0.7	104	PASS
69	198	0.00	100	34.1	13979	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	44.7	18345	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	41002	PASS
199	198	5	9	6.8	2796	PASS
275	198	10	60	30.5	12522	PASS
365	198	1	100	3.9	1591	PASS
441	198	0.01	100	13.1	5381	PASS
442	198	50	100	84.7	34749	PASS
443	442	15	24	19.8	6895	PASS

m/z	Abundance
43.90	583.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

m/z	Abundance
39.00	1006.0
44.00	493.0
50.00	3353.0
51.00	13590.0
52.00	697.0
55.90	454.0
57.00	1002.0
63.00	615.0
69.00	15208.0
74.00	1466.0
75.00	2428.0
76.10	812.0
77.00	16528.0
78.00	1115.0
79.00	1167.0
80.00	846.0
81.00	1171.0
81.90	318.0
82.90	311.0
86.00	349.0
91.90	377.0
93.00	1912.0
98.00	1301.0
99.00	1320.0
101.00	610.0
103.90	431.0
105.00	412.0
107.00	4687.0
108.00	759.0
110.00	9544.0
110.90	1432.0
117.00	3492.0
118.00	335.0
123.00	634.0
127.00	18600.0
128.00	1433.0
129.00	6975.0
130.00	652.0
135.00	546.0
137.00	353.0
141.00	910.0
147.00	447.0
148.00	987.0
155.00	505.0
156.00	701.0
161.00	385.0
167.00	1665.0
168.00	1261.0
174.00	367.0
175.00	722.0
176.90	411.0
179.00	1340.0
180.00	1000.0
180.90	455.0
185.00	650.0
186.00	5040.0
187.00	1438.0
188.90	338.0
192.00	421.0
193.00	444.0
195.90	1157.0
198.00	38320.0
199.00	2714.0
204.00	1358.0
205.00	2222.0
206.00	9484.0
207.00	1640.0
208.00	355.0
211.00	457.0
217.00	2401.0
217.90	335.0
221.00	1964.0
223.00	583.0
224.00	5177.0
225.00	1414.0
227.00	2084.0
228.00	339.0
228.90	480.0
243.10	344.0
244.00	4308.0
245.00	601.0
246.00	772.0
255.00	22232.0
256.00	3238.0
257.90	1295.0
265.00	474.0
273.00	624.0
274.00	1830.0
275.00	10709.0
276.00	1507.0
277.00	814.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

296.00	2826.0
296.90	438.0
314.90	321.0
323.00	1010.0
334.00	675.0
354.00	305.0
364.90	1157.0
372.00	490.0
403.00	348.0
423.00	1805.0
424.00	563.0
441.00	3985.0
442.10	25488.0
443.10	5220.0
444.10	478.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

m/z	Abundance
39.00	1138.0
43.90	566.0
50.00	3618.0
51.00	14191.0
52.00	711.0
55.90	471.0
56.90	1061.0
63.00	563.0
68.20	313.0
69.00	16440.0
74.00	1533.0
75.00	2546.0
76.00	924.0
77.00	18152.0
78.00	1253.0
79.00	1199.0
80.00	893.0
81.00	1261.0
85.90	390.0
90.90	307.0
92.00	340.0
93.00	1996.0
98.00	1553.0
99.00	1351.0
100.90	748.0
103.90	433.0
105.00	454.0
107.00	5805.0
108.00	843.0
110.00	11221.0
111.00	1563.0
116.00	317.0
117.00	4362.0
117.90	339.0
121.90	405.0
122.90	527.0
127.00	21872.0
128.00	1743.0
129.00	8272.0
129.90	701.0
135.00	554.0
137.00	353.0
140.90	1098.0
141.90	352.0
147.00	590.0
148.00	1136.0
148.90	325.0
152.90	354.0
155.00	665.0
156.00	879.0
159.90	391.0
160.90	538.0
164.90	335.0
166.00	315.0
167.00	2136.0
168.00	1392.0
174.00	452.0
175.00	878.0
175.90	350.0
176.90	462.0
178.90	1715.0
180.00	1152.0
181.00	555.0
185.00	762.0
186.00	6388.0
187.00	1875.0
188.90	397.0
191.90	538.0
192.90	606.0
196.00	1533.0
198.00	49792.0
199.00	3302.0
203.00	321.0
204.00	1635.0
205.00	2841.0
206.00	12687.0
207.00	1794.0
208.00	429.0
211.00	530.0
217.00	3192.0
217.90	442.0
221.00	2352.0
221.80	483.0
223.00	776.0
224.00	7081.0
225.00	1744.0
227.00	2675.0
228.00	419.0
229.00	565.0
242.00	399.0
243.00	469.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

244.00	5847.0
245.10	691.0
246.00	1121.0
255.00	30576.0
256.00	4211.0
256.90	363.0
258.00	1708.0
264.90	689.0
273.00	998.0
274.00	2420.0
275.00	15737.0
276.00	2110.0
277.00	1248.0
296.00	3976.0
297.00	529.0
303.00	505.0
315.00	460.0
316.00	316.0
323.00	1324.0
334.00	942.0
346.00	306.0
352.00	437.0
353.10	315.0
354.00	557.0
365.00	2018.0
366.00	320.0
372.00	806.0
402.00	341.0
403.00	564.0
421.00	392.0
422.00	439.0
423.10	2556.0
424.00	761.0
441.10	6256.0
442.10	41936.0
443.10	8235.0
444.10	721.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

m/z	Abundance
39.00	1138.0
43.90	566.0
50.00	3618.0
51.00	14191.0
52.00	711.0
55.90	471.0
56.90	1061.0
63.00	563.0
68.20	313.0
69.00	16440.0
74.00	1533.0
75.00	2546.0
76.00	924.0
77.00	18152.0
78.00	1253.0
79.00	1199.0
80.00	893.0
81.00	1261.0
85.90	390.0
90.90	307.0
92.00	340.0
93.00	1996.0
98.00	1553.0
99.00	1351.0
100.90	748.0
103.90	433.0
105.00	454.0
107.00	5805.0
108.00	843.0
110.00	11221.0
111.00	1563.0
116.00	317.0
117.00	4362.0
117.90	339.0
121.90	405.0
122.90	527.0
127.00	21872.0
128.00	1743.0
129.00	8272.0
129.90	701.0
135.00	554.0
137.00	353.0
140.90	1098.0
141.90	352.0
147.00	590.0
148.00	1136.0
148.90	325.0
152.90	354.0
155.00	665.0
156.00	879.0
159.90	391.0
160.90	538.0
164.90	335.0
166.00	315.0
167.00	2136.0
168.00	1392.0
174.00	452.0
175.00	878.0
175.90	350.0
176.90	462.0
178.90	1715.0
180.00	1152.0
181.00	555.0
185.00	762.0
186.00	6388.0
187.00	1875.0
188.90	397.0
191.90	538.0
192.90	606.0
196.00	1533.0
198.00	49792.0
199.00	3302.0
203.00	321.0
204.00	1635.0
205.00	2841.0
206.00	12687.0
207.00	1794.0
208.00	429.0
211.00	530.0
217.00	3192.0
217.90	442.0
221.00	2352.0
221.80	483.0
223.00	776.0
224.00	7081.0
225.00	1744.0
227.00	2675.0
228.00	419.0
229.00	565.0
242.00	399.0
243.00	469.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

244.00	5847.0
245.10	691.0
246.00	1121.0
255.00	30576.0
256.00	4211.0
256.90	363.0
258.00	1708.0
264.90	689.0
273.00	998.0
274.00	2420.0
275.00	15737.0
276.00	2110.0
277.00	1248.0
296.00	3976.0
297.00	529.0
303.00	505.0
315.00	460.0
316.00	316.0
323.00	1324.0
334.00	942.0
346.00	306.0
352.00	437.0
353.10	315.0
354.00	557.0
365.00	2018.0
366.00	320.0
372.00	806.0
402.00	341.0
403.00	564.0
421.00	392.0
422.00	439.0
423.10	2556.0
424.00	761.0
441.10	6256.0
442.10	41936.0
443.10	8235.0
444.10	721.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

m/z	Abundance
39.00	1138.0
43.90	566.0
50.00	3618.0
51.00	14191.0
52.00	711.0
55.90	471.0
56.90	1061.0
63.00	563.0
68.20	313.0
69.00	16440.0
74.00	1533.0
75.00	2546.0
76.00	924.0
77.00	18152.0
78.00	1253.0
79.00	1199.0
80.00	893.0
81.00	1261.0
85.90	390.0
90.90	307.0
92.00	340.0
93.00	1996.0
98.00	1553.0
99.00	1351.0
100.90	748.0
103.90	433.0
105.00	454.0
107.00	5805.0
108.00	843.0
110.00	11221.0
111.00	1563.0
116.00	317.0
117.00	4362.0
117.90	339.0
121.90	405.0
122.90	527.0
127.00	21872.0
128.00	1743.0
129.00	8272.0
129.90	701.0
135.00	554.0
137.00	353.0
140.90	1098.0
141.90	352.0
147.00	590.0
148.00	1136.0
148.90	325.0
152.90	354.0
155.00	665.0
156.00	879.0
159.90	391.0
160.90	538.0
164.90	335.0
166.00	315.0
167.00	2136.0
168.00	1392.0
174.00	452.0
175.00	878.0
175.90	350.0
176.90	462.0
178.90	1715.0
180.00	1152.0
181.00	555.0
185.00	762.0
186.00	6388.0
187.00	1875.0
188.90	397.0
191.90	538.0
192.90	606.0
196.00	1533.0
198.00	49792.0
199.00	3302.0
203.00	321.0
204.00	1635.0
205.00	2841.0
206.00	12687.0
207.00	1794.0
208.00	429.0
211.00	530.0
217.00	3192.0
217.90	442.0
221.00	2352.0
221.80	483.0
223.00	776.0
224.00	7081.0
225.00	1744.0
227.00	2675.0
228.00	419.0
229.00	565.0
242.00	399.0
243.00	469.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

244.00	5847.0
245.10	691.0
246.00	1121.0
255.00	30576.0
256.00	4211.0
256.90	363.0
258.00	1708.0
264.90	689.0
273.00	998.0
274.00	2420.0
275.00	15737.0
276.00	2110.0
277.00	1248.0
296.00	3976.0
297.00	529.0
303.00	505.0
315.00	460.0
316.00	316.0
323.00	1324.0
334.00	942.0
346.00	306.0
352.00	437.0
353.10	315.0
354.00	557.0
365.00	2018.0
366.00	320.0
372.00	806.0
402.00	341.0
403.00	564.0
421.00	392.0
422.00	439.0
423.10	2556.0
424.00	761.0
441.10	6256.0
442.10	41936.0
443.10	8235.0
444.10	721.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

m/z	Abundance
39.00	658.0
43.90	464.0
50.00	2201.0
51.00	9075.0
52.00	473.0
57.00	651.0
63.00	404.0
69.00	10290.0
74.00	1058.0
75.00	1585.0
76.00	568.0
77.00	11764.0
78.00	760.0
79.00	807.0
79.90	595.0
81.00	892.0
93.00	1333.0
98.00	1036.0
99.00	825.0
101.00	525.0
104.00	310.0
104.90	314.0
107.00	3764.0
107.90	577.0
110.00	7153.0
111.00	1197.0
117.00	2910.0
123.00	383.0
127.00	14565.0
128.00	1028.0
129.00	5718.0
130.00	427.0
135.00	520.0
140.90	794.0
147.00	382.0
148.00	806.0
155.00	383.0
156.00	602.0
161.00	389.0
167.00	1492.0
168.00	867.0
175.00	646.0
176.90	351.0
178.90	1169.0
180.00	831.0
181.00	320.0
185.00	624.0
186.00	4507.0
187.00	1302.0
188.90	323.0
191.90	352.0
193.00	419.0
196.00	1074.0
198.00	34896.0
199.00	2372.0
204.00	1240.0
205.00	1973.0
206.00	8807.0
207.00	1498.0
208.00	363.0
211.00	386.0
217.00	2351.0
221.00	1950.0
223.00	562.0
224.00	5056.0
225.00	1305.0
227.00	2190.0
228.90	503.0
243.00	322.0
244.00	4318.0
245.00	596.0
246.00	807.0
255.00	23280.0
256.00	3460.0
258.00	1223.0
265.00	548.0
272.90	740.0
274.00	1866.0
275.00	11120.0
276.00	1513.0
277.00	911.0
296.00	3020.0
297.00	486.0
303.00	383.0
315.00	387.0
323.00	1090.0
334.00	788.0
354.00	440.0
365.00	1598.0
372.10	634.0
403.00	374.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

421.00	347.0
422.00	337.0
423.00	2288.0
424.10	613.0
441.10	5902.0
442.10	36824.0
443.10	7232.0
444.10	739.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67