


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN121318\
Data File : BN003886.D
Acq On : 13 Dec 2018 11:44
Operator : JU/SJ
Sample : DFTPP67
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN121218MA.M
Title : SVOA CALIBRATION
Last Update : Fri Dec 14 04:33:02 2018



AutoFind: Scans 2487, 2488, 2489; Background Corrected with Scan 2482

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.0	12285	PASS
68	69	0.00	2	0.7	104	PASS
69	198	0.00	100	34.1	13979	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	44.7	18345	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	41002	PASS
199	198	5	9	6.8	2796	PASS
275	198	10	60	30.5	12522	PASS
365	198	1	100	3.9	1591	PASS
441	198	0.01	100	13.1	5381	PASS
442	198	50	100	84.7	34749	PASS
443	442	15	24	19.8	6895	PASS

m/z	Abundance
43.90	583.0

Instrument :
BNA_N
ClientSampleId :

m/z	Abundance
39.00	1006.0
44.00	493.0
50.00	3353.0
51.00	13590.0
52.00	697.0
55.90	454.0
57.00	1002.0
63.00	615.0
69.00	15208.0
74.00	1466.0
75.00	2428.0
76.10	812.0
77.00	16528.0
78.00	1115.0
79.00	1167.0
80.00	846.0
81.00	1171.0
81.90	318.0
82.90	311.0
86.00	349.0
91.90	377.0
93.00	1912.0
98.00	1301.0
99.00	1320.0
101.00	610.0
103.90	431.0
105.00	412.0
107.00	4687.0
108.00	759.0
110.00	9544.0
110.90	1432.0
117.00	3492.0
118.00	335.0
123.00	634.0
127.00	18600.0
128.00	1433.0
129.00	6975.0
130.00	652.0
135.00	546.0
137.00	353.0
141.00	910.0
147.00	447.0
148.00	987.0
155.00	505.0
156.00	701.0
161.00	385.0
167.00	1665.0
168.00	1261.0
174.00	367.0
175.00	722.0
176.90	411.0
179.00	1340.0
180.00	1000.0
180.90	455.0
185.00	650.0
186.00	5040.0
187.00	1438.0
188.90	338.0
192.00	421.0
193.00	444.0
195.90	1157.0
198.00	38320.0
199.00	2714.0
204.00	1358.0
205.00	2222.0
206.00	9484.0
207.00	1640.0
208.00	355.0
211.00	457.0
217.00	2401.0
217.90	335.0
221.00	1964.0
223.00	583.0
224.00	5177.0
225.00	1414.0
227.00	2084.0
228.00	339.0
228.90	480.0
243.10	344.0
244.00	4308.0
245.00	601.0
246.00	772.0
255.00	22232.0
256.00	3238.0
257.90	1295.0
265.00	474.0
273.00	624.0
274.00	1830.0
275.00	10709.0
276.00	1507.0
277.00	814.0

Instrument :
BNA_N
ClientSampleId :

296.00	2826.0
296.90	438.0
314.90	321.0
323.00	1010.0
334.00	675.0
354.00	305.0
364.90	1157.0
372.00	490.0
403.00	348.0
423.00	1805.0
424.00	563.0
441.00	3985.0
442.10	25488.0
443.10	5220.0
444.10	478.0

Instrument :
BNA_N
ClientSampleId :

m/z	Abundance
39.00	1138.0
43.90	566.0
50.00	3618.0
51.00	14191.0
52.00	711.0
55.90	471.0
56.90	1061.0
63.00	563.0
68.20	313.0
69.00	16440.0
74.00	1533.0
75.00	2546.0
76.00	924.0
77.00	18152.0
78.00	1253.0
79.00	1199.0
80.00	893.0
81.00	1261.0
85.90	390.0
90.90	307.0
92.00	340.0
93.00	1996.0
98.00	1553.0
99.00	1351.0
100.90	748.0
103.90	433.0
105.00	454.0
107.00	5805.0
108.00	843.0
110.00	11221.0
111.00	1563.0
116.00	317.0
117.00	4362.0
117.90	339.0
121.90	405.0
122.90	527.0
127.00	21872.0
128.00	1743.0
129.00	8272.0
129.90	701.0
135.00	554.0
137.00	353.0
140.90	1098.0
141.90	352.0
147.00	590.0
148.00	1136.0
148.90	325.0
152.90	354.0
155.00	665.0
156.00	879.0
159.90	391.0
160.90	538.0
164.90	335.0
166.00	315.0
167.00	2136.0
168.00	1392.0
174.00	452.0
175.00	878.0
175.90	350.0
176.90	462.0
178.90	1715.0
180.00	1152.0
181.00	555.0
185.00	762.0
186.00	6388.0
187.00	1875.0
188.90	397.0
191.90	538.0
192.90	606.0
196.00	1533.0
198.00	49792.0
199.00	3302.0
203.00	321.0
204.00	1635.0
205.00	2841.0
206.00	12687.0
207.00	1794.0
208.00	429.0
211.00	530.0
217.00	3192.0
217.90	442.0
221.00	2352.0
221.80	483.0
223.00	776.0
224.00	7081.0
225.00	1744.0
227.00	2675.0
228.00	419.0
229.00	565.0
242.00	399.0
243.00	469.0

Instrument :
BNA_N
ClientSampleId :

244.00	5847.0
245.10	691.0
246.00	1121.0
255.00	30576.0
256.00	4211.0
256.90	363.0
258.00	1708.0
264.90	689.0
273.00	998.0
274.00	2420.0
275.00	15737.0
276.00	2110.0
277.00	1248.0
296.00	3976.0
297.00	529.0
303.00	505.0
315.00	460.0
316.00	316.0
323.00	1324.0
334.00	942.0
346.00	306.0
352.00	437.0
353.10	315.0
354.00	557.0
365.00	2018.0
366.00	320.0
372.00	806.0
402.00	341.0
403.00	564.0
421.00	392.0
422.00	439.0
423.10	2556.0
424.00	761.0
441.10	6256.0
442.10	41936.0
443.10	8235.0
444.10	721.0

Instrument :
BNA_N
ClientSampleId :

m/z	Abundance
39.00	1138.0
43.90	566.0
50.00	3618.0
51.00	14191.0
52.00	711.0
55.90	471.0
56.90	1061.0
63.00	563.0
68.20	313.0
69.00	16440.0
74.00	1533.0
75.00	2546.0
76.00	924.0
77.00	18152.0
78.00	1253.0
79.00	1199.0
80.00	893.0
81.00	1261.0
85.90	390.0
90.90	307.0
92.00	340.0
93.00	1996.0
98.00	1553.0
99.00	1351.0
100.90	748.0
103.90	433.0
105.00	454.0
107.00	5805.0
108.00	843.0
110.00	11221.0
111.00	1563.0
116.00	317.0
117.00	4362.0
117.90	339.0
121.90	405.0
122.90	527.0
127.00	21872.0
128.00	1743.0
129.00	8272.0
129.90	701.0
135.00	554.0
137.00	353.0
140.90	1098.0
141.90	352.0
147.00	590.0
148.00	1136.0
148.90	325.0
152.90	354.0
155.00	665.0
156.00	879.0
159.90	391.0
160.90	538.0
164.90	335.0
166.00	315.0
167.00	2136.0
168.00	1392.0
174.00	452.0
175.00	878.0
175.90	350.0
176.90	462.0
178.90	1715.0
180.00	1152.0
181.00	555.0
185.00	762.0
186.00	6388.0
187.00	1875.0
188.90	397.0
191.90	538.0
192.90	606.0
196.00	1533.0
198.00	49792.0
199.00	3302.0
203.00	321.0
204.00	1635.0
205.00	2841.0
206.00	12687.0
207.00	1794.0
208.00	429.0
211.00	530.0
217.00	3192.0
217.90	442.0
221.00	2352.0
221.80	483.0
223.00	776.0
224.00	7081.0
225.00	1744.0
227.00	2675.0
228.00	419.0
229.00	565.0
242.00	399.0
243.00	469.0

Instrument :
BNA_N
ClientSampleId :

244.00	5847.0
245.10	691.0
246.00	1121.0
255.00	30576.0
256.00	4211.0
256.90	363.0
258.00	1708.0
264.90	689.0
273.00	998.0
274.00	2420.0
275.00	15737.0
276.00	2110.0
277.00	1248.0
296.00	3976.0
297.00	529.0
303.00	505.0
315.00	460.0
316.00	316.0
323.00	1324.0
334.00	942.0
346.00	306.0
352.00	437.0
353.10	315.0
354.00	557.0
365.00	2018.0
366.00	320.0
372.00	806.0
402.00	341.0
403.00	564.0
421.00	392.0
422.00	439.0
423.10	2556.0
424.00	761.0
441.10	6256.0
442.10	41936.0
443.10	8235.0
444.10	721.0

Instrument :
BNA_N
ClientSampleId :

m/z	Abundance
39.00	1138.0
43.90	566.0
50.00	3618.0
51.00	14191.0
52.00	711.0
55.90	471.0
56.90	1061.0
63.00	563.0
68.20	313.0
69.00	16440.0
74.00	1533.0
75.00	2546.0
76.00	924.0
77.00	18152.0
78.00	1253.0
79.00	1199.0
80.00	893.0
81.00	1261.0
85.90	390.0
90.90	307.0
92.00	340.0
93.00	1996.0
98.00	1553.0
99.00	1351.0
100.90	748.0
103.90	433.0
105.00	454.0
107.00	5805.0
108.00	843.0
110.00	11221.0
111.00	1563.0
116.00	317.0
117.00	4362.0
117.90	339.0
121.90	405.0
122.90	527.0
127.00	21872.0
128.00	1743.0
129.00	8272.0
129.90	701.0
135.00	554.0
137.00	353.0
140.90	1098.0
141.90	352.0
147.00	590.0
148.00	1136.0
148.90	325.0
152.90	354.0
155.00	665.0
156.00	879.0
159.90	391.0
160.90	538.0
164.90	335.0
166.00	315.0
167.00	2136.0
168.00	1392.0
174.00	452.0
175.00	878.0
175.90	350.0
176.90	462.0
178.90	1715.0
180.00	1152.0
181.00	555.0
185.00	762.0
186.00	6388.0
187.00	1875.0
188.90	397.0
191.90	538.0
192.90	606.0
196.00	1533.0
198.00	49792.0
199.00	3302.0
203.00	321.0
204.00	1635.0
205.00	2841.0
206.00	12687.0
207.00	1794.0
208.00	429.0
211.00	530.0
217.00	3192.0
217.90	442.0
221.00	2352.0
221.80	483.0
223.00	776.0
224.00	7081.0
225.00	1744.0
227.00	2675.0
228.00	419.0
229.00	565.0
242.00	399.0
243.00	469.0

Instrument :
BNA_N
ClientSampleId :

244.00	5847.0
245.10	691.0
246.00	1121.0
255.00	30576.0
256.00	4211.0
256.90	363.0
258.00	1708.0
264.90	689.0
273.00	998.0
274.00	2420.0
275.00	15737.0
276.00	2110.0
277.00	1248.0
296.00	3976.0
297.00	529.0
303.00	505.0
315.00	460.0
316.00	316.0
323.00	1324.0
334.00	942.0
346.00	306.0
352.00	437.0
353.10	315.0
354.00	557.0
365.00	2018.0
366.00	320.0
372.00	806.0
402.00	341.0
403.00	564.0
421.00	392.0
422.00	439.0
423.10	2556.0
424.00	761.0
441.10	6256.0
442.10	41936.0
443.10	8235.0
444.10	721.0

Instrument :
BNA_N
ClientSampleId :

m/z	Abundance
39.00	658.0
43.90	464.0
50.00	2201.0
51.00	9075.0
52.00	473.0
57.00	651.0
63.00	404.0
69.00	10290.0
74.00	1058.0
75.00	1585.0
76.00	568.0
77.00	11764.0
78.00	760.0
79.00	807.0
79.90	595.0
81.00	892.0
93.00	1333.0
98.00	1036.0
99.00	825.0
101.00	525.0
104.00	310.0
104.90	314.0
107.00	3764.0
107.90	577.0
110.00	7153.0
111.00	1197.0
117.00	2910.0
123.00	383.0
127.00	14565.0
128.00	1028.0
129.00	5718.0
130.00	427.0
135.00	520.0
140.90	794.0
147.00	382.0
148.00	806.0
155.00	383.0
156.00	602.0
161.00	389.0
167.00	1492.0
168.00	867.0
175.00	646.0
176.90	351.0
178.90	1169.0
180.00	831.0
181.00	320.0
185.00	624.0
186.00	4507.0
187.00	1302.0
188.90	323.0
191.90	352.0
193.00	419.0
196.00	1074.0
198.00	34896.0
199.00	2372.0
204.00	1240.0
205.00	1973.0
206.00	8807.0
207.00	1498.0
208.00	363.0
211.00	386.0
217.00	2351.0
221.00	1950.0
223.00	562.0
224.00	5056.0
225.00	1305.0
227.00	2190.0
228.90	503.0
243.00	322.0
244.00	4318.0
245.00	596.0
246.00	807.0
255.00	23280.0
256.00	3460.0
258.00	1223.0
265.00	548.0
272.90	740.0
274.00	1866.0
275.00	11120.0
276.00	1513.0
277.00	911.0
296.00	3020.0
297.00	486.0
303.00	383.0
315.00	387.0
323.00	1090.0
334.00	788.0
354.00	440.0
365.00	1598.0
372.10	634.0
403.00	374.0

Instrument :
BNA_N
ClientSampleId :

421.00	347.0
422.00	337.0
423.00	2288.0
424.10	613.0
441.10	5902.0
442.10	36824.0
443.10	7232.0
444.10	739.0

Instrument :
BNA_N
ClientSampleId :