

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121518\
 Data File : BN003963.D
 Acq On : 15 Dec 2018 17:28
 Operator : JU/SJ
 Sample : J6227-19DL 5X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F9E68DL

Quant Time: Dec 15 23:05:12 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 15 21:58:36 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	2093	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	10391	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	6133	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	15019	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	19418	0.40	ng/ul	0.00
20) Perylene-d12	23.72	264	17802	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	689	0.04	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	1907	0.04	ng/ul	0.00
Target Compounds				Ovalue		
5) 2-Methylnaphthalene	12.29	142	1170	0.057	ng/ul	100
12) Phenanthrene	17.25	178	4303	0.087	ng/ul	95
15) Fluoranthene	19.27	202	1198	0.021	ng/ul	80
17) Pyrene	19.64	202	5777	0.083	ng/ul	96
18) Benzo(a)anthracene	21.38	228	2213	0.036	ng/ul#	66
19) Chrysene	21.44	228	4157	0.059	ng/ul#	77
23) Benzo(a)pyrene	23.61	252	1342	0.021	ng/ul#	9
26) Benzo(g,h,i)perylene	26.80	276	1930	0.030	ng/ul#	53

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121518\
Data File : BN003963.D
Acq On : 15 Dec 2018 17:28
Operator : JU/SJ
Sample : J6227-19DL 5X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9E68DL

Quant Time: Dec 15 23:05:12 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 15 21:58:36 2018
Response via : Initial Calibration

