

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121522\
 Data File : BN023215.D
 Acq On : 15 Dec 2022 14:18
 Operator : CG/JU
 Sample : SSTD02049
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020212

Quant Time: Dec 15 22:36:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 15 22:31:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.004	152	290652	20.000	ng/u1	0.00
20) Naphthalene-d8	10.822	136	1370033	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	898827	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.398	188	1997598	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	1980022	20.000	ng/u1	0.00
88) Perylene-d12	24.039	264	2129819	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.346	96	55300	6.496	ng/uL	0.00
4) Pyridine-d5	3.775	84	395036	16.408	ng/u1	0.00
7) Phenol-d5	7.157	99	531295	19.354	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	323115	19.631	ng/u1	0.00
11) 2-Chlorophenol-d4	7.528	132	416085	20.697	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	429801	19.832	ng/u1	0.00
21) Nitrobenzene-d5	9.175	128	213224	20.107	ng/u1	0.00
24) 2-Nitrophenol-d4	9.899	143	225236	19.035	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.440	165	423552	20.997	ng/u1	0.00
31) 4-Chloroaniline-d4	10.963	131	682973	22.227	ng/u1	0.00
46) Dimethylphthalate-d6	14.063	166	1343411	20.549	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	1522430	21.135	ng/u1	0.00
54) 4-Nitrophenol-d4	14.845	143	284287	21.813	ng/u1	0.00
60) Fluorene-d10	15.645	176	1135970	21.186	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	219323	17.273	ng/u1	0.00
73) Anthracene-d10	17.498	188	1820704	20.815	ng/u1	0.00
81) Pyrene-d10	19.786	212	2206872	21.507	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.874	264	2183120	20.391	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.381	88	57402	6.336	ng/uL	100
5) Pyridine	3.793	79	416990	17.468	ng/u1	100
6) Benzaldehyde	7.134	77	207169	14.916	ng/u1	100
8) Phenol	7.187	94	537608	18.803	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.422	93	436317	19.584	ng/u1	100
12) 2-Chlorophenol	7.563	128	428558	19.964	ng/u1	100
13) 2-Methylphenol	8.452	108	417013	19.595	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.534	45	561824	19.104	ng/u1	100
16) Acetophenone	8.834	105	688280	19.898	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.822	70	348000	18.877	ng/u1	100
18) 4-Methylphenol	8.781	108	462750	20.025	ng/u1	100
19) Hexachloroethane	9.093	117	165985	18.263	ng/u1	100
22) Nitrobenzene	9.216	77	511450	17.801	ng/u1	100
23) Isophorone	9.746	82	990718	18.027	ng/u1	100
25) 2-Nitrophenol	9.934	139	245479	18.870	ng/u1	100
26) 2,4-Dimethylphenol	9.993	107	496037	17.776	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.234	93	626224	19.760	ng/u1	100
29) 2,4-Dichlorophenol	10.469	162	419898	20.480	ng/u1	100
30) Naphthalene	10.875	128	1473372	20.262	ng/u1	100
32) 4-Chloroaniline	10.987	127	685200	21.353	ng/u1	100
33) Hexachlorobutadiene	11.163	225	240238	19.036	ng/u1	100
34) Caprolactam	11.757	113	82758	10.640	ng/u1	100
35) 4-Chloro-3-methylphenol	12.104	107	483552	18.083	ng/u1	100
36) 2-Methylnaphthalene	12.481	142	1000885	20.204	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121522\
 Data File : BN023215.D
 Acq On : 15 Dec 2022 14:18
 Operator : CG/JU
 Sample : SST02049
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SST020212


Quant Time: Dec 15 22:36:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 15 22:31:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.698	142	1044011	21.024	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.845	216	504021	20.892	ng/ul	100
40) Hexachlorocyclopentadiene	12.828	237	280445	19.910	ng/ul	100
41) 2,4,6-Trichlorophenol	13.087	196	347903	20.310	ng/ul	100
42) 2,4,5-Trichlorophenol	13.157	196	385229	21.237	ng/ul	100
43) 1,1'-Biphenyl	13.487	154	1351945	20.492	ng/ul	100
44) 2-Chloronaphthalene	13.528	162	1042684	20.282	ng/ul	100
45) 2-Nitroaniline	13.734	65	310249	16.617	ng/ul	100
47) Dimethylphthalate	14.110	163	1348311	19.622	ng/ul	100
48) 2,6-Dinitrotoluene	14.228	165	269768	19.394	ng/ul	100
50) Acenaphthylene	14.375	152	1665618	20.698	ng/ul	100
51) 3-Nitroaniline	14.557	138	257159	17.170	ng/ul	100
52) Acenaphthene	14.716	153	1163915	20.385	ng/ul	100
53) 2,4-Dinitrophenol	14.757	184	137719	14.167	ng/ul	100
55) 4-Nitrophenol	14.857	109	214194	16.580	ng/ul	100
56) Dibenzofuran	15.051	168	1595731	21.153	ng/ul	100
57) 2,4-Dinitrotoluene	15.010	165	398600	19.678	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.275	232	335907	21.622	ng/ul	100
59) Diethylphthalate	15.469	149	1355300	18.609	ng/ul	100
61) Fluorene	15.698	166	1325492	21.040	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.692	204	631847	21.277	ng/ul	100
63) 4-Nitroaniline	15.722	138	254975	19.090	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.775	198	218985	16.686	ng/ul	100
67) N-Nitrosodiphenylamine	15.904	169	1126770	18.840	ng/ul	100
68) 4-Bromophenyl-phenylether	16.586	248	391836	19.919	ng/ul	100
69) Hexachlorobenzene	16.704	284	452248	19.667	ng/ul	100
70) Atrazine	16.857	200	397227	18.633	ng/ul	100
71) Pentachlorophenol	17.045	266	284954	19.504	ng/ul	100
72) Phenanthrene	17.439	178	2137904	19.856	ng/ul	100
74) Anthracene	17.533	178	2187986	20.062	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.451	216	501316	19.030	ng/ul	100
76) Pentachlorobenzene	14.969	250	508465	17.884	ng/ul	100
77) Carbazole	17.804	167	2052277	20.253	ng/ul	100
78) Di-n-butylphthalate	18.363	149	2357996	16.351	ng/ul	100
80) Fluoranthene	19.451	202	2514261	19.632	ng/ul	100
82) Pyrene	19.816	202	2685508	20.471	ng/ul	100
83) Butylbenzylphthalate	20.710	149	1142230	17.105	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.498	252	908370	20.792	ng/ul	100
85) Benzo(a)anthracene	21.568	228	2695454	20.615	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.486	149	1618010	16.416	ng/ul	100
87) Chrysene	21.621	228	2574537	20.792	ng/ul	100
89) Di-n-octyl phthalate	22.427	149	2768339	14.779	ng/ul	100
90) Benzo(b)fluoranthene	23.280	252	2796055	19.372	ng/ul	100
91) Benzo(k)fluoranthene	23.333	252	2698369	19.695	ng/ul	100
93) Benzo(a)pyrene	23.927	252	2436974	19.919	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.598	276	3024025	21.734	ng/ul	100
95) Dibenzo(a,h)anthracene	26.621	278	2525428	20.251	ng/ul	100
96) Benzo(g,h,i)perylene	27.392	276	2317649	19.104	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 15 22:36:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 15 22:31:27 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 6.336 ng/uL
 RT: 3.381 min Scan# 50
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020212


Tgt Ion: 88 Resp: 57402
 Ion Ratio Lower Upper
 88 100
 43 28.8 23.0 34.6
 58 69.7 55.8 83.6



#5
 Pyridine
 Concen: 17.468 ng/uL
 RT: 3.793 min Scan# 120
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18

Tgt Ion: 79 Resp: 416990
 Ion Ratio Lower Upper
 79 100
 52 70.9 56.7 85.1
 51 32.5 26.0 39.0









#6
Benzaldehyde
Concen: 14.916 ng/ul
RT: 7.134 min Scan# 688
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion: 77 Resp: 207169
Ion Ratio Lower Upper
77 100
105 97.5 78.0 117.0
106 91.5 73.2 109.8



#8
Phenol
Concen: 18.803 ng/ul
RT: 7.187 min Scan# 697
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion: 94 Resp: 537608
Ion Ratio Lower Upper
94 100
65 27.2 21.8 32.6
66 35.5 28.4 42.6









#10
Bis(2-Chloroethyl)ether
Concen: 19.584 ng/ul
RT: 7.422 min Scan# 737
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Instrument :
BNA_N
ClientSampleId :
SSTD020212

Tgt Ion: 93 Resp: 436317




Ion	Ratio	Lower	Upper
93	100		
63	73.6	58.9	88.3
95	32.8	26.2	39.4




#12
2-Chlorophenol
Concen: 19.964 ng/ul
RT: 7.563 min Scan# 761
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:128 Resp: 428558

Ion	Ratio	Lower	Upper
128	100		
64	42.5	34.0	51.0
130	31.2	25.0	37.4









#13
2-Methylphenol
Concen: 19.595 ng/ul
RT: 8.452 min Scan# 912
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:108 Resp: 417013
Ion Ratio Lower Upper
108 100
107 89.0 71.2 106.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.104 ng/ul
RT: 8.534 min Scan# 926
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion: 45 Resp: 561824
Ion Ratio Lower Upper
45 100
77 15.8 12.6 19.0
79 11.0 8.8 13.2









#16
 Acetophenone
 Concen: 19.898 ng/ul
 RT: 8.834 min Scan# 977
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18




Instrument :
 BNA_N
 ClientSampleId :
 SST020212


Tgt Ion: 105 Resp: 688280
 Ion Ratio Lower Upper
 105 100
 77 76.7 61.4 92.0
 51 25.0 20.0 30.0



#17
 N-Nitroso-di-n-propylamine
 Concen: 18.877 ng/ul
 RT: 8.822 min Scan# 975
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18

Tgt Ion: 70 Resp: 348000
 Ion Ratio Lower Upper
 70 100
 42 56.6 45.3 67.9
 101 10.1 8.1 12.1
 130 22.4 17.9 26.9









#18
4-Methylphenol
Concen: 20.025 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:108 Resp: 462750
Ion Ratio Lower Upper
108 100
107 109.5 87.6 131.4



#19
Hexachloroethane
Concen: 18.263 ng/ul
RT: 9.093 min Scan# 1021
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:117 Resp: 165985
Ion Ratio Lower Upper
117 100
201 93.2 74.6 111.8
199 57.9 46.3 69.5









#22
Nitrobenzene
Concen: 17.801 ng/ul
RT: 9.216 min Scan# 1042
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion: 77 Resp: 511450
Ion Ratio Lower Upper
77 100
123 46.4 37.1 55.7
65 12.6 10.1 15.1



#23
Isophorone
Concen: 18.027 ng/ul
RT: 9.746 min Scan# 1132
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion: 82 Resp: 990718
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 18.1 14.5 21.7









#25
2-Nitrophenol
Concen: 18.870 ng/ul
RT: 9.934 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:139 Resp: 245479
Ion Ratio Lower Upper
139 100
65 47.1 37.7 56.5
109 28.3 22.6 34.0



#26
2,4-Dimethylphenol
Concen: 17.776 ng/ul
RT: 9.993 min Scan# 1174
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:107 Resp: 496037
Ion Ratio Lower Upper
107 100
121 51.6 41.3 61.9
122 90.1 72.1 108.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.760 ng/ul
 RT: 10.234 min Scan# 1215
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020212


Tgt Ion: 93 Resp: 626224
 Ion Ratio Lower Upper
 93 100
 95 32.2 25.8 38.6
 123 11.2 9.0 13.4



#29
 2,4-Dichlorophenol
 Concen: 20.480 ng/ul
 RT: 10.469 min Scan# 1255
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18

Tgt Ion: 162 Resp: 419898
 Ion Ratio Lower Upper
 162 100
 164 65.1 52.1 78.1
 98 36.6 29.3 43.9





#30

Naphthalene

Concen: 20.262 ng/ul

RT: 10.875 min Scan# 1324

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:128 Resp: 1473372

Ion Ratio Lower Upper

128 100

129 10.9 8.7 13.1

127 12.9 10.3 15.5



#32

4-Chloroaniline

Concen: 21.353 ng/ul

RT: 10.987 min Scan# 1343

Delta R.T. 0.000 min

Lab File: BN023215.D




Acq: 15 Dec 2022 14:18


Tgt Ion:127 Resp: 685200

Ion Ratio Lower Upper

127 100

129 32.8 26.2 39.4





#33

Hexachlorobutadiene

Concen: 19.036 ng/ul

RT: 11.163 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:225 Resp: 240238

Ion Ratio Lower Upper

225 100

223 61.1 48.9 73.3

227 63.7 51.0 76.4



#34

Caprolactam

Concen: 10.640 ng/ul

RT: 11.757 min Scan# 1474

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18




Tgt Ion:113 Resp: 82758


Ion Ratio Lower Upper

113 100

55 154.6 123.7 185.5

56 114.6 91.7 137.5









#35
4-Chloro-3-methylphenol
Concen: 18.083 ng/ul
RT: 12.104 min Scan# 1533
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:107 Resp: 483552
Ion Ratio Lower Upper
107 100
144 27.2 21.8 32.6
142 78.3 62.6 94.0



#36
2-Methylnaphthalene
Concen: 20.204 ng/ul
RT: 12.481 min Scan# 1597
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:142 Resp: 1000885
Ion Ratio Lower Upper
142 100
141 87.4 69.9 104.9





#37

1-Methylnaphthalene

Concen: 21.024 ng/ul

RT: 12.698 min Scan# 1634

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:142 Resp: 1044011

Ion Ratio Lower Upper

142 100

141 88.6 70.9 106.3

116 3.8 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.892 ng/ul

RT: 12.845 min Scan# 1659

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18

Tgt Ion:216 Resp: 504021




Ion Ratio Lower Upper


216 100

214 76.4 61.1 91.7

179 20.7 16.6 24.8

108 19.3 15.4 23.2









#40
Hexachlorocyclopentadiene
Concen: 19.910 ng/ul
RT: 12.828 min Scan# 1656
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:237 Resp: 280445
Ion Ratio Lower Upper
237 100
235 60.1 48.1 72.1
272 13.5 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 20.310 ng/ul
RT: 13.087 min Scan# 1700
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:196 Resp: 347903
Ion Ratio Lower Upper
196 100
198 96.1 76.9 115.3
200 29.7 23.8 35.6









#42
 2,4,5-Trichlorophenol
 Concen: 21.237 ng/ul
 RT: 13.157 min Scan# 1712
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020212

Tgt Ion:196 Resp: 385229




Ion	Ratio	Lower	Upper
196	100		
198	93.5	74.8	112.2
200	30.2	24.2	36.2




#43
 1,1'-Biphenyl
 Concen: 20.492 ng/ul
 RT: 13.487 min Scan# 1768
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18

Tgt Ion:154 Resp: 1351945

Ion	Ratio	Lower	Upper
154	100		
153	40.0	32.0	48.0
76	13.7	11.0	16.4









#44
2-Chloronaphthalene
Concen: 20.282 ng/ul
RT: 13.528 min Scan# 1775
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:162 Resp: 1042684
Ion Ratio Lower Upper
162 100
164 31.7 25.4 38.0
127 39.8 31.8 47.8



#45
2-Nitroaniline
Concen: 16.617 ng/ul
RT: 13.734 min Scan# 1810
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion: 65 Resp: 310249
Ion Ratio Lower Upper
65 100
92 71.8 57.4 86.2
138 111.3 89.0 133.6









#47
Dimethylphthalate
Concen: 19.622 ng/ul
RT: 14.110 min Scan# 1874
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:163 Resp: 1348311
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.4 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 19.394 ng/ul
RT: 14.228 min Scan# 1894
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:165 Resp: 269768
Ion Ratio Lower Upper
165 100
89 59.6 47.7 71.5
121 22.3 17.8 26.8









#50
Acenaphthylene
Concen: 20.698 ng/ul
RT: 14.375 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:152 Resp: 1665618
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4
153 14.2 11.4 17.0



#51
3-Nitroaniline
Concen: 17.170 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:138 Resp: 257159
Ion Ratio Lower Upper
138 100
108 10.4 8.3 12.5
92 115.6 92.5 138.7





#52

Acenaphthene

Concen: 20.385 ng/ul

RT: 14.716 min Scan# 1977

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:153 Resp: 1163915

Ion Ratio Lower Upper

153 100

152 48.6 38.9 58.3

154 88.1 70.5 105.7



#53

2,4-Dinitrophenol

Concen: 14.167 ng/ul

RT: 14.757 min Scan# 1984

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18




Tgt Ion:184 Resp: 137719


Ion Ratio Lower Upper

184 100

63 71.4 57.1 85.7

154 59.6 47.7 71.5









#55
4-Nitrophenol
Concen: 16.580 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:109 Resp: 214194
Ion Ratio Lower Upper
109 100
139 144.0 115.2 172.8
65 136.2 109.0 163.4



#56
Dibenzofuran
Concen: 21.153 ng/ul
RT: 15.051 min Scan# 2034
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:168 Resp: 1595731
Ion Ratio Lower Upper
168 100
139 36.5 29.2 43.8









#57
2,4-Dinitrotoluene
Concen: 19.678 ng/ul
RT: 15.010 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:165 Resp: 398600
Ion Ratio Lower Upper
165 100
63 45.7 36.6 54.8
182 3.1 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.622 ng/ul
RT: 15.275 min Scan# 2072
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:232 Resp: 335907
Ion Ratio Lower Upper
232 100
131 47.4 37.9 56.9
130 2.4 1.9 2.9
166 29.8 23.8 35.8









#59
Diethylphthalate
Concen: 18.609 ng/ul
RT: 15.469 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:149 Resp: 1355300
Ion Ratio Lower Upper
149 100
177 21.2 17.0 25.4
150 12.5 10.0 15.0



#61
Fluorene
Concen: 21.040 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:166 Resp: 1325492
Ion Ratio Lower Upper
166 100
165 97.3 77.8 116.8
167 13.2 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 21.277 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:204 Resp: 631847
Ion Ratio Lower Upper
204 100
206 32.6 26.1 39.1
141 60.2 48.2 72.2



#63
4-Nitroaniline
Concen: 19.090 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:138 Resp: 254975
Ion Ratio Lower Upper
138 100
92 49.3 39.4 59.2
108 74.4 59.5 89.3









#66
4,6-Dinitro-2-methylphenol
Concen: 16.686 ng/ul
RT: 15.775 min Scan# 2157
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:198 Resp: 218985
Ion Ratio Lower Upper
198 100
182 4.2 3.4 5.0
77 24.6 19.7 29.5



#67
N-Nitrosodiphenylamine
Concen: 18.840 ng/ul
RT: 15.904 min Scan# 2179
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:169 Resp: 1126770
Ion Ratio Lower Upper
169 100
168 68.4 54.7 82.1
167 36.5 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 19.919 ng/ul
RT: 16.586 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:248 Resp: 391836
Ion Ratio Lower Upper
248 100
250 101.2 81.0 121.4
141 76.0 60.8 91.2



#69
Hexachlorobenzene
Concen: 19.667 ng/ul
RT: 16.704 min Scan# 2315
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:284 Resp: 452248
Ion Ratio Lower Upper
284 100
142 36.2 29.0 43.4
249 29.6 23.7 35.5









#70
 Atrazine
 Concen: 18.633 ng/ul
 RT: 16.857 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020212


Tgt Ion:200 Resp: 397227
 Ion Ratio Lower Upper
 200 100
 173 26.6 21.3 31.9
 215 49.8 39.8 59.8



#71
 Pentachlorophenol
 Concen: 19.504 ng/ul
 RT: 17.045 min Scan# 2373
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18

Tgt Ion:266 Resp: 284954
 Ion Ratio Lower Upper
 266 100
 264 63.1 50.5 75.7
 268 65.7 52.6 78.8









#72
 Phenanthrene
 Concen: 19.856 ng/u1
 RT: 17.439 min Scan# 2440
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020212

Tgt Ion:178 Resp: 2137904




Ion	Ratio	Lower	Upper
178	100		
179	15.4	12.3	18.5
176	19.2	15.4	23.0




#74
 Anthracene
 Concen: 20.062 ng/u1
 RT: 17.533 min Scan# 2456
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18

Tgt Ion:178 Resp: 2187986

Ion	Ratio	Lower	Upper
178	100		
179	14.7	11.8	17.6
176	18.4	14.7	22.1









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 19.030 ng/uL
 RT: 13.451 min Scan# 1762
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18




Instrument :
 BNA_N
 ClientSampleId :
 SST020212


Tgt Ion:216 Resp: 501316
 Ion Ratio Lower Upper
 216 100
 214 77.4 61.9 92.9
 218 49.3 39.4 59.2



#76
 Pentachlorobenzene
 Concen: 17.884 ng/uL
 RT: 14.969 min Scan# 2020
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18

Tgt Ion:250 Resp: 508465
 Ion Ratio Lower Upper
 250 100
 248 61.7 49.4 74.0
 252 66.1 52.9 79.3









#77
 Carbazole
 Concen: 20.253 ng/ul
 RT: 17.804 min Scan# 2502
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020212


Tgt Ion:167 Resp: 2052277
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.3 25.9
 139 13.3 10.6 16.0



#78
 Di-n-butylphthalate
 Concen: 16.351 ng/ul
 RT: 18.363 min Scan# 2597
 Delta R.T. 0.000 min
 Lab File: BN023215.D
 Acq: 15 Dec 2022 14:18

Tgt Ion:149 Resp: 2357996
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.2
 104 5.1 4.1 6.1





#80

Fluoranthene

Concen: 19.632 ng/ul

RT: 19.451 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:202 Resp: 2514261

Ion Ratio Lower Upper

202 100

101 13.0 10.4 15.6

100 9.5 7.6 11.4



#82

Pyrene

Concen: 20.471 ng/ul

RT: 19.816 min Scan# 2844

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18




Tgt Ion:202 Resp: 2685508


Ion Ratio Lower Upper

202 100

101 15.0 12.0 18.0

100 11.9 9.5 14.3





#83

Butylbenzylphthalate

Concen: 17.105 ng/ul

RT: 20.710 min Scan# 2996

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:149 Resp: 1142230

Ion Ratio Lower Upper

149 100

91 72.4 57.9 86.9

206 19.5 15.6 23.4



#84

3,3'-Dichlorobenzidine

Concen: 20.792 ng/ul

RT: 21.498 min Scan# 3130

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18




Tgt Ion:252 Resp: 908370


Ion Ratio Lower Upper

252 100

254 62.9 50.3 75.5

126 11.7 9.4 14.0





#85

Benzo(a)anthracene

Concen: 20.615 ng/ul

RT: 21.568 min Scan# 3142

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:228 Resp: 2695454

Ion Ratio Lower Upper

228 100

229 19.4 15.5 23.3

226 26.7 21.4 32.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 16.416 ng/ul

RT: 21.486 min Scan# 3128

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18




Tgt Ion:149 Resp: 1618010


Ion Ratio Lower Upper

149 100

167 27.4 21.9 32.9

279 4.2 3.4 5.0





#87

Chrysene

Concen: 20.792 ng/ul

RT: 21.621 min Scan# 3151

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:228 Resp: 2574537

Ion Ratio Lower Upper

228 100

226 30.7 24.6 36.8

229 20.6 16.5 24.7



#89

Di-n-octyl phthalate

Concen: 14.779 ng/ul




RT: 22.427 min Scan# 3288


Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18

Tgt Ion:149 Resp: 2768339









#90
Benzo(b)fluoranthene
Concen: 19.372 ng/ul
RT: 23.280 min Scan# 3433
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:252 Resp: 2796055
Ion Ratio Lower Upper
252 100
253 21.0 16.8 25.2
125 11.0 8.8 13.2



#91
Benzo(k)fluoranthene
Concen: 19.695 ng/ul
RT: 23.333 min Scan# 3442
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:252 Resp: 2698369
Ion Ratio Lower Upper
252 100
253 22.4 17.9 26.9
125 11.3 9.0 13.6









#93
Benzo(a)pyrene
Concen: 19.919 ng/ul
RT: 23.927 min Scan# 3543
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18




Instrument :
BNA_N
ClientSampleId :
SSTD020212


Tgt Ion:252 Resp: 2436974
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 11.8 9.4 14.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 21.734 ng/ul
RT: 26.598 min Scan# 3997
Delta R.T. 0.000 min
Lab File: BN023215.D
Acq: 15 Dec 2022 14:18

Tgt Ion:276 Resp: 3024025
Ion Ratio Lower Upper
276 100
138 25.7 20.6 30.8
277 22.7 18.2 27.2





#95

Dibenzo(a,h)anthracene

Concen: 20.251 ng/ul

RT: 26.621 min Scan# 4001

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18

Instrument :

BNA_N

ClientSampleId :

SSTD020212





Tgt Ion:278 Resp: 2525428

Ion Ratio Lower Upper

278 100

139 16.6 13.3 19.9

279 23.2 18.6 27.8



#96

Benzo(g,h,i)perylene

Concen: 19.104 ng/ul

RT: 27.392 min Scan# 4132

Delta R.T. 0.000 min

Lab File: BN023215.D

Acq: 15 Dec 2022 14:18

Tgt Ion:276 Resp: 2317649

Ion Ratio Lower Upper

276 100

138 24.0 19.2 28.8

277 24.1 19.3 28.9

