

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121522\
 Data File : BN023226.D
 Acq On : 15 Dec 2022 22:21
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020371

Quant Time: Dec 16 00:53:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 15 23:22:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.005	152	278271	20.000	ng/u1	0.00
20) Naphthalene-d8	10.822	136	1354502	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	907705	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.398	188	2006391	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	2007926	20.000	ng/u1	0.00
88) Perylene-d12	24.033	264	2195069	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.346	96	54359	8.082	ng/uL	0.00
4) Pyridine-d5	3.775	84	387221	19.503	ng/u1	0.00
7) Phenol-d5	7.158	99	511307	19.756	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	307057	20.388	ng/u1	0.00
11) 2-Chlorophenol-d4	7.528	132	405019	20.768	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	420931	20.257	ng/u1	0.00
21) Nitrobenzene-d5	9.175	128	211744	20.233	ng/u1	0.00
24) 2-Nitrophenol-d4	9.899	143	225934	20.498	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.440	165	422315	20.296	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	686975	20.518	ng/u1	0.00
46) Dimethylphthalate-d6	14.063	166	1352836	19.888	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	1542741	20.223	ng/u1	0.00
54) 4-Nitrophenol-d4	14.845	143	287381	18.916	ng/u1	0.00
60) Fluorene-d10	15.645	176	1154303	19.921	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	233741	19.790	ng/u1	0.00
73) Anthracene-d10	17.498	188	1838312	20.444	ng/u1	0.00
81) Pyrene-d10	19.786	212	2263916	20.727	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.874	264	2154241	20.022	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.381	88	53822	7.721	ng/uL	97
5) Pyridine	3.793	79	393221	19.383	ng/u1	99
6) Benzaldehyde	7.134	77	200194	21.595	ng/u1	98
8) Phenol	7.187	94	521009	19.929	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.422	93	419242	20.647	ng/u1	98
12) 2-Chlorophenol	7.558	128	412858	20.612	ng/u1	98
13) 2-Methylphenol	8.446	108	401353	20.178	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.540	45	529569	20.592	ng/u1	98
16) Acetophenone	8.834	105	668721	21.270	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.822	70	343163	21.797	ng/u1	99
18) 4-Methylphenol	8.781	108	453364	20.539	ng/u1	99
19) Hexachloroethane	9.093	117	160306	20.780	ng/u1	98
22) Nitrobenzene	9.216	77	497773	19.837	ng/u1	99
23) Isophorone	9.746	82	970626	20.507	ng/u1	98
25) 2-Nitrophenol	9.934	139	248823	20.698	ng/u1	99
26) 2,4-Dimethylphenol	9.993	107	493733	20.289	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.228	93	605702	20.005	ng/u1	98
29) 2,4-Dichlorophenol	10.469	162	412471	20.032	ng/u1	98
30) Naphthalene	10.875	128	1432586	19.797	ng/u1	99
32) 4-Chloroaniline	10.987	127	678587	20.314	ng/u1	98
33) Hexachlorobutadiene	11.157	225	241603	20.582	ng/u1	98
34) Caprolactam	11.757	113	92463	11.549	ng/u1	98
35) 4-Chloro-3-methylphenol	12.104	107	482633	20.694	ng/u1	95
36) 2-Methylnaphthalene	12.481	142	986482	20.227	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121522\
 Data File : BN023226.D
 Acq On : 15 Dec 2022 22:21
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020371

Quant Time: Dec 16 00:53:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 15 23:22:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.699	142	1020386	20.303	ng/ul	94
39) 1,2,4,5-Tetrachloroben...	12.846	216	508430	19.803	ng/ul	96
40) Hexachlorocyclopentadiene	12.828	237	288546	18.685	ng/ul	97
41) 2,4,6-Trichlorophenol	13.087	196	350519	19.849	ng/ul	98
42) 2,4,5-Trichlorophenol	13.157	196	392429	20.078	ng/ul	99
43) 1,1'-Biphenyl	13.487	154	1345940	19.623	ng/ul	98
44) 2-Chloronaphthalene	13.528	162	1063558	19.912	ng/ul	99
45) 2-Nitroaniline	13.734	65	313701	20.267	ng/ul	98
47) Dimethylphthalate	14.110	163	1354495	19.734	ng/ul	99
48) 2,6-Dinitrotoluene	14.228	165	279922	21.064	ng/ul	97
50) Acenaphthylene	14.375	152	1671328	19.923	ng/ul	98
51) 3-Nitroaniline	14.557	138	263479	19.472	ng/ul	97
52) Acenaphthene	14.716	153	1159574	19.883	ng/ul	97
53) 2,4-Dinitrophenol	14.757	184	149124	17.610	ng/ul	95
55) 4-Nitrophenol	14.857	109	213584	19.008	ng/ul	99
56) Dibenzofuran	15.051	168	1599747	19.701	ng/ul	99
57) 2,4-Dinitrotoluene	15.010	165	409546	20.878	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.275	232	335914	19.851	ng/ul	99
59) Diethylphthalate	15.469	149	1357767	19.885	ng/ul	98
61) Fluorene	15.698	166	1305754	19.858	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.692	204	628505	19.943	ng/ul	96
63) 4-Nitroaniline	15.722	138	253284	19.221	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.775	198	234539	20.341	ng/ul	100
67) N-Nitrosodiphenylamine	15.904	169	1144310	20.608	ng/ul	97
68) 4-Bromophenyl-phenylether	16.587	248	395789	20.444	ng/ul	97
69) Hexachlorobenzene	16.704	284	476027	21.031	ng/ul	96
70) Atrazine	16.857	200	416314	20.516	ng/ul	97
71) Pentachlorophenol	17.045	266	296503	19.838	ng/ul	96
72) Phenanthrene	17.439	178	2172942	20.467	ng/ul	100
74) Anthracene	17.534	178	2193169	20.704	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.451	216	505539	20.261	ng/ul	97
76) Pentachlorobenzene	14.969	250	529199	20.930	ng/ul	95
77) Carbazole	17.804	167	2029011	20.801	ng/ul	100
78) Di-n-butylphthalate	18.363	149	2396351	20.690	ng/ul	100
80) Fluoranthene	19.451	202	2629516	20.699	ng/ul	99
82) Pyrene	19.816	202	2729496	20.454	ng/ul	99
83) Butylbenzylphthalate	20.710	149	1157797	20.966	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.498	252	912781	20.686	ng/ul	96
85) Benzo(a)anthracene	21.563	228	2709114	20.138	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.486	149	1670965	21.400	ng/ul	100
87) Chrysene	21.622	228	2566965	20.219	ng/ul	99
89) Di-n-octyl phthalate	22.427	149	2876143	21.636	ng/ul	100
90) Benzo(b)fluoranthene	23.280	252	2730280	19.720	ng/ul	99
91) Benzo(k)fluoranthene	23.333	252	2778536	21.089	ng/ul	99
93) Benzo(a)pyrene	23.927	252	2385380	20.048	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.604	276	2917704	19.357	ng/ul	99
95) Dibenzo(a,h)anthracene	26.621	278	2409150	19.602	ng/ul	99
96) Benzo(g,h,i)perylene	27.392	276	2298430	19.272	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 7.721 ng/uL
 RT: 3.381 min Scan# 50
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020371


Tgt Ion: 88 Resp: 53822
 Ion Ratio Lower Upper
 88 100
 43 26.1 23.0 34.6
 58 67.5 55.8 83.6



#5
 Pyridine
 Concen: 19.383 ng/uL
 RT: 3.793 min Scan# 120
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21

Tgt Ion: 79 Resp: 393221
 Ion Ratio Lower Upper
 79 100
 52 70.0 56.7 85.1
 51 33.3 26.0 39.0









#6
Benzaldehyde
Concen: 21.595 ng/ul
RT: 7.134 min Scan# 688
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion: 77 Resp: 200194
Ion Ratio Lower Upper
77 100
105 95.4 78.0 117.0
106 92.5 73.2 109.8



#8
Phenol
Concen: 19.929 ng/ul
RT: 7.187 min Scan# 697
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion: 94 Resp: 521009
Ion Ratio Lower Upper
94 100
65 26.5 21.8 32.6
66 34.4 28.4 42.6









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.647 ng/ul
 RT: 7.422 min Scan# 71
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020371

Tgt Ion: 93 Resp: 419242




Ion	Ratio	Lower	Upper
93	100		
63	72.3	58.9	88.3
95	31.6	26.2	39.4




#12
 2-Chlorophenol
 Concen: 20.612 ng/ul
 RT: 7.558 min Scan# 760
 Delta R.T. -0.006 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21

Tgt Ion: 128 Resp: 412858

Ion	Ratio	Lower	Upper
128	100		
64	44.2	34.0	51.0
130	31.6	25.0	37.4





#13

2-Methylphenol

Concen: 20.178 ng/ul

RT: 8.446 min Scan# 911

Delta R.T. -0.006 min

Lab File: BN023226.D

Acq: 15 Dec 2022 22:21

Instrument :

BNA_N

ClientSampleId :





SSTD020371

Tgt Ion:108 Resp: 401353

Ion Ratio Lower Upper

108 100

107 88.5 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.592 ng/ul

RT: 8.540 min Scan# 927

Delta R.T. 0.006 min

Lab File: BN023226.D

Acq: 15 Dec 2022 22:21




Tgt Ion: 45 Resp: 529569


Ion Ratio Lower Upper

45 100

77 15.4 12.6 19.0

79 12.5 8.8 13.2









#16
 Acetophenone
 Concen: 21.270 ng/ul
 RT: 8.834 min Scan# 977
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21




Instrument :
 BNA_N
 ClientSampleId :
 SST020371


Tgt Ion:105 Resp: 668721
 Ion Ratio Lower Upper
 105 100
 77 75.5 61.4 92.0
 51 23.9 20.0 30.0



#17
 N-Nitroso-di-n-propylamine
 Concen: 21.797 ng/ul
 RT: 8.822 min Scan# 975
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21

Tgt Ion: 70 Resp: 343163
 Ion Ratio Lower Upper
 70 100
 42 56.1 45.3 67.9
 101 9.8 8.1 12.1
 130 22.7 17.9 26.9









#18
4-Methylphenol
Concen: 20.539 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:108 Resp: 453364
Ion Ratio Lower Upper
108 100
107 109.0 87.6 131.4



#19
Hexachloroethane
Concen: 20.780 ng/ul
RT: 9.093 min Scan# 1021
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:117 Resp: 160306
Ion Ratio Lower Upper
117 100
201 94.1 74.6 111.8
199 59.7 46.3 69.5









#22
Nitrobenzene
Concen: 19.837 ng/ul
RT: 9.216 min Scan# 1042
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion: 77 Resp: 497773
Ion Ratio Lower Upper
77 100
123 46.9 37.1 55.7
65 12.4 10.1 15.1



#23
Isophorone
Concen: 20.507 ng/ul
RT: 9.746 min Scan# 1132
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion: 82 Resp: 970626
Ion Ratio Lower Upper
82 100
95 7.0 5.4 8.2
138 19.1 14.5 21.7









#25
2-Nitrophenol
Concen: 20.698 ng/ul
RT: 9.934 min Scan# 1164
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:139 Resp: 248823
Ion Ratio Lower Upper
139 100
65 46.4 37.7 56.5
109 27.9 22.6 34.0



#26
2,4-Dimethylphenol
Concen: 20.289 ng/ul
RT: 9.993 min Scan# 1174
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:107 Resp: 493733
Ion Ratio Lower Upper
107 100
121 53.1 41.3 61.9
122 87.4 72.1 108.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.005 ng/ul
 RT: 10.228 min Scan# 1214
 Delta R.T. -0.006 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020371


Tgt Ion: 93 Resp: 605702
 Ion Ratio Lower Upper
 93 100
 95 31.1 25.8 38.6
 123 11.3 9.0 13.4



#29
 2,4-Dichlorophenol
 Concen: 20.032 ng/ul
 RT: 10.469 min Scan# 1255
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21

Tgt Ion: 162 Resp: 412471
 Ion Ratio Lower Upper
 162 100
 164 66.0 52.1 78.1
 98 35.0 29.3 43.9









#30
Naphthalene
Concen: 19.797 ng/ul
RT: 10.875 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:128 Resp: 1432586
Ion Ratio Lower Upper
128 100
129 11.2 8.7 13.1
127 13.2 10.3 15.5



#32
4-Chloroaniline
Concen: 20.314 ng/ul
RT: 10.987 min Scan# 1343
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:127 Resp: 678587
Ion Ratio Lower Upper
127 100
129 31.4 26.2 39.4





#33

Hexachlorobutadiene

Concen: 20.582 ng/ul

RT: 11.157 min Scan# 11

Delta R.T. -0.006 min

Lab File: BN023226.D

Acq: 15 Dec 2022 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020371





Tgt Ion:225 Resp: 241603

Ion Ratio Lower Upper

225 100

223 61.0 48.9 73.3

227 61.0 51.0 76.4



#34

Caprolactam

Concen: 11.549 ng/ul

RT: 11.757 min Scan# 1474

Delta R.T. 0.000 min

Lab File: BN023226.D

Acq: 15 Dec 2022 22:21




Tgt Ion:113 Resp: 92463


Ion Ratio Lower Upper

113 100

55 151.5 123.7 185.5

56 114.0 91.7 137.5









#35
4-Chloro-3-methylphenol
Concen: 20.694 ng/ul
RT: 12.104 min Scan# 1533
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:107 Resp: 482633
Ion Ratio Lower Upper
107 100
144 25.7 21.8 32.6
142 83.4 62.6 94.0



#36
2-Methylnaphthalene
Concen: 20.227 ng/ul
RT: 12.481 min Scan# 1597
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:142 Resp: 986482
Ion Ratio Lower Upper
142 100
141 88.0 69.9 104.9





#37

1-Methylnaphthalene

Concen: 20.303 ng/ul

RT: 12.699 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN023226.D

Acq: 15 Dec 2022 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020371





Tgt Ion:142 Resp: 1020386

Ion Ratio Lower Upper

142 100

141 94.5 70.9 106.3

116 3.7 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.803 ng/ul

RT: 12.846 min Scan# 1659

Delta R.T. 0.000 min

Lab File: BN023226.D

Acq: 15 Dec 2022 22:21

Tgt Ion:216 Resp: 508430




Ion Ratio Lower Upper


216 100

214 80.4 61.1 91.7

179 21.5 16.6 24.8

108 18.4 15.4 23.2









#40
Hexachlorocyclopentadiene
Concen: 18.685 ng/ul
RT: 12.828 min Scan# 1656
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:237 Resp: 288546
Ion Ratio Lower Upper
237 100
235 62.6 48.1 72.1
272 13.7 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 19.849 ng/ul
RT: 13.087 min Scan# 1700
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:196 Resp: 350519
Ion Ratio Lower Upper
196 100
198 98.2 76.9 115.3
200 31.3 23.8 35.6









#42
2,4,5-Trichlorophenol
Concen: 20.078 ng/ul
RT: 13.157 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:196 Resp: 392429
Ion Ratio Lower Upper
196 100
198 92.5 74.8 112.2
200 30.5 24.2 36.2



#43
1,1'-Biphenyl
Concen: 19.623 ng/ul
RT: 13.487 min Scan# 1768
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:154 Resp: 1345940
Ion Ratio Lower Upper
154 100
153 39.1 32.0 48.0
76 13.0 11.0 16.4









#44
2-Chloronaphthalene
Concen: 19.912 ng/ul
RT: 13.528 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:162 Resp: 1063558
Ion Ratio Lower Upper
162 100
164 32.0 25.4 38.0
127 39.3 31.8 47.8



#45
2-Nitroaniline
Concen: 20.267 ng/ul
RT: 13.734 min Scan# 1810
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion: 65 Resp: 313701
Ion Ratio Lower Upper
65 100
92 70.1 57.4 86.2
138 113.9 89.0 133.6









#47
Dimethylphthalate
Concen: 19.734 ng/ul
RT: 14.110 min Scan# 1874
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:163 Resp: 1354495
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.0 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 21.064 ng/ul
RT: 14.228 min Scan# 1894
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:165 Resp: 279922
Ion Ratio Lower Upper
165 100
89 57.6 47.7 71.5
121 21.2 17.8 26.8









#50
Acenaphthylene
Concen: 19.923 ng/ul
RT: 14.375 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:152 Resp: 1671328
Ion Ratio Lower Upper
152 100
151 19.5 16.2 24.4
153 13.5 11.4 17.0



#51
3-Nitroaniline
Concen: 19.472 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:138 Resp: 263479
Ion Ratio Lower Upper
138 100
108 10.6 8.3 12.5
92 111.6 92.5 138.7









#52
Acenaphthene
Concen: 19.883 ng/ul
RT: 14.716 min Scan# 1977
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:153 Resp: 1159574
Ion Ratio Lower Upper
153 100
152 46.8 38.9 58.3
154 90.4 70.5 105.7



#53
2,4-Dinitrophenol
Concen: 17.610 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:184 Resp: 149124
Ion Ratio Lower Upper
184 100
63 64.5 57.1 85.7
154 59.7 47.7 71.5









#55
4-Nitrophenol
Concen: 19.008 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:109 Resp: 213584
Ion Ratio Lower Upper
109 100
139 146.9 115.2 172.8
65 135.9 109.0 163.4



#56
Dibenzofuran
Concen: 19.701 ng/ul
RT: 15.051 min Scan# 2034
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:168 Resp: 1599747
Ion Ratio Lower Upper
168 100
139 37.2 29.2 43.8









#57
2,4-Dinitrotoluene
Concen: 20.878 ng/ul
RT: 15.010 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:165 Resp: 409546
Ion Ratio Lower Upper
165 100
63 42.0 36.6 54.8
182 2.8 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.851 ng/ul
RT: 15.275 min Scan# 2072
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:232 Resp: 335914
Ion Ratio Lower Upper
232 100
131 47.4 37.9 56.9
130 2.6 1.9 2.9
166 31.1 23.8 35.8









#59
Diethylphthalate
Concen: 19.885 ng/ul
RT: 15.469 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:149 Resp: 1357767
Ion Ratio Lower Upper
149 100
177 22.6 17.0 25.4
150 12.7 10.0 15.0



#61
Fluorene
Concen: 19.858 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:166 Resp: 1305754
Ion Ratio Lower Upper
166 100
165 94.8 77.8 116.8
167 13.4 10.6 15.8









#62
 4-Chlorophenyl-phenylether
 Concen: 19.943 ng/ul
 RT: 15.692 min Scan# 2143
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020371

Tgt Ion:204 Resp: 628505




Ion	Ratio	Lower	Upper
204	100		
206	31.5	26.1	39.1
141	56.4	48.2	72.2




#63
 4-Nitroaniline
 Concen: 19.221 ng/ul
 RT: 15.722 min Scan# 2148
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21

Tgt Ion:138 Resp: 253284

Ion	Ratio	Lower	Upper
138	100		
92	49.2	39.4	59.2
108	72.6	59.5	89.3









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.341 ng/ul
 RT: 15.775 min Scan# 2157
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020371


Tgt Ion:198 Resp: 234539
 Ion Ratio Lower Upper
 198 100
 182 4.0 3.4 5.0
 77 24.8 19.7 29.5



#67
 N-Nitrosodiphenylamine
 Concen: 20.608 ng/ul
 RT: 15.904 min Scan# 2179
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21

Tgt Ion:169 Resp: 1144310
 Ion Ratio Lower Upper
 169 100
 168 65.3 54.7 82.1
 167 36.1 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 20.444 ng/ul
RT: 16.587 min Scan# 2295
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:248 Resp: 395789
Ion Ratio Lower Upper
248 100
250 98.4 81.0 121.4
141 78.4 60.8 91.2



#69
Hexachlorobenzene
Concen: 21.031 ng/ul
RT: 16.704 min Scan# 2315
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:284 Resp: 476027
Ion Ratio Lower Upper
284 100
142 32.6 29.0 43.4
249 30.2 23.7 35.5









#70
 Atrazine
 Concen: 20.516 ng/ul
 RT: 16.857 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21




Instrument :
 BNA_N
 ClientSampleId :
 SST020371


Tgt Ion:200 Resp: 416314
 Ion Ratio Lower Upper
 200 100
 173 28.0 21.3 31.9
 215 51.6 39.8 59.8



#71
 Pentachlorophenol
 Concen: 19.838 ng/ul
 RT: 17.045 min Scan# 2373
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21

Tgt Ion:266 Resp: 296503
 Ion Ratio Lower Upper
 266 100
 264 65.3 50.5 75.7
 268 62.4 52.6 78.8









#72
 Phenanthrene
 Concen: 20.467 ng/u1
 RT: 17.439 min Scan# 2440
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020371

Tgt Ion:178 Resp: 2172942




Ion	Ratio	Lower	Upper
178	100		
179	15.7	12.3	18.5
176	19.2	15.4	23.0




#74
 Anthracene
 Concen: 20.704 ng/u1
 RT: 17.534 min Scan# 2456
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21

Tgt Ion:178 Resp: 2193169

Ion	Ratio	Lower	Upper
178	100		
179	15.4	11.8	17.6
176	18.3	14.7	22.1









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 20.261 ng/uL
 RT: 13.451 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21

Instrument :
 BNA_N
 ClientSampleId :
 SST020371

Tgt Ion:216 Resp: 505539




Ion	Ratio	Lower	Upper
216	100		
214	79.0	61.9	92.9
218	46.7	39.4	59.2




#76
 Pentachlorobenzene
 Concen: 20.930 ng/uL
 RT: 14.969 min Scan# 2020
 Delta R.T. 0.000 min
 Lab File: BN023226.D
 Acq: 15 Dec 2022 22:21

Tgt Ion:250 Resp: 529199

Ion	Ratio	Lower	Upper
250	100		
248	62.7	49.4	74.0
252	60.0	52.9	79.3





#77

Carbazole

Concen: 20.801 ng/ul

RT: 17.804 min Scan# 2502

Delta R.T. 0.000 min

Lab File: BN023226.D

Acq: 15 Dec 2022 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020371





Tgt Ion:167 Resp: 2029011

Ion Ratio Lower Upper

167 100

166 21.7 17.3 25.9

139 13.3 10.6 16.0



#78

Di-n-butylphthalate

Concen: 20.690 ng/ul

RT: 18.363 min Scan# 2597

Delta R.T. 0.000 min

Lab File: BN023226.D

Acq: 15 Dec 2022 22:21




Tgt Ion:149 Resp: 2396351


Ion Ratio Lower Upper

149 100

150 9.2 7.4 11.2

104 5.2 4.1 6.1









#80
Fluoranthene
Concen: 20.699 ng/u1
RT: 19.451 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:202 Resp: 2629516
Ion Ratio Lower Upper
202 100
101 12.3 10.4 15.6
100 9.5 7.6 11.4



#82
Pyrene
Concen: 20.454 ng/u1
RT: 19.816 min Scan# 2844
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:202 Resp: 2729496
Ion Ratio Lower Upper
202 100
101 14.7 12.0 18.0
100 12.0 9.5 14.3





#83

Butylbenzylphthalate

Concen: 20.966 ng/ul

RT: 20.710 min Scan# 2996

Delta R.T. 0.000 min

Lab File: BN023226.D

Acq: 15 Dec 2022 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020371





Tgt Ion:149 Resp: 1157797

Ion Ratio Lower Upper

149 100

91 71.0 57.9 86.9

206 21.0 15.6 23.4



#84

3,3'-Dichlorobenzidine

Concen: 20.686 ng/ul

RT: 21.498 min Scan# 3130

Delta R.T. 0.000 min

Lab File: BN023226.D

Acq: 15 Dec 2022 22:21




Tgt Ion:252 Resp: 912781


Ion Ratio Lower Upper

252 100

254 66.9 50.3 75.5

126 11.9 9.4 14.0









#85
Benzo(a)anthracene
Concen: 20.138 ng/ul
RT: 21.563 min Scan# 3141
Delta R.T. -0.006 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:228 Resp: 2709114
Ion Ratio Lower Upper
228 100
229 19.2 15.5 23.3
226 25.3 21.4 32.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.400 ng/ul
RT: 21.486 min Scan# 3128
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:149 Resp: 1670965
Ion Ratio Lower Upper
149 100
167 27.7 21.9 32.9
279 4.2 3.4 5.0





#87

Chrysene

Concen: 20.219 ng/ul

RT: 21.622 min Scan# 3151

Delta R.T. 0.000 min

Lab File: BN023226.D

Acq: 15 Dec 2022 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020371





Tgt Ion:228 Resp: 2566965

Ion Ratio Lower Upper

228 100

226 30.6 24.6 36.8

229 19.5 16.5 24.7



#89

Di-n-octyl phthalate

Concen: 21.636 ng/ul




RT: 22.427 min Scan# 3288


Delta R.T. 0.000 min

Lab File: BN023226.D

Acq: 15 Dec 2022 22:21

Tgt Ion:149 Resp: 2876143









#90
Benzo(b)fluoranthene
Concen: 19.720 ng/ul
RT: 23.280 min Scan# 3433
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:252 Resp: 2730280
Ion Ratio Lower Upper
252 100
253 22.0 16.8 25.2
125 10.9 8.8 13.2



#91
Benzo(k)fluoranthene
Concen: 21.089 ng/ul
RT: 23.333 min Scan# 3442
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:252 Resp: 2778536
Ion Ratio Lower Upper
252 100
253 22.0 17.9 26.9
125 11.3 9.0 13.6









#93
Benzo(a)pyrene
Concen: 20.048 ng/ul
RT: 23.927 min Scan# 31
Delta R.T. 0.000 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21




Instrument :
BNA_N
ClientSampleId :
SSTD020371


Tgt Ion:252 Resp: 2385380
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 11.8 9.4 14.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 19.357 ng/ul
RT: 26.604 min Scan# 3998
Delta R.T. 0.006 min
Lab File: BN023226.D
Acq: 15 Dec 2022 22:21

Tgt Ion:276 Resp: 2917704
Ion Ratio Lower Upper
276 100
138 25.2 20.6 30.8
277 23.7 18.2 27.2





#95

Dibenzo(a,h)anthracene

Concen: 19.602 ng/ul

RT: 26.621 min Scan# 4001

Delta R.T. 0.000 min

Lab File: BN023226.D

Acq: 15 Dec 2022 22:21

Instrument :

BNA_N

ClientSampleId :

SSTD020371





Tgt Ion:278 Resp: 2409150

Ion Ratio Lower Upper

278 100

139 17.2 13.3 19.9

279 23.5 18.6 27.8



#96

Benzo(g,h,i)perylene

Concen: 19.272 ng/ul

RT: 27.392 min Scan# 4132

Delta R.T. 0.000 min

Lab File: BN023226.D

Acq: 15 Dec 2022 22:21

Tgt Ion:276 Resp: 2298430

Ion Ratio Lower Upper

276 100

138 23.1 19.2 28.8

277 22.2 19.3 28.9

