

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121522\
 Data File : BN023229.D
 Acq On : 16 Dec 2022 00:12
 Operator : CG/JU
 Sample : N5951-10
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SVOC-GPC2-BLANK

Quant Time: Dec 16 00:55:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 15 23:22:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.005	152	228990	20.000	ng/ul	0.00
20) Naphthalene-d8	10.822	136	1021954	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.651	164	695666	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.398	188	1610115	20.000	ng/ul	0.00
79) Chrysene-d12	21.580	240	1626158	20.000	ng/ul	0.00
88) Perylene-d12	24.027	264	1722159	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	278	0.050	ng/uL	0.01
4) Pyridine-d5	3.781	84	316	0.019	ng/ul	0.00
7) Phenol-d5	0.000	99	0	0.000	ng/ul	
9) Bis-(2-Chloroethyl)eth...	0.000	67	0	0.000	ng/ul	
11) 2-Chlorophenol-d4	0.000	132	0	0.000	ng/ul	
15) 4-Methylphenol-d8	0.000	113	0	0.000	ng/ul	
21) Nitrobenzene-d5	0.000	128	0	0.000	ng/ul	
24) 2-Nitrophenol-d4	0.000	143	0	0.000	ng/ul	
28) 2,4-Dichlorophenol-d3	0.000	165	0	0.000	ng/ul	
31) 4-Chloroaniline-d4	10.828	131	1384	0.055	ng/ul	-0.14
46) Dimethylphthalate-d6	14.057	166	136	0.003	ng/ul	0.00
49) Acenaphthylene-d8	14.351	160	63	0.001	ng/ul	0.00
54) 4-Nitrophenol-d4	0.000	143	0	0.000	ng/ul	
60) Fluorene-d10	15.639	176	767	0.017	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	0.000	200	0	0.000	ng/ul	
73) Anthracene-d10	17.398	188	1610115	22.313	ng/ul	-0.10
81) Pyrene-d10	19.786	212	704	0.008	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.027	264	1722159	20.402	ng/ul	0.15

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121522\
Data File : BN023229.D
Acq On : 16 Dec 2022 00:12
Operator : CG/JU
Sample : N5951-10
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SVOC-GPC2-BLANK

Quant Time: Dec 16 00:55:06 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 15 23:22:36 2022
Response via : Initial Calibration

