

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN121525\  
 Data File : BN038428.D  
 Acq On : 16 Dec 2025 06:31  
 Operator : RC/JU  
 Sample : Q3839-26  
 Misc :  
 ALS Vial : 31 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 C0AT8

Quant Time: Dec 16 07:16:42 2025  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN121025.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Dec 10 13:56:50 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc  | Units | Dev(Min)       |
|-------------------------------|--------|------|----------|-------|-------|----------------|
| -----                         |        |      |          |       |       |                |
| Internal Standards            |        |      |          |       |       |                |
| 1) 1,4-Dichlorobenzene-d4     | 7.710  | 152  | 4631     | 0.400 | ng    | -0.01          |
| 7) Naphthalene-d8             | 10.509 | 136  | 12375    | 0.400 | ng    | #-0.02         |
| 13) Acenaphthene-d10          | 14.387 | 164  | 6551     | 0.400 | ng    | -0.01          |
| 19) Phenanthrene-d10          | 17.136 | 188  | 12461    | 0.400 | ng    | #-0.01         |
| 29) Chrysene-d12              | 21.331 | 240  | 9867     | 0.400 | ng    | 0.00           |
| 35) Perylene-d12              | 23.619 | 264  | 10049    | 0.400 | ng    | -0.01          |
| System Monitoring Compounds   |        |      |          |       |       |                |
| 4) 2-Fluorophenol             | 5.283  | 112  | 1515     | 0.131 | ng    | 0.00           |
| 5) Phenol-d6                  | 6.879  | 99   | 1019     | 0.074 | ng    | 0.00           |
| 8) Nitrobenzene-d5            | 8.864  | 82   | 3481     | 0.333 | ng    | -0.01          |
| 11) 2-Methylnaphthalene-d10   | 12.116 | 152  | 5443     | 0.312 | ng    | -0.01          |
| 14) 2,4,6-Tribromophenol      | 15.882 | 330  | 987      | 0.340 | ng    | -0.01          |
| 15) 2-Fluorobiphenyl          | 13.003 | 172  | 13140    | 0.416 | ng    | 0.00           |
| 27) Fluoranthene-d10          | 19.173 | 212  | 11441    | 0.371 | ng    | -0.01          |
| 31) Terphenyl-d14             | 19.777 | 244  | 10028    | 0.455 | ng    | -0.01          |
| Target Compounds              |        |      |          |       |       |                |
| 34) Bis(2-ethylhexyl)phtha... | 21.259 | 149  | 1240     | 0.059 | ng    | Qvalue<br># 98 |
| -----                         |        |      |          |       |       |                |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN121525\  
Data File : BN038428.D  
Acq On : 16 Dec 2025 06:31  
Operator : RC/JU  
Sample : Q3839-26  
Misc :  
ALS Vial : 31 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
COAT8

Quant Time: Dec 16 07:16:42 2025  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN121025.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Dec 10 13:56:50 2025  
Response via : Initial Calibration

