

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121525\
 Data File : BN038398.D
 Acq On : 15 Dec 2025 10:54
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 15 11:53:05 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	84	0.00
2	1,4-Dioxane	0.400	0.390	2.5	80	0.00
3	n-Nitrosodimethylamine	0.400	0.351	12.3	74	0.00
4 S	2-Fluorophenol	0.400	0.396	1.0	83	0.00
5 S	Phenol-d6	0.400	0.391	2.3	84	0.00
6	bis(2-Chloroethyl)ether	0.400	0.377	5.8	78	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	88	-0.01
8 S	Nitrobenzene-d5	0.400	0.376	6.0	85	-0.01
9	Naphthalene	0.400	0.378	5.5	84	-0.01
10	Hexachlorobutadiene	0.400	0.379	5.3	83	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.388	3.0	88	-0.01
12	2-Methylnaphthalene	0.400	0.383	4.3	86	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	95	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.405	-1.3	102	-0.01
15 S	2-Fluorobiphenyl	0.400	0.451	-12.7	104	0.00
16	Acenaphthylene	0.400	0.375	6.3	92	-0.01
17	Acenaphthene	0.400	0.361	9.8	87	-0.01
18	Fluorene	0.400	0.375	6.3	92	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	99	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.212	47.0#	61	0.00
21	4-Bromophenyl-phenylether	0.400	0.363	9.3	93	-0.01
22	Hexachlorobenzene	0.400	0.365	8.8	92	-0.01
23	Atrazine	0.400	0.353	11.8	93	-0.01
24	Pentachlorophenol	0.400	0.351	12.3	100	0.00
25	Phenanthrene	0.400	0.365	8.8	92	-0.01
26	Anthracene	0.400	0.367	8.3	96	-0.01
27 SURR	Fluoranthene-d10	0.400	0.383	4.3	95	0.00
28	Fluoranthene	0.400	0.387	3.3	97	-0.01
29 I	Chrysene-d12	0.400	0.400	0.0	102	0.00
30	Pyrene	0.400	0.370	7.5	97	-0.01
31 S	Terphenyl-d14	0.400	0.378	5.5	98	-0.01
32	Benzo(a)anthracene	0.400	0.391	2.3	104	0.00
33	Chrysene	0.400	0.385	3.8	97	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.372	7.0	95	0.00
35 I	Perylene-d12	0.400	0.400	0.0	105	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.376	6.0	98	-0.03
37	Benzo(b)fluoranthene	0.400	0.374	6.5	101	-0.02
38	Benzo(k)fluoranthene	0.400	0.375	6.3	100	-0.02
39 C	Benzo(a)pyrene	0.400	0.372	7.0	100	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.378	5.5	100	-0.03
41	Benzo(g,h,i)perylene	0.400	0.383	4.3	101	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4