

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121618\
 Data File : BN004026.D
 Acq On : 17 Dec 2018 09:32
 Operator : JU/SJ
 Sample : J6229-05
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F9F03

Manual Integrations
 APPROVED

Sohil
 12/17/2018 4:28:56 PM

Quant Time: Dec 17 13:58:26 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 17 06:22:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	1789	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	9277	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	5705	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.22	188	14793m	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	16564m	0.40	ng/ul	0.00
20) Perylene-d12	23.72	264	14680m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	2868	0.20	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	9527m	0.22	ng/ul	0.00
Target Compounds				Ovalue		
5) 2-Methylnaphthalene	12.29	142	456	0.025	ng/ul	99
7) Acenaphthylene	14.19	152	10323	0.402	ng/ul	99
8) Acenaphthene	14.53	153	522	0.023	ng/ul#	93
9) Fluorene	15.52	166	1309	0.050	ng/ul#	91
12) Phenanthrene	17.26	178	48763m	1.004	ng/ul	
13) Anthracene	17.35	178	12765m	0.320	ng/ul	
15) Fluoranthene	19.27	202	189093m	3.298	ng/ul	
17) Pyrene	19.64	202	180242	3.034	ng/ul	97
18) Benzo(a)anthracene	21.38	228	108735m	2.078	ng/ul	
19) Chrysene	21.44	228	122287m	2.043	ng/ul	
21) Benzo(b)fluoranthene	23.02	252	138138m	2.275	ng/ul	
22) Benzo(k)fluoranthene	23.06	252	48796m	0.825	ng/ul	
23) Benzo(a)pyrene	23.62	252	76502m	1.455	ng/ul	
24) Indeno(1,2,3-cd)pyrene	26.08	276	53258m	0.883	ng/ul	
25) Dibenzo(a,h)anthracene	26.08	278	16294m	0.336	ng/ul	
26) Benzo(a,h,i)perylene	26.81	276	47030m	0.897	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121618\
Data File : BN004026.D
Acq On : 17 Dec 2018 09:32
Operator : JU/SJ
Sample : J6229-05
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9F03

Quant Time: Dec 17 13:58:26 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN11918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 17 06:22:23 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
12/17/2018 4:28:56 PM

