

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121622\
 Data File : BN023243.D
 Acq On : 16 Dec 2022 17:03
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020373

Quant Time: Dec 16 21:44:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 15 23:22:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.005	152	253926	20.000	ng/u1	0.00
20) Naphthalene-d8	10.822	136	1212931	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	815776	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.398	188	1878403	20.000	ng/u1	0.00
79) Chrysene-d12	21.586	240	1800196	20.000	ng/u1	0.00
88) Perylene-d12	24.039	264	1861313	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.346	96	46447	7.568	ng/uL	0.00
4) Pyridine-d5	3.775	84	350311	19.336	ng/u1	0.00
7) Phenol-d5	7.158	99	452726	19.170	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	273061	19.869	ng/u1	0.00
11) 2-Chlorophenol-d4	7.528	132	366689	20.605	ng/u1	0.00
15) 4-Methylphenol-d8	8.716	113	380380	20.060	ng/u1	0.00
21) Nitrobenzene-d5	9.175	128	192082	20.497	ng/u1	0.00
24) 2-Nitrophenol-d4	9.899	143	206030	20.873	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.440	165	389139	20.884	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	617614	20.599	ng/u1	0.00
46) Dimethylphthalate-d6	14.063	166	1236200	20.221	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	1425397	20.791	ng/u1	0.00
54) 4-Nitrophenol-d4	14.845	143	262174	19.201	ng/u1	0.00
60) Fluorene-d10	15.645	176	1062397	20.401	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	210495	19.037	ng/u1	0.00
73) Anthracene-d10	17.498	188	1716393	20.389	ng/u1	0.00
81) Pyrene-d10	19.792	212	2003091	20.455	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.880	264	1904698	20.877	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.381	88	49571	7.793	ng/uL	94
5) Pyridine	3.793	79	358477	19.364	ng/u1	99
6) Benzaldehyde	7.134	77	182022	21.517	ng/u1	97
8) Phenol	7.187	94	467032	19.577	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.422	93	370595	20.001	ng/u1	99
12) 2-Chlorophenol	7.558	128	378442	20.706	ng/u1	98
13) 2-Methylphenol	8.446	108	361894	19.938	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.534	45	471309	20.084	ng/u1	100
16) Acetophenone	8.834	105	601593	20.969	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.816	70	305394	21.258	ng/u1	98
18) 4-Methylphenol	8.781	108	409189	20.315	ng/u1	99
19) Hexachloroethane	9.093	117	147389	20.938	ng/u1	92
22) Nitrobenzene	9.216	77	452295	20.128	ng/u1	99
23) Isophorone	9.746	82	876693	20.684	ng/u1	99
25) 2-Nitrophenol	9.928	139	226569	21.046	ng/u1	96
26) 2,4-Dimethylphenol	9.993	107	447605	20.540	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.228	93	535030	19.733	ng/u1	98
29) 2,4-Dichlorophenol	10.463	162	378796	20.544	ng/u1	98
30) Naphthalene	10.875	128	1304240	20.127	ng/u1	99
32) 4-Chloroaniline	10.987	127	611109	20.430	ng/u1	99
33) Hexachlorobutadiene	11.157	225	222698	21.186	ng/u1	98
34) Caprolactam	11.757	113	82757	11.543	ng/u1	96
35) 4-Chloro-3-methylphenol	12.104	107	437161	20.932	ng/u1	94
36) 2-Methylnaphthalene	12.481	142	894001	20.470	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121622\
 Data File : BN023243.D
 Acq On : 16 Dec 2022 17:03
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020373

Quant Time: Dec 16 21:44:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 15 23:22:36 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.698	142	920734	20.458	ng/ul	94
39) 1,2,4,5-Tetrachloroben...	12.845	216	459038	19.894	ng/ul	99
40) Hexachlorocyclopentadiene	12.828	237	274821	19.801	ng/ul	96
41) 2,4,6-Trichlorophenol	13.087	196	318842	20.090	ng/ul	92
42) 2,4,5-Trichlorophenol	13.157	196	362834	20.656	ng/ul	98
43) 1,1'-Biphenyl	13.487	154	1229676	19.949	ng/ul	99
44) 2-Chloronaphthalene	13.528	162	957098	19.938	ng/ul	99
45) 2-Nitroaniline	13.734	65	285828	20.547	ng/ul	98
47) Dimethylphthalate	14.110	163	1224830	19.855	ng/ul	99
48) 2,6-Dinitrotoluene	14.228	165	255745	21.413	ng/ul	97
50) Acenaphthylene	14.375	152	1537060	20.387	ng/ul	99
51) 3-Nitroaniline	14.557	138	243283	20.005	ng/ul	96
52) Acenaphthene	14.716	153	1051363	20.059	ng/ul	99
53) 2,4-Dinitrophenol	14.757	184	135689	17.829	ng/ul	94
55) 4-Nitrophenol	14.857	109	197613	19.569	ng/ul	94
56) Dibenzofuran	15.051	168	1472092	20.172	ng/ul	100
57) 2,4-Dinitrotoluene	15.010	165	377288	21.401	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.275	232	319555	21.013	ng/ul	97
59) Diethylphthalate	15.469	149	1244754	20.284	ng/ul	98
61) Fluorene	15.698	166	1213872	20.541	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.692	204	581620	20.535	ng/ul	98
63) 4-Nitroaniline	15.722	138	238474	20.136	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.775	198	210962	19.543	ng/ul	99
67) N-Nitrosodiphenylamine	15.904	169	1043577	20.074	ng/ul	97
68) 4-Bromophenyl-phenylether	16.586	248	368707	20.343	ng/ul	98
69) Hexachlorobenzene	16.704	284	436362	20.592	ng/ul	97
70) Atrazine	16.857	200	396143	20.852	ng/ul	98
71) Pentachlorophenol	17.045	266	283774	20.280	ng/ul	95
72) Phenanthrene	17.439	178	1973512	19.855	ng/ul	99
74) Anthracene	17.533	178	2004549	20.212	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.451	216	463730	19.852	ng/ul	100
76) Pentachlorobenzene	14.969	250	475777	20.099	ng/ul	97
77) Carbazole	17.804	167	1880490	20.592	ng/ul	100
78) Di-n-butylphthalate	18.363	149	2241536	20.672	ng/ul	99
80) Fluoranthene	19.457	202	2373119	20.837	ng/ul	98
82) Pyrene	19.822	202	2506049	20.947	ng/ul	97
83) Butylbenzylphthalate	20.716	149	1086173	21.939	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.504	252	780514	19.730	ng/ul	96
85) Benzo(a)anthracene	21.569	228	2458066	20.380	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.492	149	1553707	22.194	ng/ul	97
87) Chrysene	21.627	228	2252390	19.789	ng/ul	99
89) Di-n-octyl phthalate	22.433	149	2619070	23.236	ng/ul	100
90) Benzo(b)fluoranthene	23.286	252	2489539	21.206	ng/ul	99
91) Benzo(k)fluoranthene	23.339	252	2337183	20.920	ng/ul	99
93) Benzo(a)pyrene	23.933	252	2082691	20.643	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.609	276	2448695	19.158	ng/ul	97
95) Dibenzo(a,h)anthracene	26.627	278	2009812	19.285	ng/ul	98
96) Benzo(g,h,i)perylene	27.404	276	1887144	18.661	ng/ul	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121622\
 Data File : BN023243.D
 Acq On : 16 Dec 2022 17:03
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020373

Quant Time: Dec 16 21:44:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 15 23:22:36 2022
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 7.793 ng/uL
 RT: 3.381 min Scan# 50
 Delta R.T. 0.000 min
 Lab File: BN023243.D
 Acq: 16 Dec 2022 17:03




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020373


Tgt Ion: 88 Resp: 49571
 Ion Ratio Lower Upper
 88 100
 43 26.9 23.0 34.6
 58 64.0 55.8 83.6



#5
 Pyridine
 Concen: 19.364 ng/uL
 RT: 3.793 min Scan# 120
 Delta R.T. 0.000 min
 Lab File: BN023243.D
 Acq: 16 Dec 2022 17:03

Tgt Ion: 79 Resp: 358477
 Ion Ratio Lower Upper
 79 100
 52 71.4 56.7 85.1
 51 33.2 26.0 39.0









#6
Benzaldehyde
Concen: 21.517 ng/ul
RT: 7.134 min Scan# 688
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion: 77 Resp: 182022
Ion Ratio Lower Upper
77 100
105 96.2 78.0 117.0
106 95.6 73.2 109.8



#8
Phenol
Concen: 19.577 ng/ul
RT: 7.187 min Scan# 697
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion: 94 Resp: 467032
Ion Ratio Lower Upper
94 100
65 26.4 21.8 32.6
66 35.1 28.4 42.6









#10
Bis(2-Chloroethyl)ether
Concen: 20.001 ng/ul
RT: 7.422 min Scan# 737
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion: 93 Resp: 370595
Ion Ratio Lower Upper
93 100
63 74.3 58.9 88.3
95 32.5 26.2 39.4



#12
2-Chlorophenol
Concen: 20.706 ng/ul
RT: 7.558 min Scan# 760
Delta R.T. -0.006 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion: 128 Resp: 378442
Ion Ratio Lower Upper
128 100
64 44.0 34.0 51.0
130 32.3 25.0 37.4





#13

2-Methylphenol

Concen: 19.938 ng/ul

RT: 8.446 min Scan# 911

Delta R.T. -0.006 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03

Instrument :

BNA_N

ClientSampleId :





SSTD020373

Tgt Ion:108 Resp: 361894

Ion Ratio Lower Upper

108 100

107 87.5 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.084 ng/ul

RT: 8.534 min Scan# 926

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03




Tgt Ion: 45 Resp: 471309


Ion Ratio Lower Upper

45 100

77 15.9 12.6 19.0

79 11.3 8.8 13.2









#16
Acetophenone
Concen: 20.969 ng/ul
RT: 8.834 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:105 Resp: 601593
Ion Ratio Lower Upper
105 100
77 75.4 61.4 92.0
51 24.9 20.0 30.0



#17
N-Nitroso-di-n-propylamine
Concen: 21.258 ng/ul
RT: 8.816 min Scan# 974
Delta R.T. -0.006 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion: 70 Resp: 305394
Ion Ratio Lower Upper
70 100
42 55.0 45.3 67.9
101 9.4 8.1 12.1
130 21.5 17.9 26.9





#18

4-Methylphenol

Concen: 20.315 ng/ul

RT: 8.781 min Scan# 968

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03

Instrument :

BNA_N

ClientSampleId :





SSTD020373

Tgt Ion:108 Resp: 409189

Ion Ratio Lower Upper

108 100

107 108.5 87.6 131.4



#19

Hexachloroethane

Concen: 20.938 ng/ul

RT: 9.093 min Scan# 1021

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03




Tgt Ion:117 Resp: 147389


Ion Ratio Lower Upper

117 100

201 101.6 74.6 111.8

199 62.9 46.3 69.5









#22
Nitrobenzene
Concen: 20.128 ng/ul
RT: 9.216 min Scan# 1042
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion: 77 Resp: 452295
Ion Ratio Lower Upper
77 100
123 47.3 37.1 55.7
65 13.4 10.1 15.1



#23
Isophorone
Concen: 20.684 ng/ul
RT: 9.746 min Scan# 1132
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion: 82 Resp: 876693
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 18.5 14.5 21.7





#25

2-Nitrophenol

Concen: 21.046 ng/ul

RT: 9.928 min Scan# 1163

Delta R.T. -0.006 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:139 Resp: 226569

Ion Ratio Lower Upper

139 100

65 49.9 37.7 56.5

109 29.5 22.6 34.0



#26

2,4-Dimethylphenol

Concen: 20.540 ng/ul

RT: 9.993 min Scan# 1174

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03




Tgt Ion:107 Resp: 447605


Ion Ratio Lower Upper

107 100

121 51.1 41.3 61.9

122 87.2 72.1 108.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.733 ng/ul
 RT: 10.228 min Scan# 1214
 Delta R.T. -0.006 min
 Lab File: BN023243.D
 Acq: 16 Dec 2022 17:03




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020373


Tgt Ion: 93 Resp: 535030
 Ion Ratio Lower Upper
 93 100
 95 33.1 25.8 38.6
 123 11.8 9.0 13.4



#29
 2,4-Dichlorophenol
 Concen: 20.544 ng/ul
 RT: 10.463 min Scan# 1254
 Delta R.T. -0.006 min
 Lab File: BN023243.D
 Acq: 16 Dec 2022 17:03

Tgt Ion: 162 Resp: 378796
 Ion Ratio Lower Upper
 162 100
 164 67.4 52.1 78.1
 98 36.5 29.3 43.9









#30
Naphthalene
Concen: 20.127 ng/ul
RT: 10.875 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:128 Resp: 1304240
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.3 10.3 15.5



#32
4-Chloroaniline
Concen: 20.430 ng/ul
RT: 10.987 min Scan# 1343
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:127 Resp: 611109
Ion Ratio Lower Upper
127 100
129 32.0 26.2 39.4









#33
Hexachlorobutadiene
Concen: 21.186 ng/ul
RT: 11.157 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:225 Resp: 222698
Ion Ratio Lower Upper
225 100
223 64.5 48.9 73.3
227 63.9 51.0 76.4



#34
Caprolactam
Concen: 11.543 ng/ul
RT: 11.757 min Scan# 1474
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:113 Resp: 82757
Ion Ratio Lower Upper
113 100
55 148.3 123.7 185.5
56 111.8 91.7 137.5









#35
4-Chloro-3-methylphenol
Concen: 20.932 ng/ul
RT: 12.104 min Scan# 1533
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:107 Resp: 437161
Ion Ratio Lower Upper
107 100
144 27.8 21.8 32.6
142 85.0 62.6 94.0



#36
2-Methylnaphthalene
Concen: 20.470 ng/ul
RT: 12.481 min Scan# 1597
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:142 Resp: 894001
Ion Ratio Lower Upper
142 100
141 90.3 69.9 104.9





#37

1-Methylnaphthalene

Concen: 20.458 ng/ul

RT: 12.698 min Scan# 1634

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:142 Resp: 920734

Ion Ratio Lower Upper

142 100

141 94.3 70.9 106.3

116 3.8 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.894 ng/ul

RT: 12.845 min Scan# 1659

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03

Tgt Ion:216 Resp: 459038




Ion Ratio Lower Upper


216 100

214 76.9 61.1 91.7

179 20.3 16.6 24.8

108 17.9 15.4 23.2





#40

Hexachlorocyclopentadiene

Concen: 19.801 ng/ul

RT: 12.828 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:237 Resp: 274821

Ion Ratio Lower Upper

237 100

235 57.1 48.1 72.1

272 11.0 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 20.090 ng/ul

RT: 13.087 min Scan# 1700

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03




Tgt Ion:196 Resp: 318842


Ion Ratio Lower Upper

196 100

198 104.7 76.9 115.3

200 33.3 23.8 35.6





#42

2,4,5-Trichlorophenol

Concen: 20.656 ng/ul

RT: 13.157 min Scan# 1711

Delta R.T. 0.000 min

Lab File: BN023243.D


Acq: 16 Dec 2022 17:03

Instrument :

BNA_N

ClientSampleId :

SSTD020373






Tgt Ion:196 Resp: 362834

Ion Ratio Lower Upper

196 100

198 95.6 74.8 112.2

200 31.0 24.2 36.2



#43

1,1'-Biphenyl

Concen: 19.949 ng/ul

RT: 13.487 min Scan# 1768

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03




Tgt Ion:154 Resp: 1229676


Ion Ratio Lower Upper

154 100

153 40.2 32.0 48.0

76 13.1 11.0 16.4









#44
2-Chloronaphthalene
Concen: 19.938 ng/ul
RT: 13.528 min Scan# 1775
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:162 Resp: 957098
Ion Ratio Lower Upper
162 100
164 31.4 25.4 38.0
127 40.2 31.8 47.8



#45
2-Nitroaniline
Concen: 20.547 ng/ul
RT: 13.734 min Scan# 1810
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion: 65 Resp: 285828
Ion Ratio Lower Upper
65 100
92 70.9 57.4 86.2
138 113.6 89.0 133.6









#47
Dimethylphthalate
Concen: 19.855 ng/ul
RT: 14.110 min Scan# 1874
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:163 Resp: 1224830
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.0 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 21.413 ng/ul
RT: 14.228 min Scan# 1894
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:165 Resp: 255745
Ion Ratio Lower Upper
165 100
89 56.7 47.7 71.5
121 21.3 17.8 26.8









#50
Acenaphthylene
Concen: 20.387 ng/ul
RT: 14.375 min Scan# 1919
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:152 Resp: 1537060
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.4
153 13.1 11.4 17.0



#51
3-Nitroaniline
Concen: 20.005 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:138 Resp: 243283
Ion Ratio Lower Upper
138 100
108 10.0 8.3 12.5
92 110.9 92.5 138.7









#52
Acenaphthene
Concen: 20.059 ng/ul
RT: 14.716 min Scan# 1977
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:153 Resp: 1051363
Ion Ratio Lower Upper
153 100
152 46.1 38.9 58.3
154 88.3 70.5 105.7



#53
2,4-Dinitrophenol
Concen: 17.829 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:184 Resp: 135689
Ion Ratio Lower Upper
184 100
63 66.3 57.1 85.7
154 64.5 47.7 71.5









#55
4-Nitrophenol
Concen: 19.569 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:109 Resp: 197613
Ion Ratio Lower Upper
109 100
139 137.8 115.2 172.8
65 128.9 109.0 163.4



#56
Dibenzofuran
Concen: 20.172 ng/ul
RT: 15.051 min Scan# 2034
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:168 Resp: 1472092
Ion Ratio Lower Upper
168 100
139 36.3 29.2 43.8





#57

2,4-Dinitrotoluene

Concen: 21.401 ng/ul

RT: 15.010 min Scan# 2027

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:165 Resp: 377288

Ion Ratio Lower Upper

165 100

63 41.3 36.6 54.8

182 2.6 2.5 3.7



#58

2,3,4,6-Tetrachlorophenol

Concen: 21.013 ng/ul

RT: 15.275 min Scan# 2072

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03

Tgt Ion:232 Resp: 319555




Ion Ratio Lower Upper


232 100

131 44.7 37.9 56.9

130 2.3 1.9 2.9

166 30.4 23.8 35.8









#59
Diethylphthalate
Concen: 20.284 ng/ul
RT: 15.469 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:149 Resp: 1244754
Ion Ratio Lower Upper
149 100
177 22.4 17.0 25.4
150 12.8 10.0 15.0



#61
Fluorene
Concen: 20.541 ng/ul
RT: 15.698 min Scan# 2144
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:166 Resp: 1213872
Ion Ratio Lower Upper
166 100
165 94.5 77.8 116.8
167 13.2 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 20.535 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:204 Resp: 581620
Ion Ratio Lower Upper
204 100
206 31.6 26.1 39.1
141 58.5 48.2 72.2



#63
4-Nitroaniline
Concen: 20.136 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:138 Resp: 238474
Ion Ratio Lower Upper
138 100
92 46.8 39.4 59.2
108 74.7 59.5 89.3









#66
4,6-Dinitro-2-methylphenol
Concen: 19.543 ng/ul
RT: 15.775 min Scan# 2157
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:198 Resp: 210962
Ion Ratio Lower Upper
198 100
182 4.1 3.4 5.0
77 25.2 19.7 29.5



#67
N-Nitrosodiphenylamine
Concen: 20.074 ng/ul
RT: 15.904 min Scan# 2179
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:169 Resp: 1043577
Ion Ratio Lower Upper
169 100
168 65.5 54.7 82.1
167 35.8 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 20.343 ng/ul
RT: 16.586 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:248 Resp: 368707
Ion Ratio Lower Upper
248 100
250 101.1 81.0 121.4
141 72.9 60.8 91.2



#69
Hexachlorobenzene
Concen: 20.592 ng/ul
RT: 16.704 min Scan# 2315
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:284 Resp: 436362
Ion Ratio Lower Upper
284 100
142 33.6 29.0 43.4
249 30.1 23.7 35.5









#70
Atrazine
Concen: 20.852 ng/ul
RT: 16.857 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:200 Resp: 396143
Ion Ratio Lower Upper
200 100
173 27.0 21.3 31.9
215 47.5 39.8 59.8



#71
Pentachlorophenol
Concen: 20.280 ng/ul
RT: 17.045 min Scan# 2373
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:266 Resp: 283774
Ion Ratio Lower Upper
266 100
264 60.3 50.5 75.7
268 60.0 52.6 78.8









#72
 Phenanthrene
 Concen: 19.855 ng/ul
 RT: 17.439 min Scan# 2440
 Delta R.T. 0.000 min
 Lab File: BN023243.D
 Acq: 16 Dec 2022 17:03

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020373

Tgt Ion:178 Resp: 1973512




Ion	Ratio	Lower	Upper
178	100		
179	15.3	12.3	18.5
176	20.0	15.4	23.0




#74
 Anthracene
 Concen: 20.212 ng/ul
 RT: 17.533 min Scan# 2456
 Delta R.T. 0.000 min
 Lab File: BN023243.D
 Acq: 16 Dec 2022 17:03

Tgt Ion:178 Resp: 2004549

Ion	Ratio	Lower	Upper
178	100		
179	15.5	11.8	17.6
176	19.0	14.7	22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.852 ng/uL

RT: 13.451 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:216 Resp: 463730

Ion Ratio Lower Upper

216 100

214 77.6 61.9 92.9

218 49.3 39.4 59.2



#76

Pentachlorobenzene

Concen: 20.099 ng/uL

RT: 14.969 min Scan# 2020

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03




Tgt Ion:250 Resp: 475777


Ion Ratio Lower Upper

250 100

248 64.5 49.4 74.0

252 64.7 52.9 79.3





#77

Carbazole

Concen: 20.592 ng/ul

RT: 17.804 min Scan# 2502

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:167 Resp: 1880490

Ion Ratio Lower Upper

167 100

166 21.5 17.3 25.9

139 13.0 10.6 16.0



#78

Di-n-butylphthalate

Concen: 20.672 ng/ul

RT: 18.363 min Scan# 2597

Delta R.T. 0.000 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03




Tgt Ion:149 Resp: 2241536


Ion Ratio Lower Upper

149 100

150 8.8 7.4 11.2

104 5.0 4.1 6.1









#80
Fluoranthene
Concen: 20.837 ng/ul
RT: 19.457 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:202 Resp: 2373119
Ion Ratio Lower Upper
202 100
101 12.0 10.4 15.6
100 9.1 7.6 11.4



#82
Pyrene
Concen: 20.947 ng/ul
RT: 19.822 min Scan# 2845
Delta R.T. 0.006 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:202 Resp: 2506049
Ion Ratio Lower Upper
202 100
101 13.3 12.0 18.0
100 11.1 9.5 14.3





#83

Butylbenzylphthalate

Concen: 21.939 ng/ul

RT: 20.716 min Scan# 2997

Delta R.T. 0.006 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:149 Resp: 1086173

Ion Ratio Lower Upper

149 100

91 69.6 57.9 86.9

206 20.6 15.6 23.4



#84

3,3'-Dichlorobenzidine

Concen: 19.730 ng/ul

RT: 21.504 min Scan# 3131

Delta R.T. 0.006 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03




Tgt Ion:252 Resp: 780514


Ion Ratio Lower Upper

252 100

254 66.5 50.3 75.5

126 11.5 9.4 14.0









#85
Benzo(a)anthracene
Concen: 20.380 ng/ul
RT: 21.569 min Scan# 3142
Delta R.T. 0.000 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:228 Resp: 2458066
Ion Ratio Lower Upper
228 100
229 19.6 15.5 23.3
226 26.2 21.4 32.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 22.194 ng/ul
RT: 21.492 min Scan# 3129
Delta R.T. 0.006 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:149 Resp: 1553707
Ion Ratio Lower Upper
149 100
167 28.9 21.9 32.9
279 4.6 3.4 5.0









#87
Chrysene
Concen: 19.789 ng/ul
RT: 21.627 min Scan# 3152
Delta R.T. 0.006 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:228 Resp: 2252390
Ion Ratio Lower Upper
228 100
226 31.3 24.6 36.8
229 20.6 16.5 24.7



#89
Di-n-octyl phthalate
Concen: 23.236 ng/ul
RT: 22.433 min Scan# 3289
Delta R.T. 0.006 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:149 Resp: 2619070









#90
Benzo(b)fluoranthene
Concen: 21.206 ng/ul
RT: 23.286 min Scan# 3434
Delta R.T. 0.006 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:252 Resp: 2489539
Ion Ratio Lower Upper
252 100
253 21.4 16.8 25.2
125 11.4 8.8 13.2



#91
Benzo(k)fluoranthene
Concen: 20.920 ng/ul
RT: 23.339 min Scan# 3443
Delta R.T. 0.006 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:252 Resp: 2337183
Ion Ratio Lower Upper
252 100
253 21.7 17.9 26.9
125 11.0 9.0 13.6





#93

Benzo(a)pyrene

Concen: 20.643 ng/ul

RT: 23.933 min Scan# 3544

Delta R.T. 0.006 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:252 Resp: 2082691

Ion Ratio Lower Upper

252 100

253 22.8 17.6 26.4

125 12.2 9.4 14.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.158 ng/ul

RT: 26.609 min Scan# 3999

Delta R.T. 0.012 min

Lab File: BN023243.D

Acq: 16 Dec 2022 17:03




Tgt Ion:276 Resp: 2448695


Ion Ratio Lower Upper

276 100

138 24.3 20.6 30.8

277 24.3 18.2 27.2









#95
Dibenzo(a,h)anthracene
Concen: 19.285 ng/ul
RT: 26.627 min Scan# 4003
Delta R.T. 0.006 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Instrument :
BNA_N
ClientSampleId :
SSTD020373

Tgt Ion:278 Resp: 2009812
Ion Ratio Lower Upper
278 100
139 17.2 13.3 19.9
279 24.3 18.6 27.8



#96
Benzo(g,h,i)perylene
Concen: 18.661 ng/ul
RT: 27.404 min Scan# 4134
Delta R.T. 0.012 min
Lab File: BN023243.D
Acq: 16 Dec 2022 17:03

Tgt Ion:276 Resp: 1887144
Ion Ratio Lower Upper
276 100
138 21.2 19.2 28.8
277 23.3 19.3 28.9

