

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121622\
 Data File : BN023255.D
 Acq On : 17 Dec 2022 00:30
 Operator : CG/JU
 Sample : N5983-16
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 COLF2

Quant Time: Dec 17 00:59:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 15 23:22:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.999	152	296514	20.000	ng/u1	0.00
20) Naphthalene-d8	10.822	136	1376368	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	913010	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.398	188	2083375	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	2010602	20.000	ng/u1	0.00
88) Perylene-d12	24.039	264	1994857	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.346	96	34942	4.876	ng/uL	0.00
4) Pyridine-d5	3.775	84	424300	20.056	ng/u1	0.00
7) Phenol-d5	7.158	99	742390	26.920	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	453917	28.284	ng/u1	0.00
11) 2-Chlorophenol-d4	7.528	132	598444	28.798	ng/u1	0.00
15) 4-Methylphenol-d8	8.710	113	626859	28.311	ng/u1	0.00
21) Nitrobenzene-d5	9.169	128	314251	29.551	ng/u1	0.00
24) 2-Nitrophenol-d4	9.899	143	348058	31.075	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	605943	28.658	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	907241	26.666	ng/u1	0.00
46) Dimethylphthalate-d6	14.057	166	2165888	31.655	ng/u1	0.00
49) Acenaphthylene-d8	14.345	160	2428375	31.648	ng/u1	0.00
54) 4-Nitrophenol-d4	14.840	143	420094	27.491	ng/u1	0.00
60) Fluorene-d10	15.640	176	1892791	32.476	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	324033	26.422	ng/u1	0.00
73) Anthracene-d10	17.498	188	3003413	32.167	ng/u1	0.00
81) Pyrene-d10	19.786	212	3654361	33.412	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.880	264	3251378	33.253	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121622\
Data File : BN023255.D
Acq On : 17 Dec 2022 00:30
Operator : CG/JU
Sample : N5983-16
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
COLF2

Quant Time: Dec 17 00:59:08 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 15 23:22:36 2022
Response via : Initial Calibration

