

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121622\
 Data File : BN023256.D
 Acq On : 17 Dec 2022 01:07
 Operator : CG/JU
 Sample : N5983-01
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 COLB0

Quant Time: Dec 17 01:38:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 15 23:22:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.999	152	303673	20.000	ng/u1	0.00
20) Naphthalene-d8	10.822	136	1407682	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.651	164	938503	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.398	188	2158290	20.000	ng/u1	0.00
79) Chrysene-d12	21.580	240	2061048	20.000	ng/u1	0.00
88) Perylene-d12	24.039	264	2073972	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.346	96	33013	4.498	ng/uL	0.00
4) Pyridine-d5	3.776	84	344649	15.907	ng/u1	0.00
7) Phenol-d5	7.152	99	742835	26.301	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	441841	26.883	ng/u1	0.00
11) 2-Chlorophenol-d4	7.528	132	580062	27.256	ng/u1	0.00
15) 4-Methylphenol-d8	8.711	113	626603	27.632	ng/u1	0.00
21) Nitrobenzene-d5	9.169	128	308163	28.334	ng/u1	0.00
24) 2-Nitrophenol-d4	9.899	143	347683	30.351	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	619377	28.641	ng/u1	0.00
31) 4-Chloroaniline-d4	10.958	131	873644	25.108	ng/u1	0.00
46) Dimethylphthalate-d6	14.057	166	2226252	31.654	ng/u1	0.00
49) Acenaphthylene-d8	14.346	160	2469919	31.315	ng/u1	0.00
54) 4-Nitrophenol-d4	14.840	143	429958	27.372	ng/u1	0.00
60) Fluorene-d10	15.640	176	1926360	32.154	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	334153	26.301	ng/u1	0.00
73) Anthracene-d10	17.498	188	2961054	30.613	ng/u1	0.00
81) Pyrene-d10	19.786	212	3708506	33.077	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.874	264	3305018	32.512	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121622\
Data File : BN023256.D
Acq On : 17 Dec 2022 01:07
Operator : CG/JU
Sample : N5983-01
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
COLB0

Quant Time: Dec 17 01:38:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN121522.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 15 23:22:36 2022
Response via : Initial Calibration

