

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121625\
 Data File : BN038462.D
 Acq On : 17 Dec 2025 03:48
 Operator : RC/JU
 Sample : Q3863-10
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EB01-121225

Quant Time: Dec 17 06:43:21 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.710	152	4656	0.400	ng	-0.01
7) Naphthalene-d8	10.509	136	11829	0.400	ng	#-0.02
13) Acenaphthene-d10	14.377	164	6316	0.400	ng	-0.02
19) Phenanthrene-d10	17.136	188	11474	0.400	ng	#-0.01
29) Chrysene-d12	21.331	240	8635	0.400	ng	# 0.00
35) Perylene-d12	23.613	264	8529	0.400	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.276	112	1876	0.162	ng	-0.01
5) Phenol-d6	6.872	99	1211	0.088	ng	-0.01
8) Nitrobenzene-d5	8.865	82	4111	0.412	ng	-0.01
11) 2-Methylnaphthalene-d10	12.111	152	6168	0.370	ng	-0.02
14) 2,4,6-Tribromophenol	15.882	330	840	0.300	ng	-0.01
15) 2-Fluorobiphenyl	12.998	172	15746	0.517	ng	-0.01
27) Fluoranthene-d10	19.174	212	11427	0.403	ng	-0.01
31) Terphenyl-d14	19.773	244	8774	0.455	ng	-0.02
Target Compounds						Qvalue
34) Bis(2-ethylhexyl)phtha...	21.250	149	941	0.051	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121625\
Data File : BN038462.D
Acq On : 17 Dec 2025 03:48
Operator : RC/JU
Sample : Q3863-10
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EB01-121225

Quant Time: Dec 17 06:43:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 10 13:56:50 2025
Response via : Initial Calibration

