

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121625\
 Data File : BN038434.D
 Acq On : 16 Dec 2025 10:13
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 17 03:21:10 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	80	-0.01
2	1,4-Dioxane	0.400	0.406	-1.5	80	0.00
3	n-Nitrosodimethylamine	0.400	0.397	0.8	80	0.00
4 S	2-Fluorophenol	0.400	0.384	4.0	78	0.00
5 S	Phenol-d6	0.400	0.367	8.3	76	0.00
6	bis(2-Chloroethyl)ether	0.400	0.364	9.0	73	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	78	-0.02
8 S	Nitrobenzene-d5	0.400	0.381	4.8	77	-0.01
9	Naphthalene	0.400	0.382	4.5	76	-0.01
10	Hexachlorobutadiene	0.400	0.384	4.0	75	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.378	5.5	76	-0.02
12	2-Methylnaphthalene	0.400	0.379	5.3	76	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	84	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.359	10.3	80	-0.01
15 S	2-Fluorobiphenyl	0.400	0.459	-14.8	94	-0.01
16	Acenaphthylene	0.400	0.367	8.3	79	-0.01
17	Acenaphthene	0.400	0.363	9.3	78	-0.02
18	Fluorene	0.400	0.370	7.5	80	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	86	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.241	39.8#	60	0.00
21	4-Bromophenyl-phenylether	0.400	0.368	8.0	82	-0.01
22	Hexachlorobenzene	0.400	0.381	4.8	83	-0.02
23	Atrazine	0.400	0.346	13.5	79	-0.01
24	Pentachlorophenol	0.400	0.335	16.3	83	-0.01
25	Phenanthrene	0.400	0.377	5.8	83	-0.01
26	Anthracene	0.400	0.366	8.5	83	-0.01
27 SURR	Fluoranthene-d10	0.400	0.375	6.3	81	-0.01
28	Fluoranthene	0.400	0.386	3.5	84	-0.02
29 I	Chrysene-d12	0.400	0.400	0.0	85	0.00
30	Pyrene	0.400	0.376	6.0	82	-0.01
31 S	Terphenyl-d14	0.400	0.374	6.5	81	-0.01
32	Benzo(a)anthracene	0.400	0.372	7.0	83	-0.02
33	Chrysene	0.400	0.397	0.8	84	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.385	3.8	82	0.00
35 I	Perylene-d12	0.400	0.400	0.0	87	-0.02
36	Indeno(1,2,3-cd)pyrene	0.400	0.397	0.8	86	-0.03
37	Benzo(b)fluoranthene	0.400	0.382	4.5	85	-0.02
38	Benzo(k)fluoranthene	0.400	0.390	2.5	86	-0.02
39 C	Benzo(a)pyrene	0.400	0.388	3.0	86	-0.02
40	Dibenzo(a,h)anthracene	0.400	0.396	1.0	86	-0.03
41	Benzo(g,h,i)perylene	0.400	0.392	2.0	86	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4