

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121625\
 Data File : BN038449.D
 Acq On : 16 Dec 2025 19:15
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 17 02:54:42 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	-0.01
2	1,4-Dioxane	0.459	0.469	-2.2	85	0.00
3	n-Nitrosodimethylamine	0.575	0.570	0.9	84	0.00
4 S	2-Fluorophenol	0.996	0.982	1.4	84	-0.01
5 S	Phenol-d6	1.184	1.140	3.7	84	-0.01
6	bis(2-Chloroethyl)ether	1.158	1.096	5.4	79	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	87	-0.02
8 S	Nitrobenzene-d5	0.337	0.318	5.6	84	-0.01
9	Naphthalene	1.182	1.133	4.1	84	-0.01
10	Hexachlorobutadiene	0.210	0.199	5.2	82	-0.02
11 SURR	2-Methylnaphthalene-d10	0.564	0.540	4.3	85	-0.02
12	2-Methylnaphthalene	0.751	0.717	4.5	85	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	92	-0.02
14 S	2,4,6-Tribromophenol	0.177	0.156	11.9	86	-0.01
15 S	2-Fluorobiphenyl	1.927	2.340	-21.4	109	-0.01
16	Acenaphthylene	1.964	1.815	7.6	88	-0.01
17	Acenaphthene	1.369	1.270	7.2	87	-0.02
18	Fluorene	1.761	1.639	6.9	88	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	-0.01
20	4,6-Dinitro-2-methylphenol	0.066	0.038	42.4#	63	-0.01
21	4-Bromophenyl-phenylether	0.297	0.269	9.4	87	-0.01
22	Hexachlorobenzene	0.357	0.335	6.2	89	-0.02
23	Atrazine	0.202	0.175	13.4	85	-0.01
24	Pentachlorophenol	0.139	0.114	18.0	87	-0.01
25	Phenanthrene	1.392	1.314	5.6	89	-0.01
26	Anthracene	1.200	1.094	8.8	89	-0.01
27 SURR	Fluoranthene-d10	0.989	0.931	5.9	87	-0.01
28	Fluoranthene	1.480	1.425	3.7	90	-0.02
29 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
30	Pyrene	1.969	1.904	3.3	90	-0.02
31 S	Terphenyl-d14	0.893	0.859	3.8	88	-0.02
32	Benzo(a)anthracene	1.545	1.482	4.1	91	-0.02
33	Chrysene	1.701	1.687	0.8	89	0.00
34	Bis(2-ethylhexyl)phthalate	0.848	1.661	-95.9#	178#	-0.02
35 I	Perylene-d12	1.000	1.000	0.0	90	-0.02
36	Indeno(1,2,3-cd)pyrene	2.123	2.122	0.0	90	-0.04
37	Benzo(b)fluoranthene	1.837	1.770	3.6	89	-0.02
38	Benzo(k)fluoranthene	1.832	1.820	0.7	91	-0.02
39 C	Benzo(a)pyrene	1.503	1.477	1.7	90	-0.02
40	Dibenzo(a,h)anthracene	1.637	1.633	0.2	90	-0.04
41	Benzo(g,h,i)perylene	1.829	1.808	1.1	89	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4