

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121625\
 Data File : BN038464.D
 Acq On : 17 Dec 2025 05:00
 Operator : RC/JU
 Sample : Q3882-19
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 COAM8

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 12/17/2025
 Supervised By :Jagrut Upadhyay 12/17/2025

Quant Time: Dec 17 06:43:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 10 13:56:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.710	152	4849	0.400	ng	-0.01
7) Naphthalene-d8	10.509	136	12962	0.400	ng	#-0.02
13) Acenaphthene-d10	14.377	164	6727	0.400	ng	-0.02
19) Phenanthrene-d10	17.136	188	12863	0.400	ng	#-0.01
29) Chrysene-d12	21.331	240	11472	0.400	ng	0.00
35) Perylene-d12	23.613	264	11850	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.276	112	2378	0.197	ng	-0.01
5) Phenol-d6	6.872	99	2019	0.141	ng	-0.01
8) Nitrobenzene-d5	8.864	82	3379	0.309	ng	-0.01
11) 2-Methylnaphthalene-d10	12.111	152	5150	0.282	ng	-0.02
14) 2,4,6-Tribromophenol	15.882	330	1066	0.357	ng	-0.01
15) 2-Fluorobiphenyl	12.998	172	12637	0.390	ng	-0.01
27) Fluoranthene-d10	19.173	212	8992	0.283	ng	-0.01
31) Terphenyl-d14	19.773	244	7333	0.286	ng	-0.02
Target Compounds						Qvalue
2) 1,4-Dioxane	3.189	88	468	0.084	ng	# 80
9) Naphthalene	10.562	128	1606	0.042	ng	# 82
12) 2-Methylnaphthalene	12.182	142	1216	0.050	ng	# 94
16) Acenaphthylene	14.099	152	801	0.024	ng	# 90
17) Acenaphthene	14.441	154	784	0.034	ng	96
18) Fluorene	15.435	166	3793	0.128	ng	96
25) Phenanthrene	17.173	178	37264	0.832	ng	99
26) Anthracene	17.260	178	10271m	0.266	ng	
28) Fluoranthene	19.201	202	94958	1.996	ng	98
30) Pyrene	19.564	202	58493	1.036	ng	98
32) Benzo(a)anthracene	21.313	228	26771	0.604	ng	# 94
33) Chrysene	21.366	228	45551	0.934	ng	# 92
34) Bis(2-ethylhexyl)phtha...	21.250	149	11110	0.457	ng	98
36) Indeno(1,2,3-cd)pyrene	25.928	276	8530	0.136	ng	98
37) Benzo(b)fluoranthene	22.923	252	40268m	0.740	ng	
38) Benzo(k)fluoranthene	22.952	252	11052m	0.204	ng	
39) Benzo(a)pyrene	23.510	252	2962m	0.067	ng	
40) Dibenzo(a,h)anthracene	25.934	278	3273	0.068	ng	# 56
41) Benzo(g,h,i)perylene	26.636	276	9167	0.169	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121625\
Data File : BN038464.D
Acq On : 17 Dec 2025 05:00
Operator : RC/JU
Sample : Q3882-19
Misc :
ALS Vial : 15 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
COAM8

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 12/17/2025
Supervised By :Jagrut Upadhyay 12/17/2025


Quant Time: Dec 17 06:43:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN121025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 10 13:56:50 2025
Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.710 min Scan# 64
 Delta R.T. -0.014 min
 Lab File: BN038464.D
 Acq: 17 Dec 2025 05:00




Instrument :
 BNA_N
 ClientSampleId :
 COAM8




Tgt Ion: 152 Resp: 4849
 Ion Ratio Lower Upper
 152 100
 150 151.8 122.4 183.6
 115 59.8 47.3 70.9

Manual Integrations
 APPROVED



Reviewed By :Rahul Chavli 12/17/2025
 Supervised By :Jagrut Upadhyay 12/17/2025




#2
 1,4-Dioxane
 Concen: 0.084 ng
 RT: 3.189 min Scan# 15
 Delta R.T. -0.007 min
 Lab File: BN038464.D
 Acq: 17 Dec 2025 05:00



Tgt Ion: 88 Resp: 468
 Ion Ratio Lower Upper
 88 100
 43 0.0 24.9 37.3#
 58 81.4 61.4 92.0









#4
2-Fluorophenol
Concen: 0.197 ng
RT: 5.276 min Scan# 304
Delta R.T. -0.014 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00

Instrument :
BNA_N
ClientSampleId :
COAM8

Tgt Ion:112 Resp: 2378
Ion Ratio Lower Upper
112 100
64 54.5 44.4 66.6
63 30.6 23.4 35.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 12/17/2025
Supervised By :Jagrut Upadhyay 12/17/2025



#5
Phenol-d6
Concen: 0.141 ng
RT: 6.872 min Scan# 525
Delta R.T. -0.014 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00


Tgt Ion: 99 Resp: 2019
Ion Ratio Lower Upper
99 100
42 31.9 16.5 24.7#
71 41.0 24.9 37.3#





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.509 min Scan# 931
Delta R.T. -0.021 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00




Instrument :
BNA_N
ClientSampleId :
COAM8



Tgt Ion:136 Resp: 12962
Ion Ratio Lower Upper
136 100
137 11.9 9.1 13.7
54 9.2 5.2 7.8
68 6.9 4.1 6.1#




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 12/17/2025
Supervised By :Jagrut Upadhyay 12/17/2025



#8
Nitrobenzene-d5
Concen: 0.309 ng
RT: 8.864 min Scan# 776
Delta R.T. -0.011 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00

Tgt Ion: 82 Resp: 3379
Ion Ratio Lower Upper
82 100
128 49.4 34.0 51.0
54 54.7 44.4 66.6









#9
Naphthalene
Concen: 0.042 ng
RT: 10.562 min Scan# 936
Delta R.T. -0.011 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00

Instrument :
BNA_N
ClientSampleId :
COAM8

Tgt Ion:128 Resp: 1600
Ion Ratio Lower Upper
128 100
129 19.7 9.1 13.7
127 19.7 10.9 16.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 12/17/2025
Supervised By :Jagrut Upadhyay 12/17/2025



#11
2-Methylnaphthalene-d10
Concen: 0.282 ng
RT: 12.111 min Scan# 1142
Delta R.T. -0.015 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00


Tgt Ion:152 Resp: 5150
Ion Ratio Lower Upper
152 100
151 21.8 17.0 25.6





#12
2-Methylnaphthalene
Concen: 0.050 ng
RT: 12.182 min Scan# 1156
Delta R.T. -0.020 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00




Instrument :
BNA_N
ClientSampleId :
COAM8



Tgt Ion:142 Resp: 1216
Ion Ratio Lower Upper
142 100
141 90.0 71.4 107.2
115 47.4 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 12/17/2025
Supervised By :Jagrut Upadhyay 12/17/2025



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.377 min Scan# 1447
Delta R.T. -0.021 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00


Tgt Ion:164 Resp: 6727
Ion Ratio Lower Upper
164 100
162 108.3 86.2 129.4
160 53.8 44.6 67.0





#14
2,4,6-Tribromophenol
Concen: 0.357 ng
RT: 15.882 min Scan# 11
Delta R.T. -0.012 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00




Instrument :
BNA_N
ClientSampleId :
COAM8




Tgt Ion:330 Resp: 1066
Ion Ratio Lower Upper
330 100
332 93.9 75.8 113.6
141 40.9 31.8 47.8

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 12/17/2025
Supervised By :Jagrut Upadhyay 12/17/2025




#15
2-Fluorobiphenyl
Concen: 0.390 ng
RT: 12.998 min Scan# 1317
Delta R.T. -0.010 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00



Tgt Ion:172 Resp: 12637
Ion Ratio Lower Upper
172 100
171 35.7 28.6 42.8
170 23.6 19.3 28.9





#16

Acenaphthylene

Concen: 0.024 ng

RT: 14.099 min Scan# 1421

Delta R.T. -0.011 min

Lab File: BN038464.D

Acq: 17 Dec 2025 05:00

Instrument :

BNA_N

ClientSampleId :

COAM8

Tgt Ion:152 Resp: 801





Ion	Ratio	Lower	Upper
152	100		
151	23.7	15.4	23.2
153	16.5	10.2	15.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 12/17/2025

Supervised By :Jagrut Upadhyay 12/17/2025



#17

Acenaphthene

Concen: 0.034 ng

RT: 14.441 min Scan# 1453




Delta R.T. -0.021 min


Lab File: BN038464.D

Acq: 17 Dec 2025 05:00

Tgt Ion:154 Resp: 784


Ion	Ratio	Lower	Upper
154	100		
153	118.1	89.8	134.8
152	58.3	47.5	71.3





#18
Fluorene
Concen: 0.128 ng
RT: 15.435 min Scan# 1546
Delta R.T. -0.011 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00




Instrument :
BNA_N
ClientSampleId :
COAM8




Tgt Ion:166 Resp: 3793
Ion Ratio Lower Upper
166 100
165 96.1 80.2 120.4
167 15.5 10.6 16.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 12/17/2025
Supervised By :Jagrut Upadhyay 12/17/2025




#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.136 min Scan# 1684
Delta R.T. -0.012 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00




Tgt Ion:188 Resp: 12863
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.2 7.4 11.0#





#25
Phenanthrene
Concen: 0.832 ng
RT: 17.173 min Scan# 1687
Delta R.T. -0.012 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00




Instrument :
BNA_N
ClientSampleId :
COAM8




Tgt Ion:178 Resp: 37264
Ion Ratio Lower Upper
178 100
176 19.2 15.5 23.3
179 16.0 12.2 18.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 12/17/2025
Supervised By :Jagrut Upadhyay 12/17/2025




#26
Anthracene
Concen: 0.266 ng m
RT: 17.260 min Scan# 1694
Delta R.T. -0.025 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00



Tgt Ion:178 Resp: 10271
Ion Ratio Lower Upper
178 100
176 69.5 15.1 22.7#
179 56.8 12.0 18.0#









#27
Fluoranthene-d10
Concen: 0.283 ng
RT: 19.173 min Scan# 1937
Delta R.T. -0.014 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00

Instrument :
BNA_N
ClientSampleId :
COAM8

Tgt Ion:212 Resp: 8992
Ion Ratio Lower Upper
212 100
106 20.6 13.7 20.5
104 14.9 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 12/17/2025
Supervised By :Jagrut Upadhyay 12/17/2025



#28
Fluoranthene
Concen: 1.996 ng
RT: 19.201 min Scan# 1943
Delta R.T. -0.019 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00


Tgt Ion:202 Resp: 94958
Ion Ratio Lower Upper
202 100
101 13.3 9.6 14.4
203 17.3 13.5 20.3





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.331 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00




Instrument :
BNA_N
ClientSampleId :
COAM8




Tgt Ion:240 Resp: 11472
Ion Ratio Lower Upper
240 100
120 18.8 15.4 23.2
236 32.0 23.9 35.9

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 12/17/2025
Supervised By :Jagrut Upadhyay 12/17/2025




#30
Pyrene
Concen: 1.036 ng
RT: 19.564 min Scan# 2021
Delta R.T. -0.019 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00




Tgt Ion:202 Resp: 58493
Ion Ratio Lower Upper
202 100
200 21.2 16.9 25.3
203 19.0 14.2 21.2





#31
 Terphenyl-d14
 Concen: 0.286 ng
 RT: 19.773 min Scan# 2066
 Delta R.T. -0.019 min
 Lab File: BN038464.D
 Acq: 17 Dec 2025 05:00




Instrument :
 BNA_N
 ClientSampleId :
 COAM8




Tgt Ion:244 Resp: 7333
 Ion Ratio Lower Upper
 244 100
 212 12.2 7.9 11.9
 122 23.6 15.0 22.6

Manual Integrations
 APPROVED



Reviewed By :Rahul Chavli 12/17/2025
 Supervised By :Jagrut Upadhyay 12/17/2025




#32
 Benzo(a)anthracene
 Concen: 0.604 ng
 RT: 21.313 min Scan# 2262
 Delta R.T. -0.018 min
 Lab File: BN038464.D
 Acq: 17 Dec 2025 05:00




Tgt Ion:228 Resp: 26771
 Ion Ratio Lower Upper
 228 100
 226 29.6 23.3 34.9
 229 25.9 15.7 23.5





#33
Chrysene
Concen: 0.934 ng
RT: 21.366 min Scan# 2268
Delta R.T. -0.009 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00




Instrument :
BNA_N
ClientSampleId :
COAM8



Tgt Ion:228 Resp: 4555
Ion Ratio Lower Upper
228 100
226 30.8 25.0 37.4
229 28.4 15.5 23.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 12/17/2025
Supervised By :Jagrut Upadhyay 12/17/2025



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.457 ng
RT: 21.250 min Scan# 2255
Delta R.T. -0.018 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00


Tgt Ion:149 Resp: 11110
Ion Ratio Lower Upper
149 100
167 27.9 21.4 32.0
279 3.0 2.4 3.6





#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.613 min Scan# 21
 Delta R.T. -0.021 min
 Lab File: BN038464.D
 Acq: 17 Dec 2025 05:00




Instrument :
 BNA_N
 ClientSampleId :
 COAM8




Tgt Ion:264 Resp: 11850
 Ion Ratio Lower Upper
 264 100
 260 28.4 22.2 33.2
 265 79.9 80.2 120.2

Manual Integrations
 APPROVED



Reviewed By :Rahul Chavli 12/17/2025
 Supervised By :Jagrut Upadhyay 12/17/2025




#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.136 ng
 RT: 25.928 min Scan# 3589
 Delta R.T. -0.035 min
 Lab File: BN038464.D
 Acq: 17 Dec 2025 05:00




Tgt Ion:276 Resp: 8530
 Ion Ratio Lower Upper
 276 100
 138 27.2 23.6 35.4
 277 24.4 19.7 29.5





#37
Benzo(b)fluoranthene
Concen: 0.740 ng m
RT: 22.923 min Scan# 2567
Delta R.T. -0.023 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00




Instrument :
BNA_N
ClientSampleId :
COAM8




Tgt Ion:252 Resp: 40268
Ion Ratio Lower Upper
252 100
253 26.4 21.1 31.7
125 21.1 16.4 24.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 12/17/2025
Supervised By :Jagrut Upadhyay 12/17/2025




#38
Benzo(k)fluoranthene
Concen: 0.204 ng m
RT: 22.952 min Scan# 2571
Delta R.T. -0.038 min
Lab File: BN038464.D
Acq: 17 Dec 2025 05:00



Tgt Ion:252 Resp: 11052
Ion Ratio Lower Upper
252 100
253 34.3 21.4 32.0#
125 35.5 16.2 24.2#





#39

Benzo(a)pyrene

Concen: 0.067 ng m

RT: 23.510 min Scan# 21

Delta R.T. -0.023 min

Lab File: BN038464.D

Acq: 17 Dec 2025 05:00

Instrument :

BNA_N

ClientSampleId :

COAM8





Tgt Ion:252	Resp:	2962
Ion Ratio	Lower	Upper
252	100	
253	59.4	23.3 34.9#
125	75.8	22.4 33.6#

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 12/17/2025

Supervised By :Jagrut Upadhyay 12/17/2025



#40

Dibenzo(a,h)anthracene

Concen: 0.068 ng




RT: 25.934 min Scan# 3591


Delta R.T. -0.047 min

Lab File: BN038464.D

Acq: 17 Dec 2025 05:00


Tgt Ion:278	Resp:	3273
Ion Ratio	Lower	Upper
278	100	
139	43.2	17.8 26.8#
279	49.7	21.3 31.9#





#41
 Benzo(g,h,i)perylene
 Concen: 0.169 ng
 RT: 26.636 min Scan# 3842
 Delta R.T. -0.041 min
 Lab File: BN038464.D
 Acq: 17 Dec 2025 05:00

Instrument :
 BNA_N
 ClientSampleId :
 COAM8



Tgt Ion: 276 Resp: 9162

Ion	Ratio	Lower	Upper
276	100		
277	26.9	20.0	30.0
138	31.5	21.8	32.6

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 12/17/2025
 Supervised By :Jagrut Upadhyay 12/17/2025

