

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN121718\
Quantitation Report (Not Reviewed)

Data File : BN004060.D
Acq On : 18 Dec 2018 05:34
Operator : JU/SJ
Sample : J6270-08
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9E07

Quant Time: Dec 18 06:39:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 18 06:35:27 2018
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	1110	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	6161	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	4098	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.31	188	727474	0.40	ng/ul	0.10
16) Chrysene-d12	0.00	240	0	0.00	ng/ul	-21.40
20) Perylene-d12	23.72	264	14322	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	2615	0.28	ng/ul	0.00
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	
Target Compounds						
23) Benzo(a)pyrene	23.58	252	3544	0.069	ng/ul#	Qvalue 50


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN004060.D
Acq On : 18 Dec 2018 05:34
Operator : JU/SJ
Sample : J6270-08
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9E07

Quant Time: Dec 18 06:39:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN11918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 18 06:35:27 2018
Response via : Initial Calibration





#23

Benzo(a)pyrene

Concen: 0.069 ng/ul

RT: 23.58 min Scan# 5494

Delta R.T. -0.04 min

Lab File: BN004060.D


Acq: 18 Dec 2018 05:34

Instrument :

BNA_N

ClientSampleId :

F9E07



Tgt Ion: 252 Resp: 3544

Ion Ratio Lower Upper

252 100

253 43.2 19.0 28.4#

125 40.0 10.2 15.4#

