

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121718\
 Data File : BN004065.D
 Acq On : 18 Dec 2018 08:29
 Operator : JU/SJ
 Sample : J6270-13
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F9E12

Quant Time: Dec 18 11:10:06 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 18 06:35:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.83	152	1988	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	9896	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	5977	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.31	188	935766	0.40	ng/ul	0.10
16) Chrysene-d12	0.00	240	0	0.00	ng/ul	-21.40
20) Perylene-d12	23.73	264	16957	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	4124	0.27	ng/ul	0.00
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	
Target Compounds						
					Qvalue	
7) Acenaphthylene	14.19	152	13235	0.492	ng/ul	99
21) Benzo(b)fluoranthene	23.02	252	124258	1.771	ng/ul	98
22) Benzo(k)fluoranthene	23.02	252	124245	1.818	ng/ul	98
23) Benzo(a)pyrene	23.63	252	44016	0.725	ng/ul	97
24) Indeno(1,2,3-cd)pyrene	26.09	276	39433	0.566	ng/ul#	94
25) Dibenzo(a,h)anthracene	26.09	278	10974	0.196	ng/ul#	69
26) Benzo(g,h,i)perylene	26.82	276	36408	0.601	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121718\
Data File : BN004065.D
Acq On : 18 Dec 2018 08:29
Operator : JU/SJ
Sample : J6270-13
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9E12

Quant Time: Dec 18 11:10:06 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 18 06:35:27 2018
Response via : Initial Calibration

