

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN121718\
 Data File : BN004051.D
 Acq On : 18 Dec 2018 00:17
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.426

Quant Time: Dec 18 04:34:39 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 18 04:29:07 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	1251	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	6627	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	4030	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.22	188	10267	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	11975	0.40	ng/ul	0.00
20) Perylene-d12	23.72	264	10901	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.22	152	4237	0.42	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	13282	0.44	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Naphthalene	10.68	128	7526	0.404	ng/ul	98
5) 2-Methylnaphthalene	12.29	142	5598	0.431	ng/ul	99
7) Acenaphthylene	14.19	152	8606	0.474	ng/ul	99
8) Acenaphthene	14.53	153	6551	0.412	ng/ul	99
9) Fluorene	15.52	166	7596	0.410	ng/ul	99
11) Pentachlorophenol	16.87	266	796	0.430	ng/ul	99
12) Phenanthrene	17.26	178	13519	0.401	ng/ul	99
13) Anthracene	17.35	178	12800	0.462	ng/ul	99
15) Fluoranthene	19.28	202	17709	0.445	ng/ul	97
17) Pyrene	19.64	202	18110	0.422	ng/ul	97
18) Benzo(a)anthracene	21.38	228	17381	0.460	ng/ul	98
19) Chrysene	21.44	228	17066	0.394	ng/ul	99
21) Benzo(b)fluoranthene	23.01	252	16698	0.370	ng/ul	95
22) Benzo(k)fluoranthene	23.06	252	16650	0.379	ng/ul#	96
23) Benzo(a)pyrene	23.62	252	15696	0.402	ng/ul#	95
24) Indeno(1,2,3-cd)pyrene	26.08	276	18425	0.411	ng/ul	97
25) Dibenzo(a,h)anthracene	26.09	278	14828	0.411	ng/ul	96
26) Benzo(g,h,i)perylene	26.81	276	15656	0.402	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN121718\
Data File : BN004051.D
Acq On : 18 Dec 2018 00:17
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.426

Quant Time: Dec 18 04:34:39 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN11918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 18 04:29:07 2018
Response via : Initial Calibration

