


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121718\
Data File : BN004033.D
Acq On : 17 Dec 2018 13:39
Operator : JU/SJ
Sample : DFTPP71
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP71

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN11918.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Dec 19 01:54:36 2018



AutoFind: Scans 2488, 2489, 2490; Background Corrected with Scan 2481

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.3	33869	PASS
68	69	0.00	2	2.0	766	PASS
69	198	0.00	100	32.4	38688	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	43.6	52165	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	119528	PASS
199	198	5	9	6.8	8106	PASS
275	198	10	60	32.6	39005	PASS
365	198	1	100	4.0	4809	PASS
441	198	0.01	100	15.6	18642	PASS
442	198	50	100	99.9	119434	PASS
443	442	15	24	19.0	22720	PASS

m/z	Abundance
43.90	523.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

m/z	Abundance
38.00	515.0
39.00	3172.0
44.00	544.0
49.10	326.0
50.00	10747.0
51.00	42256.0
52.00	2147.0
56.00	1266.0
57.00	2999.0
61.00	631.0
62.00	629.0
63.00	1755.0
65.00	860.0
68.20	916.0
69.00	47728.0
73.00	395.0
74.00	4645.0
75.00	7236.0
76.00	2646.0
77.00	53368.0
78.00	3655.0
79.00	3579.0
80.00	2582.0
81.00	3821.0
82.00	958.0
83.00	999.0
85.00	667.0
86.00	959.0
87.00	460.0
91.00	785.0
92.00	935.0
93.00	5738.0
94.00	415.0
98.00	4349.0
99.00	3339.0
100.00	310.0
101.00	2121.0
102.90	767.0
103.90	1334.0
105.00	1177.0
106.10	469.0
107.00	15316.0
108.00	2456.0
110.00	32248.0
111.00	4591.0
111.90	588.0
116.00	948.0
117.00	11691.0
118.00	941.0
122.00	1060.0
123.00	1675.0
123.90	724.0
125.00	755.0
127.00	60928.0
128.00	4752.0
129.00	22944.0
130.00	1964.0
131.00	406.0
134.00	672.0
135.00	1812.0
136.00	769.0
137.00	929.0
139.90	317.0
141.00	2960.0
141.90	973.0
142.90	665.0
146.00	538.0
147.00	1457.0
148.00	3125.0
149.00	692.0
151.20	427.0
153.00	953.0
154.00	625.0
155.00	1589.0
156.00	2561.0
157.00	478.0
157.90	616.0
158.90	436.0
160.00	890.0
161.00	1457.0
161.90	403.0
165.00	1061.0
166.00	950.0
167.00	5678.0
168.00	2675.0
169.00	495.0
171.90	509.0
173.00	690.0
174.00	1206.0
175.00	2440.0
176.00	821.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

177.00	1081.0
178.00	365.0
179.00	4435.0
180.00	3239.0
181.00	1459.0
183.90	359.0
185.00	2072.0
186.00	17136.0
187.00	4967.0
188.00	519.0
189.00	1050.0
191.00	466.0
192.00	1337.0
193.00	1516.0
194.00	303.0
196.00	4097.0
198.00	131712.0
199.00	9313.0
200.00	723.0
201.50	584.0
203.00	909.0
204.00	4368.0
205.00	7827.0
206.00	33664.0
207.00	4774.0
208.00	1119.0
209.00	390.0
211.10	1361.0
215.00	348.0
216.00	721.0
217.00	8529.0
218.00	1129.0
221.00	6450.0
223.00	2109.0
224.00	18400.0
225.00	4966.0
226.10	525.0
227.00	7161.0
228.00	998.0
229.00	1612.0
231.00	678.0
234.00	475.0
235.00	599.0
236.00	345.0
237.00	581.0
240.90	447.0
242.00	979.0
243.10	1112.0
244.00	14953.0
245.00	2233.0
246.00	2729.0
247.00	547.0
249.00	614.0
252.90	302.0
255.00	80608.0
256.00	12011.0
257.00	946.0
258.00	4174.0
259.00	656.0
265.00	1726.0
273.00	2400.0
274.00	6529.0
275.00	40328.0
276.00	5603.0
277.00	3141.0
278.00	470.0
282.90	332.0
285.00	572.0
293.00	755.0
296.00	9596.0
297.00	1433.0
303.00	1277.0
304.00	321.0
314.00	541.0
314.90	1215.0
316.00	694.0
321.00	373.0
323.00	3478.0
324.00	729.0
326.90	665.0
328.00	350.0
333.00	316.0
334.00	2372.0
335.00	626.0
341.00	405.0
346.00	776.0
352.00	1002.0
353.00	896.0
354.00	1296.0
365.00	4587.0
366.00	721.0
372.00	1980.0
373.00	542.0
383.00	561.0
402.00	771.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

403.00	1167.0
404.00	422.0
421.10	955.0
422.00	953.0
423.00	6792.0
424.00	1477.0
441.10	16664.0
442.10	107424.0
443.10	20952.0
444.10	1873.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

m/z	Abundance
38.00	474.0
39.00	2929.0
44.00	489.0
50.00	9894.0
51.00	39976.0
52.00	2023.0
56.00	1280.0
57.00	2846.0
61.00	519.0
62.00	614.0
63.00	1655.0
65.00	792.0
68.10	900.0
69.00	45824.0
73.00	377.0
74.00	4345.0
75.00	7212.0
76.10	2371.0
77.00	52456.0
78.00	3321.0
79.00	3505.0
80.00	2595.0
81.00	3625.0
82.00	912.0
82.90	892.0
85.00	638.0
85.90	947.0
86.90	442.0
91.00	854.0
92.00	940.0
93.00	5709.0
93.90	398.0
98.00	4422.0
99.00	3445.0
101.00	2090.0
103.00	780.0
104.00	1294.0
105.00	1275.0
107.00	15720.0
108.00	2570.0
110.00	32560.0
111.00	4700.0
112.00	598.0
116.00	1017.0
117.00	11913.0
118.00	917.0
121.90	1031.0
123.00	1685.0
124.00	787.0
125.00	681.0
127.00	63032.0
128.00	4885.0
129.00	23376.0
130.00	1959.0
131.00	362.0
134.00	756.0
135.00	1996.0
136.00	705.0
137.00	938.0
139.90	317.0
141.00	3183.0
142.00	1043.0
142.90	623.0
146.00	562.0
147.00	1591.0
148.00	3468.0
149.00	742.0
151.10	405.0
152.90	931.0
153.90	782.0
155.00	1726.0
156.00	2660.0
157.10	557.0
157.90	569.0
158.90	455.0
159.90	978.0
161.00	1428.0
161.90	462.0
165.00	1189.0
165.90	978.0
167.00	5863.0
168.00	2779.0
169.00	464.0
170.80	317.0
171.90	570.0
173.00	725.0
174.00	1374.0
175.00	2405.0
176.00	733.0
176.90	1218.0
177.90	439.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

178.90	4838.0
180.00	3286.0
181.00	1672.0
183.90	349.0
185.00	2382.0
186.00	19040.0
187.00	5149.0
188.00	504.0
188.90	1103.0
191.00	554.0
192.00	1402.0
193.00	1650.0
194.00	371.0
196.00	4607.0
198.00	144832.0
199.00	9689.0
200.00	795.0
201.40	788.0
203.00	923.0
204.00	4800.0
205.00	8593.0
206.00	36712.0
207.00	4923.0
208.00	1140.0
209.00	377.0
210.20	607.0
211.00	1445.0
214.90	356.0
216.00	726.0
217.00	9263.0
218.00	1246.0
221.00	7355.0
223.00	2097.0
224.00	20256.0
225.00	5309.0
226.00	542.0
227.00	8047.0
228.00	1087.0
229.00	1732.0
231.00	761.0
233.90	549.0
235.00	610.0
236.00	380.0
237.00	730.0
238.90	336.0
241.00	502.0
242.00	1125.0
243.10	1385.0
244.00	18016.0
245.00	2344.0
246.00	3175.0
246.90	652.0
249.00	689.0
252.90	365.0
255.00	91704.0
256.00	13468.0
257.00	1011.0
258.00	4858.0
259.00	809.0
265.00	1953.0
273.00	2880.0
274.00	7605.0
275.00	47152.0
276.00	6254.0
277.00	3421.0
278.00	605.0
283.00	376.0
285.00	703.0
293.00	812.0
296.00	11585.0
297.00	1610.0
303.00	1390.0
303.90	355.0
314.00	665.0
315.00	1429.0
316.00	863.0
320.90	456.0
323.00	4457.0
324.00	775.0
327.00	753.0
328.00	496.0
331.90	325.0
333.00	399.0
334.00	3009.0
335.00	683.0
341.00	460.0
346.00	970.0
352.00	1416.0
353.00	1000.0
354.00	1489.0
365.00	5837.0
366.00	984.0
371.00	355.0
372.00	2464.0
373.00	695.0

Instrument :

BNA_N

ClientSampleId :

DFTPP71

382.90	702.0
390.00	312.0
402.00	1016.0
403.00	1440.0
404.00	625.0
421.00	1153.0
422.00	1177.0
423.00	8748.0
424.00	1817.0
441.10	22608.0
442.10	143296.0
443.10	27168.0
444.10	2511.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

m/z	Abundance
39.00	1475.0
44.00	522.0
50.00	4823.0
51.00	19376.0
52.00	937.0
56.00	619.0
57.00	1491.0
63.00	906.0
65.00	425.0
68.10	483.0
69.00	22512.0
74.00	2452.0
75.00	3606.0
76.00	1171.0
77.00	26312.0
78.00	1633.0
79.00	1685.0
80.00	1274.0
81.00	1770.0
81.90	452.0
83.00	463.0
84.90	303.0
85.90	574.0
91.00	398.0
92.00	473.0
93.00	2936.0
98.00	2333.0
98.90	1696.0
101.00	1106.0
102.90	359.0
103.90	689.0
105.00	648.0
107.00	7817.0
108.00	1294.0
110.00	16408.0
111.00	2308.0
116.00	504.0
117.00	6308.0
118.00	442.0
121.90	546.0
122.90	886.0
123.90	401.0
124.90	354.0
127.00	32536.0
128.00	2427.0
129.00	12308.0
130.00	1008.0
133.90	366.0
134.90	913.0
135.90	438.0
136.90	498.0
141.00	1665.0
142.00	555.0
142.90	355.0
147.00	916.0
147.90	1755.0
148.90	384.0
153.00	499.0
154.00	397.0
154.90	947.0
156.00	1425.0
157.00	345.0
160.00	478.0
160.90	776.0
164.90	648.0
166.00	553.0
167.00	3382.0
168.00	1493.0
172.90	359.0
174.00	769.0
175.00	1408.0
176.00	464.0
176.90	684.0
178.90	2680.0
180.00	1690.0
180.90	925.0
185.00	1206.0
186.00	10472.0
187.00	3061.0
189.00	616.0
191.90	907.0
193.00	965.0
196.00	2744.0
198.00	82040.0
199.00	5318.0
200.00	398.0
201.30	385.0
203.00	532.0
204.00	2679.0
205.00	4753.0
206.00	20584.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

207.10	2826.0
208.00	706.0
210.20	384.0
210.90	881.0
215.90	470.0
217.00	5451.0
218.00	704.0
221.00	4171.0
223.00	1286.0
224.00	11811.0
225.00	3037.0
226.00	312.0
227.00	4848.0
228.00	648.0
229.00	954.0
231.10	469.0
234.00	316.0
235.00	320.0
237.00	381.0
242.00	601.0
243.00	825.0
244.00	10786.0
245.00	1409.0
246.00	1831.0
246.90	373.0
248.90	422.0
255.00	54928.0
256.00	7895.0
257.00	606.0
258.00	3037.0
259.00	447.0
265.00	1245.0
273.00	1849.0
274.00	4628.0
275.00	29536.0
276.00	3755.0
277.00	2213.0
278.00	398.0
285.00	388.0
292.90	567.0
296.00	7467.0
297.00	1027.0
303.00	960.0
314.00	377.0
315.00	913.0
316.00	509.0
323.00	2770.0
324.00	536.0
326.90	536.0
334.00	1814.0
335.00	480.0
341.00	375.0
345.90	613.0
352.00	900.0
353.00	713.0
354.00	1136.0
365.00	4005.0
366.00	597.0
372.00	1739.0
373.00	362.0
382.90	495.0
402.00	731.0
403.00	1008.0
404.00	381.0
421.00	865.0
422.00	824.0
423.00	6664.0
424.00	1210.0
441.10	16656.0
442.10	107584.0
443.10	20040.0
444.00	1853.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71