

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121720\
 Data File : BN013206.D
 Acq On : 18 Dec 2020 05:43
 Operator : CG/JU
 Sample : L4960-05DL 20X
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BES-12SRDL

Quant Time: Dec 18 06:22:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 17 12:34:52 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.338	152	809	0.40	ng	0.00
7) Naphthalene-d8	10.099	136	2579	0.40	ng	-0.01
13) Acenaphthene-d10	14.003	164	1999	0.40	ng	0.00
19) Phenanthrene-d10	16.763	188	3655	0.40	ng	#-0.01
29) Chrysene-d12	21.006	240	3455	0.40	ng	# 0.00
36) Perylene-d12	23.092	264	3839	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	0.000	112	0d	0.00	ng	
5) Phenol-d6	0.000	99	0d	0.00	ng	
8) Nitrobenzene-d5	8.515	82	41	0.02	ng	0.00
11) 2-Methylnaphthalene-d10	0.000	152	0d	0.00	ng	
14) 2,4,6-Tribromophenol	0.000	330	0d	0.00	ng	
15) 2-Fluorobiphenyl	0.000	172	0d	0.00	ng	
27) Fluoranthene-d10	0.000	212	0d	0.00	ng	
31) Terphenyl-d14	19.441	244	152	0.02	ng	0.00
Target Compounds						Qvalue
9) Naphthalene	10.137	128	18862	2.39	ng	97
12) 2-Methylnaphthalene	11.773	142	5027	0.94	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121720\
Data File : BN013206.D
Acq On : 18 Dec 2020 05:43
Operator : CG/JU
Sample : L4960-05DL 20X
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BES-12SRDL

Quant Time: Dec 18 06:22:26 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 17 12:34:52 2020
Response via : Initial Calibration

