

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121720\
 Data File : BN013181.D
 Acq On : 17 Dec 2020 11:25
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 17 14:35:40 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 17 12:34:52 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2	1,4-Dioxane	0.698	0.557	20.2	102	0.00
3	n-Nitrosodimethylamine	1.158	0.851	26.5#	86	0.00
4 S	2-Fluorophenol	1.639	1.420	13.4	96	0.00
5 S	Phenol-d6	1.663	1.447	13.0	111	0.00
6	bis(2-Chloroethyl)ether	1.036	0.983	5.1	137	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
8 S	Nitrobenzene-d5	0.413	0.359	13.1	103	0.00
9	Naphthalene	1.226	1.060	13.5	98	0.00
10	Hexachlorobutadiene	0.262	0.242	7.6	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.705	0.593	15.9	96	0.00
12	2-Methylnaphthalene	0.831	0.666	19.9	92	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.252	0.199	21.0	92	0.00
15 S	2-Fluorobiphenyl	1.677	1.490	11.2	95	0.00
16	Acenaphthylene	2.105	1.801	14.4	92	0.00
17	Acenaphthene	1.333	1.192	10.6	95	0.00
18	Fluorene	1.755	1.529	12.9	95	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
20	4,6-Dinitro-2-methylphenol	0.075	0.042	44.0#	78	0.00
21	4-Bromophenyl-phenylether	0.078	0.074	5.1	109	0.00
22	Hexachlorobenzene	0.275	0.228	17.1	85	0.00
23	Atrazine	0.239	0.184	23.0	89	0.00
24	Pentachlorophenol	0.116	0.084	27.6#	85	0.00
25	Phenanthrene	1.301	1.086	16.5	93	0.00
26	Anthracene	1.154	0.907	21.4	90	0.00
27 SURR	Fluoranthene-d10	1.316	1.094	16.9	92	0.00
28	Fluoranthene	1.586	1.317	17.0	91	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
30	Pyrene	1.557	1.363	12.5	90	0.00
31 S	Terphenyl-d14	1.016	0.872	14.2	87	0.00
32	Benzo(a)anthracene	1.406	1.172	16.6	86	0.00
33	Chrysene	1.493	1.316	11.9	92	0.00
34	Bis(2-ethylhexyl)phthalate	1.445	0.625	56.7#	61	0.00
35	Indeno(1,2,3-cd)pyrene	2.061	1.768	14.2	80	0.00
36 I	Perylene-d12	1.000	1.000	0.0	92	0.00
37	Benzo(b)fluoranthene	1.412	1.207	14.5	80	0.00
38	Benzo(k)fluoranthene	1.613	1.447	10.3	88	0.00
39 C	Benzo(a)pyrene	1.403	1.213	13.5	80	0.00
40	Dibenzo(a,h)anthracene	1.521	1.341	11.8	82	0.00
41	Benzo(g,h,i)perylene	1.621	1.498	7.6	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4