

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121819\
 Data File : BN009020.D
 Acq On : 18 Dec 2019 18:16
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 19 01:02:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 19 00:50:46 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	141	0.00
2	1,4-Dioxane	0.375	0.465	-24.0	148	0.00
3	n-Nitrosodimethylamine	0.518	0.585	-12.9	165#	0.00
4 S	2-Fluorophenol	1.017	1.086	-6.8	140	0.00
5 S	Phenol-d6	1.406	1.315	6.5	124	0.00
6	bis(2-Chloroethyl)ether	1.325	1.333	-0.6	128	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	124	0.00
8 S	Nitrobenzene-d5	0.339	0.343	-1.2	121	0.00
9	Naphthalene	1.091	1.147	-5.1	116	0.00
10	Hexachlorobutadiene	0.209	0.237	-13.4	125	0.00
11 SURR	2-Methylnaphthalene-d10	0.684	0.675	1.3	109	0.00
12	2-Methylnaphthalene	0.813	0.811	0.2	109	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.192	0.189	1.6	101	0.00
15 S	2-Fluorobiphenyl	1.519	1.606	-5.7	106	0.00
16	Acenaphthylene	1.830	1.910	-4.4	101	0.00
17	Acenaphthene	1.215	1.281	-5.4	100	0.00
18	Fluorene	1.634	1.726	-5.6	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
20	4-Bromophenyl-phenylether	0.239	0.255	-6.7	103	0.00
21	Hexachlorobenzene	0.273	0.283	-3.7	100	0.00
22	Pentachlorophenol	0.098	0.082	16.3	90	0.00
23	Phenanthrene	1.259	1.280	-1.7	97	0.00
24	Anthracene	1.109	1.123	-1.3	99	0.00
25 SURR	Fluoranthene-d10	1.211	1.191	1.7	97	0.00
26	Fluoranthene	1.498	1.502	-0.3	96	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
28	Pvrene	1.402	1.672	-19.3	96	0.00
29 S	Terphenyl-d14	0.913	1.061	-16.2	98	0.00
30	Benzo(a)anthracene	1.418	1.526	-7.6	89	0.00
31	Chrysene	1.470	1.536	-4.5	84	0.00
32	Bis(2-ethylhexyl)phthalate	0.664	0.798	-20.2	112	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	1.436	7.7	83	0.00
34 I	Perylene-d12	1.000	1.000	0.0	87	0.00
35	Benzo(b)fluoranthene	1.432	1.529	-6.8	79	0.00
36	Benzo(k)fluoranthene	1.442	1.508	-4.6	79	0.00
37 C	Benzo(a)pyrene	1.286	1.360	-5.8	81	0.00
38	Dibenzo(a,h)anthracene	1.206	1.245	-3.2	82	0.00
39	Benzo(a,h,i)perylene	1.191	1.258	-5.6	82	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121819\
Data File : BN009020.D
Acq On : 18 Dec 2019 18:16
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 19 01:02:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 19 00:50:46 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			