

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121819\
 Data File : BN009027.D
 Acq On : 18 Dec 2019 22:27
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 19 01:30:12 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 19 00:50:46 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	131	0.00
2	1,4-Dioxane	0.375	0.487	-29.9#	144	0.00
3	n-Nitrosodimethylamine	0.518	0.579	-11.8	151#	0.00
4 S	2-Fluorophenol	1.017	1.241	-22.0	148	0.00
5 S	Phenol-d6	1.406	1.526	-8.5	134	0.00
6	bis(2-Chloroethyl)ether	1.325	1.348	-1.7	120	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
8 S	Nitrobenzene-d5	0.339	0.341	-0.6	117	0.01
9	Naphthalene	1.091	1.145	-4.9	113	0.00
10	Hexachlorobutadiene	0.209	0.239	-14.4	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.684	0.684	0.0	108	0.00
12	2-Methylnaphthalene	0.813	0.824	-1.4	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.192	0.182	5.2	101	0.00
15 S	2-Fluorobiphenyl	1.519	1.514	0.3	103	0.00
16	Acenaphthylene	1.830	1.965	-7.4	107	0.00
17	Acenaphthene	1.215	1.270	-4.5	103	0.00
18	Fluorene	1.634	1.701	-4.1	101	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
20	4-Bromophenyl-phenylether	0.239	0.244	-2.1	103	0.00
21	Hexachlorobenzene	0.273	0.287	-5.1	106	0.00
22	Pentachlorophenol	0.098	0.081	17.3	93	0.00
23	Phenanthrene	1.259	1.266	-0.6	101	0.00
24	Anthracene	1.109	1.160	-4.6	107	0.00
25 SURR	Fluoranthene-d10	1.211	1.223	-1.0	104	0.00
26	Fluoranthene	1.498	1.504	-0.4	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
28	Pvrene	1.402	1.826	-30.2#	101	0.00
29 S	Terphenyl-d14	0.913	1.158	-26.8#	103	0.00
30	Benzo(a)anthracene	1.418	1.580	-11.4	89	0.00
31	Chrysene	1.470	1.536	-4.5	81	0.00
32	Bis(2-ethylhexyl)phthalate	0.664	1.141	-71.8#	155#	0.00
33	Indeno(1,2,3-cd)pyrene	1.556	1.371	11.9	76	0.00
34 I	Perylene-d12	1.000	1.000	0.0	79	0.00
35	Benzo(b)fluoranthene	1.432	1.553	-8.4	74	0.00
36	Benzo(k)fluoranthene	1.442	1.480	-2.6	72	0.00
37 C	Benzo(a)pyrene	1.286	1.377	-7.1	75	0.00
38	Dibenzo(a,h)anthracene	1.206	1.254	-4.0	76	0.00
39	Benzo(a,h,i)perylene	1.191	1.271	-6.7	77	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN121819\
Data File : BN009027.D
Acq On : 18 Dec 2019 22:27
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 19 01:30:12 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN120919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 19 00:50:46 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			