

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121820\
 Data File : BN013210.D
 Acq On : 18 Dec 2020 14:53
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 19 01:13:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 18 16:11:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	84	0.00
2	1,4-Dioxane	0.400	0.474	-18.5	99	0.00
3	n-Nitrosodimethylamine	0.400	0.421	-5.2	81	0.00
4 S	2-Fluorophenol	0.400	0.397	0.8	72	0.00
5 S	Phenol-d6	0.400	0.383	4.3	80	0.00
6	bis(2-Chloroethyl)ether	0.400	0.309	22.8	73	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	65	0.00
8 S	Nitrobenzene-d5	0.400	0.429	-7.2	73	0.00
9	Naphthalene	0.400	0.397	0.8	64	0.00
10	Hexachlorobutadiene	0.400	0.473	-18.2	76	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.407	-1.7	67	0.00
12	2-Methylnaphthalene	0.400	0.364	9.0	60	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	64	0.00
14 S	2,4,6-Tribromophenol	0.400	0.404	-1.0	71	0.00
15 S	2-Fluorobiphenyl	0.400	0.420	-5.0	68	0.00
16	Acenaphthylene	0.400	0.403	-0.8	65	0.00
17	Acenaphthene	0.400	0.405	-1.3	65	0.00
18	Fluorene	0.400	0.403	-0.8	67	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	67	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.367	8.3	58	0.00
21	4-Bromophenyl-phenylether	0.400	0.520	-30.0#	95	0.00
22	Hexachlorobenzene	0.400	0.389	2.8	63	0.00
23	Atrazine	0.400	0.352	12.0	64	0.00
24	Pentachlorophenol	0.400	0.345	13.8	64	0.00
25	Phenanthrene	0.400	0.380	5.0	67	0.00
26	Anthracene	0.400	0.360	10.0	65	0.00
27 SURR	Fluoranthene-d10	0.400	0.376	6.0	65	0.00
28	Fluoranthene	0.400	0.381	4.8	66	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	65	0.00
30	Pyrene	0.400	0.400	0.0	66	0.00
31 S	Terphenyl-d14	0.400	0.405	-1.3	66	0.00
32	Benzo(a)anthracene	0.400	0.341	14.8	56	0.00
33	Chrysene	0.400	0.453	-13.2	76	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.116	71.0#	43	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.383	4.3	57	0.00
36 I	Perylene-d12	0.400	0.400	0.0	60	0.00
37	Benzo(b)fluoranthene	0.400	0.412	-3.0	62	0.00
38	Benzo(k)fluoranthene	0.400	0.432	-8.0	68	0.00
39 C	Benzo(a)pyrene	0.400	0.396	1.0	59	0.00
40	Dibenzo(a,h)anthracene	0.400	0.383	4.3	58	0.00
41	Benzo(g,h,i)perylene	0.400	0.418	-4.5	62	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4