

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121820\
 Data File : BN013221.D
 Acq On : 18 Dec 2020 22:42
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 19 01:12:24 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 18 16:11:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	89	0.00
2	1,4-Dioxane	0.400	0.391	2.3	86	0.00
3	n-Nitrosodimethylamine	0.400	0.386	3.5	78	0.00
4 S	2-Fluorophenol	0.400	0.382	4.5	73	0.00
5 S	Phenol-d6	0.400	0.383	4.3	85	0.00
6	bis(2-Chloroethyl)ether	0.400	0.325	18.8	81	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	73	-0.01
8 S	Nitrobenzene-d5	0.400	0.416	-4.0	80	-0.01
9	Naphthalene	0.400	0.394	1.5	72	-0.01
10	Hexachlorobutadiene	0.400	0.471	-17.7	85	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.385	3.8	71	0.00
12	2-Methylnaphthalene	0.400	0.365	8.8	68	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	65	0.00
14 S	2,4,6-Tribromophenol	0.400	0.416	-4.0	74	-0.01
15 S	2-Fluorobiphenyl	0.400	0.470	-17.5	77	-0.01
16	Acenaphthylene	0.400	0.403	-0.8	66	0.00
17	Acenaphthene	0.400	0.412	-3.0	68	0.00
18	Fluorene	0.400	0.409	-2.2	69	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	68	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.398	0.5	69	-0.01
21	4-Bromophenyl-phenylether	0.400	0.487	-21.7	90	0.00
22	Hexachlorobenzene	0.400	0.379	5.3	62	0.00
23	Atrazine	0.400	0.357	10.8	66	-0.01
24	Pentachlorophenol	0.400	0.350	12.5	65	0.00
25	Phenanthrene	0.400	0.387	3.3	69	0.00
26	Anthracene	0.400	0.373	6.8	69	-0.01
27 SURR	Fluoranthene-d10	0.400	0.390	2.5	69	0.00
28	Fluoranthene	0.400	0.392	2.0	69	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	67	-0.01
30	Pyrene	0.400	0.404	-1.0	69	0.00
31 S	Terphenyl-d14	0.400	0.408	-2.0	69	0.00
32	Benzo(a)anthracene	0.400	0.379	5.3	65	-0.01
33	Chrysene	0.400	0.421	-5.2	73	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.116	71.0#	44	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.410	-2.5	64	0.00
36 I	Perylene-d12	0.400	0.400	0.0	62	-0.01
37	Benzo(b)fluoranthene	0.400	0.442	-10.5	70	-0.01
38	Benzo(k)fluoranthene	0.400	0.419	-4.7	70	0.00
39 C	Benzo(a)pyrene	0.400	0.440	-10.0	68	0.00
40	Dibenzo(a,h)anthracene	0.400	0.411	-2.7	65	-0.02
41	Benzo(g,h,i)perylene	0.400	0.428	-7.0	66	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4