

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121820\
 Data File : BN013210.D
 Acq On : 18 Dec 2020 14:53
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Dec 19 01:13:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 18 16:11:20 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.346	152	549	0.40	ng	0.00
7) Naphthalene-d8	10.112	136	1509	0.40	ng	# 0.00
13) Acenaphthene-d10	14.004	164	896	0.40	ng	0.00
19) Phenanthrene-d10	16.775	188	2027	0.40	ng	# 0.00
29) Chrysene-d12	21.006	240	1989	0.40	ng	# 0.00
36) Perylene-d12	23.096	264	2221	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.962	112	893	0.40	ng	0.00
5) Phenol-d6	6.549	99	873	0.38	ng	0.00
8) Nitrobenzene-d5	8.528	82	668	0.43	ng	0.00
11) 2-Methylnaphthalene-d10	11.724	152	1082	0.41	ng	0.00
14) 2,4,6-Tribromophenol	15.539	330	228	0.40	ng	0.00
15) 2-Fluorobiphenyl	12.633	172	1577	0.42	ng	0.00
27) Fluoranthene-d10	18.821	212	2506	0.38	ng	0.00
31) Terphenyl-d14	19.446	244	2046	0.41	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.933	88	454	0.47	ng	96
3) n-Nitrosodimethylamine	3.279	42	669	0.42	ng	# 94
6) bis(2-Chloroethyl)ether	6.798	93	439	0.31	ng	# 83
9) Naphthalene	10.163	128	1836	0.40	ng	# 87
10) Hexachlorobutadiene	10.416	225	468	0.47	ng	# 100
12) 2-Methylnaphthalene	11.804	142	1141	0.36	ng	# 95
16) Acenaphthylene	13.724	152	1899	0.40	ng	100
17) Acenaphthene	14.067	154	1208	0.40	ng	87
18) Fluorene	15.069	166	1586	0.40	ng	98
20) 4,6-Dinitro-2-methylph...	15.260	198	101	0.37	ng	# 1
21) 4-Bromophenyl-phenylether	16.021	248	206	0.52	ng	# 92
22) Hexachlorobenzene	16.082	284	543	0.39	ng	94
23) Atrazine	16.276	200	426	0.35	ng	# 85
24) Pentachlorophenol	16.447	266	202	0.35	ng	92
25) Phenanthrene	16.812	178	2509	0.38	ng	99
26) Anthracene	16.922	178	2104	0.36	ng	100
28) Fluoranthene	18.856	202	3059	0.38	ng	99
30) Pyrene	19.223	202	3097	0.40	ng	99
32) Benzo(a)anthracene	20.995	228	2385	0.34	ng	96
33) Chrysene	21.038	228	3360	0.45	ng	98
34) Bis(2-ethylhexyl)phtha...	20.921	149	1358	0.12	ng	96
35) Indeno(1,2,3-cd)pyrene	25.132	276	3923	0.38	ng	94
37) Benzo(b)fluoranthene	22.479	252	3234	0.41	ng	# 89
38) Benzo(k)fluoranthene	22.517	252	3867	0.43	ng	# 88
39) Benzo(a)pyrene	23.007	252	3088	0.40	ng	# 83
40) Dibenzo(a,h)anthracene	25.153	278	3239	0.38	ng	# 83
41) Benzo(g,h,i)perylene	25.748	276	3762	0.42	ng	# 91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121820\
Data File : BN013210.D
Acq On : 18 Dec 2020 14:53
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Dec 19 01:13:28 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 18 16:11:20 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.346 min Scan# 567
Delta R.T. -0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 549
Ion Ratio Lower Upper
152 100
150 141.9 116.6 174.8
115 72.9 52.9 79.3



#2
1,4-Dioxane
Concen: 0.47 ng
RT: 2.933 min Scan# 19
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion: 88 Resp: 454
Ion Ratio Lower Upper
88 100
58 74.4 59.4 89.2
43 33.7 32.3 48.5









#3
n-Nitrosodimethylamine
Concen: 0.42 ng
RT: 3.279 min Scan# 62
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 669
Ion Ratio Lower Upper
42 100
74 85.7 64.4 96.6
44 0.0 2.8 4.2#



#4
2-Fluorophenol
Concen: 0.40 ng
RT: 4.962 min Scan# 271
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion: 112 Resp: 893
Ion Ratio Lower Upper
112 100
64 66.5 49.5 74.3
63 39.6 32.0 48.0









#5
Phenol-d6
Concen: 0.38 ng
RT: 6.549 min Scan# 468
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion: 99 Resp: 873




Ion	Ratio	Lower	Upper
99	100		
42	33.2	22.4	33.6
71	45.4	37.8	56.6




#6
bis(2-Chloroethyl)ether
Concen: 0.31 ng
RT: 6.798 min Scan# 499
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion: 93 Resp: 439

Ion	Ratio	Lower	Upper
93	100		
63	96.1	63.1	94.7#
95	39.2	26.1	39.1#









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.112 min Scan# 829
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	1509
Ion	Ratio	Lower	Upper
136	100		
137	19.4	10.6	15.8#
54	16.2	8.6	12.8#
68	12.7	5.7	8.5#



#8
Nitrobenzene-d5
Concen: 0.43 ng
RT: 8.528 min Scan# 704
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:	82	Resp:	668
Ion	Ratio	Lower	Upper
82	100		
128	52.2	30.6	46.0#
54	73.6	48.3	72.5#









#9
Naphthalene
Concen: 0.40 ng
RT: 10.163 min Scan# 833
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 1836
Ion Ratio Lower Upper
128 100
129 18.8 9.8 14.8#
127 19.3 12.2 18.4#



#10
Hexachlorobutadiene
Concen: 0.47 ng
RT: 10.416 min Scan# 853
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:225 Resp: 468
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.5 51.0 76.6









#11
2-Methylnaphthalene-d10
Concen: 0.41 ng
RT: 11.724 min Scan# 988
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 1082
Ion Ratio Lower Upper
152 100
151 18.4 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.36 ng
RT: 11.804 min Scan# 1006
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:142 Resp: 1141
Ion Ratio Lower Upper
142 100
141 87.5 69.4 104.2
115 53.5 35.7 53.5#









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.004 min Scan# 1282
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:164 Resp: 896




Ion	Ratio	Lower	Upper
164	100		
162	107.8	80.6	120.8
160	56.5	39.5	59.3




#14
2,4,6-Tribromophenol
Concen: 0.40 ng
RT: 15.539 min Scan# 1403
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:330 Resp: 228

Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	44.7	34.4	51.6









#15
2-Fluorobiphenyl
Concen: 0.42 ng
RT: 12.633 min Scan# 1174
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	1577
Ion	Ratio	Lower	Upper
172	100		
171	38.2	28.2	42.4
170	30.6	20.0	30.0#



#16
Acenaphthylene
Concen: 0.40 ng
RT: 13.724 min Scan# 1260
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:	152	Resp:	1899
Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.8	23.8
153	13.4	10.6	16.0









#17
Acenaphthene
Concen: 0.40 ng
RT: 14.067 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 1208
Ion Ratio Lower Upper
154 100
153 119.3 83.6 125.4
152 61.7 43.9 65.9



#18
Fluorene
Concen: 0.40 ng
RT: 15.069 min Scan# 1366
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:166 Resp: 1586
Ion Ratio Lower Upper
166 100
165 100.1 78.6 118.0
167 13.4 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 16.775 min Scan# 1503
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	188	Resp:	2027
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	11.7	5.0	7.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.37 ng
RT: 15.260 min Scan# 1381
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:	198	Resp:	101
Ion	Ratio	Lower	Upper
198	100		
51	150.0	43.5	65.3#
105	98.5	27.2	40.8#









#21
4-Bromophenyl-phenylether
Concen: 0.52 ng
RT: 16.021 min Scan# 1441
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:248 Resp: 206




Ion	Ratio	Lower	Upper
248	100		
250	96.4	77.3	115.9
141	86.9	57.8	86.8#




#22
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.082 min Scan# 1446
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:284 Resp: 543

Ion	Ratio	Lower	Upper
284	100		
142	35.0	32.0	48.0
249	32.2	27.0	40.6









#23
Atrazine
Concen: 0.35 ng
RT: 16.276 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	200	Resp:	426
Ion	Ratio	Lower	Upper
200	100		
173	38.8	22.8	34.2#
215	55.3	38.1	57.1



#24
Pentachlorophenol
Concen: 0.35 ng
RT: 16.447 min Scan# 1476
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:	266	Resp:	202
Ion	Ratio	Lower	Upper
266	100		
264	64.9	50.9	76.3
268	75.2	51.5	77.3









#25
Phenanthrene
Concen: 0.38 ng
RT: 16.812 min Scan# 1506
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	178	Resp:	2509
Ion	Ratio	Lower	Upper
178	100		
176	19.0	14.8	22.2
179	15.5	12.4	18.6



#26
Anthracene
Concen: 0.36 ng
RT: 16.922 min Scan# 1515
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:	178	Resp:	2104
Ion	Ratio	Lower	Upper
178	100		
176	18.5	14.8	22.2
179	15.9	12.5	18.7









#27
Fluoranthene-d10
Concen: 0.38 ng
RT: 18.821 min Scan# 1767
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 2506
Ion Ratio Lower Upper
212 100
106 7.6 6.8 10.2
104 4.6 4.1 6.1



#28
Fluoranthene
Concen: 0.38 ng
RT: 18.856 min Scan# 1774
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:202 Resp: 3059
Ion Ratio Lower Upper
202 100
101 6.7 5.8 8.6
203 17.7 13.7 20.5









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.006 min Scan# 2100
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 1989
Ion Ratio Lower Upper
240 100
120 13.3 5.3 7.9#
236 31.7 21.8 32.8



#30
Pyrene
Concen: 0.40 ng
RT: 19.223 min Scan# 1848
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:202 Resp: 3097
Ion Ratio Lower Upper
202 100
200 20.5 16.6 24.8
203 17.5 13.8 20.8









#31
Terphenyl-d14
Concen: 0.41 ng
RT: 19.446 min Scan# 1893
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 2046
Ion Ratio Lower Upper
244 100
212 13.1 7.3 10.9#
122 14.0 7.3 10.9#



#32
Benzo(a)anthracene
Concen: 0.34 ng
RT: 20.995 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:228 Resp: 2385
Ion Ratio Lower Upper
228 100
226 29.0 22.3 33.5
229 23.2 15.7 23.5









#33
Chrysene
Concen: 0.45 ng
RT: 21.038 min Scan# 2103
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	228	Resp:	3360
Ion	Ratio	Lower	Upper
228	100		
226	31.2	24.2	36.2
229	20.7	15.7	23.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.12 ng
RT: 20.921 min Scan# 2092
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:	149	Resp:	1358
Ion	Ratio	Lower	Upper
149	100		
167	26.2	23.2	34.8
279	4.9	3.9	5.9









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.38 ng
RT: 25.132 min Scan# 3028
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:276 Resp: 3923
Ion Ratio Lower Upper
276 100
138 13.6 13.3 19.9
277 22.8 20.2 30.4



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.096 min Scan# 2433
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:264 Resp: 2221
Ion Ratio Lower Upper
264 100
260 27.9 19.8 29.6
265 94.6 51.4 77.0#









#37
Benzo(b)fluoranthene
Concen: 0.41 ng
RT: 22.479 min Scan# 2253
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 3234
Ion Ratio Lower Upper
252 100
253 31.1 20.1 30.1#
125 12.0 7.2 10.8#



#38
Benzo(k)fluoranthene
Concen: 0.43 ng
RT: 22.517 min Scan# 2264
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Tgt Ion:252 Resp: 3867
Ion Ratio Lower Upper
252 100
253 31.3 19.8 29.8#
125 11.7 7.1 10.7#





#39

Benzo(a)pyrene

Concen: 0.40 ng

RT: 23.007 min Scan# 2407

Delta R.T. 0.000 min

Lab File: BN013210.D

Acq: 18 Dec 2020 14:53





Tgt Ion:252 Resp: 3088

Ion Ratio Lower Upper

252 100

253 35.9 21.3 31.9#

125 15.7 8.3 12.5#



#40

Dibenzo(a,h)anthracene

Concen: 0.38 ng

RT: 25.153 min Scan# 3034

Delta R.T. 0.000 min

Lab File: BN013210.D

Acq: 18 Dec 2020 14:53




Tgt Ion:278 Resp: 3239


Ion Ratio Lower Upper

278 100

139 17.2 9.3 13.9#

279 36.3 21.6 32.4#





#41
Benzo(g,h,i)perylene
Concen: 0.42 ng
RT: 25.748 min Scan# 3208
Delta R.T. 0.000 min
Lab File: BN013210.D
Acq: 18 Dec 2020 14:53

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	276	Resp:	3762
Ion	Ratio	Lower	Upper
276	100		
277	28.8	19.8	29.8
138	18.8	11.7	17.5#

