

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN121918\
Quantitation Report (Not Reviewed)

Data File : BN004113.D
Acq On : 19 Dec 2018 14:56
Operator : JU/SJ
Sample : J6297-06
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9E41

Quant Time: Dec 20 04:45:56 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 20 04:42:11 2018
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.83	152	1424	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	7407	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	4742	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.31	188	667442	0.40	ng/ul	0.10
16) Chrysene-d12	0.00	240	0	0.00	ng/ul	-21.40
20) Perylene-d12	23.72	264	12382	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	2626	0.23	ng/ul	0.00
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	
Target Compounds						
23) Benzo(a)pyrene	23.57	252	2705	0.061	ng/ul#	Qvalue 60




(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN004113.D
Acq On : 19 Dec 2018 14:56
Operator : JU/SJ
Sample : J6297-06
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9E41

Quant Time: Dec 20 04:45:56 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 20 04:42:11 2018
Response via : Initial Calibration





#23

Benzo(a)pyrene

Concen: 0.061 ng/ul

RT: 23.57 min Scan# 5493

Delta R.T. -0.05 min

Lab File: BN004113.D

Acq: 19 Dec 2018 14:56

Instrument :

BNA_N

ClientSampleId :

F9E41

Tgt Ion: 252 Resp: 2705

Ion Ratio Lower Upper

252 100

253 40.6 19.0 28.4#

125 33.1 10.2 15.4#

