

Data File : BN004116.D
Acq On : 19 Dec 2018 16:42
Operator : JU/SJ
Sample : J6297-08
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Dec 20 06:00:13 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 20 05:49:27 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.83	152	1286	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	6788	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	4347	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	11483m	0.40	ng/ul	0.00
16) Chrysene-d12	21.40	240	12313m	0.40	ng/ul	0.00
20) Perylene-d12	23.72	264	11060m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	2511	0.24	ng/ul	0.00
14) Fluoranthene-d10	19.24	212	8444m	0.25	ng/ul	0.00
Target Compounds				Qvalue		
7) Acenaphthylene	14.19	152	411	0.021	ng/ul#	60
15) Fluoranthene	19.27	202	1069m	0.024	ng/ul	
17) Pyrene	19.64	202	1169	0.026	ng/ul#	80
18) Benzo(a)anthracene	21.39	228	984	0.025	ng/ul#	82
19) Chrysene	21.44	228	1161	0.026	ng/ul#	82
21) Benzo(b)fluoranthene	23.02	252	1670m	0.037	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN004116.D
Acq On : 19 Dec 2018 16:42
Operator : JU/SJ
Sample : J6297-08
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Dec 20 06:00:13 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 20 05:49:27 2018
Response via : Initial Calibration

