

Data File : BN004125.D
Acq On : 19 Dec 2018 21:59
Operator : JU/SJ
Sample : PB115526BL
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK26

Quant Time: Dec 20 05:53:45 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 20 05:49:27 2018
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.83	152	1311	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	6951	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	4395	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.31	188	682639	0.40	ng/ul	0.10
16) Chrysene-d12	0.00	240	0	0.00	ng/ul	-21.40
20) Perylene-d12	23.72	264	12695	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	2613	0.24	ng/ul	0.00
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	
Target Compounds						
9) Fluorene	15.46	166	1058	0.052	ng/ul#	16
23) Benzo(a)pyrene	23.57	252	2802	0.062	ng/ul#	43


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN004125.D
Acq On : 19 Dec 2018 21:59
Operator : JU/SJ
Sample : PB115526BL
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK26

Quant Time: Dec 20 05:53:45 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 20 05:49:27 2018
Response via : Initial Calibration









#9
Fluorene
 Concen: 0.052 ng/ul
 RT: 15.46 min Scan# 3314
 Delta R.T. -0.06 min
 Lab File: BN004125.D
 Acq: 19 Dec 2018 21:59

Instrument :
 BNA_N
 ClientSampleId :
 SBLK26

Tgt Ion:166 Resp: 1058
 Ion Ratio Lower Upper
 166 100
 165 11.0 78.3 117.5#
 167 37.1 11.3 16.9#



#23
Benzo(a)pyrene
 Concen: 0.062 ng/ul
 RT: 23.57 min Scan# 5494
 Delta R.T. -0.05 min
 Lab File: BN004125.D
 Acq: 19 Dec 2018 21:59

Tgt Ion:252 Resp: 2802
 Ion Ratio Lower Upper
 252 100
 253 44.5 19.0 28.4#
 125 46.0 10.2 15.4#

