

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN121918\
Quantitation Report (Not Reviewed)

Data File : BN004130.D
Acq On : 20 Dec 2018 00:55
Operator : JU/SJ
Sample : J6299-05
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9F48

Quant Time: Dec 20 05:54:23 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 20 05:49:27 2018
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.83	152	1324	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	7089	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	4530	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.31	188	547637	0.40	ng/ul	0.10
16) Chrysene-d12	0.00	240	0	0.00	ng/ul	-21.40
20) Perylene-d12	23.71	264	13078	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	2155	0.20	ng/ul	0.00
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	
Target Compounds						
23) Benzo(a)pyrene	23.57	252	2223	0.047	ng/ul#	Qvalue 44


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN004130.D
Acq On : 20 Dec 2018 00:55
Operator : JU/SJ
Sample : J6299-05
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9F48

Quant Time: Dec 20 05:54:23 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN11918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 20 05:49:27 2018
Response via : Initial Calibration





#23

Benzo(a)pyrene

Concen: 0.047 ng/ul

RT: 23.57 min Scan# 5491

Delta R.T. -0.05 min

Lab File: BN004130.D


Acq: 20 Dec 2018 00:55

Instrument :

BNA_N

ClientSampleId :

F9F48



Tgt Ion: 252 Resp: 2223

Ion Ratio Lower Upper

252 100

253 46.3 19.0 28.4#

125 42.1 10.2 15.4#

