

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN121918\
Quantitation Report (Not Reviewed)

Data File : BN004133.D
Acq On : 20 Dec 2018 02:40
Operator : JU/SJ
Sample : J6299-07
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9F50

Quant Time: Dec 20 07:20:24 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 20 07:15:48 2018
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.83	152	1353	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	7196	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	4584	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.31	188	563801	0.40	ng/ul	0.09
16) Chrysene-d12	0.00	240	0	0.00	ng/ul	-21.40
20) Perylene-d12	23.71	264	13386	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.22	152	2198	0.20	ng/ul	0.00
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	
Target Compounds						
23) Benzo(a)pyrene	23.57	252	2368	0.049	ng/ul#	Qvalue 53


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN004133.D
Acq On : 20 Dec 2018 02:40
Operator : JU/SJ
Sample : J6299-07
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9F50

Quant Time: Dec 20 07:20:24 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN11918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 20 07:15:48 2018
Response via : Initial Calibration





#23

Benzo(a)pyrene

Concen: 0.049 ng/ul

RT: 23.57 min Scan# 5492

Delta R.T. -0.05 min

Lab File: BN004133.D


Acq: 20 Dec 2018 02:40

Instrument :

BNA_N

ClientSampleId :

F9F50



Tgt Ion: 252 Resp: 2368

Ion Ratio Lower Upper

252 100

253 42.6 19.0 28.4#

125 37.3 10.2 15.4#

