


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN121920\
 Data File : BN013223.D
 Acq On : 19 Dec 2020 09:04
 Operator : CG/JU
 Sample : DFTPP31
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP31

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN120320.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Dec 21 10:25:14 2020



AutoFind: Scans 2503, 2504, 2505; Background Corrected with Scan 2497

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	61.3	73202	PASS
68	69	0.00	2	2.2	1437	FAIL*
69	198	0.00	100	53.7	64112	PASS
70	69	0.00	2	0.5	308	PASS
127	198	10	80	51.0	60914	PASS
197	198	0.00	2	0.4	445	PASS
198	198	100	100	100.0	119357	PASS
199	198	5	9	6.8	8155	PASS
275	198	10	60	28.4	33944	PASS
365	198	1	100	4.6	5467	PASS
441	198	0.01	100	12.6	14983	PASS
442	198	50	100	73.4	87592	PASS
443	442	15	24	20.0	17547	PASS

Scan 2497 (17.175 min): BN013223.D
DFTPP31

m/z	Abundance
40.10	365.0
43.90	411.0
73.00	1160.0
146.90	394.0
206.90	1048.0
207.90	467.0
221.10	447.0
281.00	434.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

m/z	Abundance
36.90	453.0
38.00	869.0
39.10	5182.0
40.00	463.0
44.00	913.0
45.00	326.0
49.10	1102.0
50.10	17976.0
51.10	60840.0
52.00	3706.0
54.90	445.0
56.00	1537.0
57.00	3247.0
58.10	365.0
61.10	776.0
62.10	1137.0
63.10	2242.0
64.00	319.0
65.00	1294.0
67.90	1138.0
69.00	53784.0
73.00	1187.0
74.00	8055.0
75.00	10423.0
76.00	3907.0
77.10	58352.0
78.10	3732.0
79.00	5278.0
79.90	4136.0
81.00	5150.0
81.90	1243.0
83.00	1083.0
85.00	611.0
86.10	1327.0
87.10	871.0
90.90	1272.0
92.00	1190.0
93.00	9765.0
94.00	639.0
95.90	332.0
97.10	390.0
98.00	8165.0
99.00	5442.0
99.80	448.0
101.00	2004.0
102.90	784.0
104.00	1987.0
105.00	1418.0
106.10	416.0
107.00	17424.0
108.00	2231.0
109.10	398.0
110.00	25704.0
110.90	5275.0
112.00	948.0
113.00	356.0
116.10	1026.0
117.00	26952.0
118.00	2393.0
122.10	1445.0
123.00	1638.0
124.00	1208.0
125.00	636.0
126.10	443.0
127.00	48752.0
128.00	3992.0
129.00	23944.0
130.00	2003.0
131.00	414.0
133.90	1022.0
134.90	1874.0
136.00	820.0
137.00	1370.0
141.00	3594.0
142.00	1001.0
143.00	693.0
145.90	557.0
147.00	1721.0
147.90	5130.0
149.00	1020.0
150.10	513.0
150.90	381.0
152.90	803.0
153.90	535.0
155.00	1731.0
156.00	1995.0
157.10	640.0
158.00	476.0
159.00	757.0
159.90	1191.0
160.90	1338.0
161.90	486.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

165.00	976.0
165.90	1228.0
167.00	7991.0
168.00	4802.0
168.90	590.0
171.00	383.0
171.90	488.0
173.00	571.0
174.00	1425.0
175.00	2134.0
175.90	411.0
176.70	944.0
177.20	854.0
179.00	4209.0
180.00	3869.0
181.00	1289.0
183.90	310.0
185.00	2736.0
186.00	14191.0
187.00	4961.0
188.10	382.0
189.00	1492.0
190.90	575.0
192.00	1391.0
193.00	1769.0
193.90	310.0
195.10	307.0
195.90	2434.0
196.70	419.0
197.00	458.0
198.00	84416.0
199.00	6066.0
200.00	531.0
203.00	1160.0
204.00	3821.0
205.00	6598.0
206.00	25312.0
207.10	4095.0
208.00	1005.0
209.00	601.0
209.90	525.0
211.10	965.0
215.00	449.0
215.90	415.0
217.00	7373.0
217.80	782.0
221.10	2805.0
222.00	1270.0
223.00	1337.0
224.00	13340.0
225.10	3289.0
225.90	516.0
227.00	7657.0
228.00	1132.0
229.10	1402.0
231.10	441.0
234.00	837.0
235.00	377.0
236.00	303.0
236.90	836.0
239.90	344.0
240.90	547.0
242.10	766.0
243.00	758.0
244.10	8727.0
245.10	1559.0
246.00	2511.0
247.10	486.0
249.00	467.0
253.10	317.0
254.00	425.0
255.00	50296.0
256.00	9297.0
257.00	826.0
258.00	3590.0
258.90	675.0
265.00	1518.0
265.90	316.0
273.00	1735.0
274.00	4392.0
275.00	21544.0
276.10	2396.0
277.00	2890.0
278.00	498.0
293.00	388.0
296.00	8468.0
297.10	724.0
303.00	540.0
315.10	780.0
315.90	534.0
323.00	1530.0
323.90	392.0
326.90	480.0
334.00	919.0
345.90	418.0
351.80	330.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

352.90	480.0
354.00	411.0
364.90	2377.0
366.00	511.0
371.90	688.0
373.10	373.0
422.90	2143.0
423.90	590.0
441.10	7679.0
442.00	43040.0
443.00	9699.0
443.90	961.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

m/z	Abundance
37.00	582.0
38.00	1530.0
39.10	7668.0
40.00	434.0
44.00	1022.0
49.10	940.0
50.00	27528.0
51.00	91168.0
52.10	4957.0
54.90	530.0
56.00	2194.0
57.00	4349.0
58.10	366.0
61.00	1132.0
62.00	1588.0
63.00	3485.0
64.00	620.0
65.00	1169.0
68.00	1786.0
69.00	77096.0
70.00	470.0
73.00	1439.0
74.10	10742.0
75.00	13946.0
76.00	5091.0
77.00	87976.0
78.10	6558.0
79.00	8227.0
80.00	5166.0
81.00	7533.0
82.00	1775.0
82.90	1352.0
85.00	1260.0
86.00	2534.0
86.90	1059.0
91.00	1721.0
92.00	2119.0
93.00	14478.0
94.10	934.0
96.10	939.0
98.00	12653.0
99.00	9178.0
99.80	420.0
101.00	3336.0
103.00	1250.0
104.00	2781.0
105.00	2085.0
105.80	520.0
107.00	27160.0
108.00	4412.0
109.00	1003.0
110.00	38600.0
111.00	6966.0
112.00	963.0
115.00	353.0
116.00	1641.0
117.00	39480.0
118.00	2497.0
119.00	400.0
122.00	1697.0
122.90	3038.0
124.00	1506.0
125.00	1464.0
127.00	72248.0
128.00	5996.0
129.00	35272.0
129.90	2827.0
131.00	475.0
132.80	439.0
134.00	1009.0
135.00	2537.0
136.00	1327.0
137.10	2029.0
138.00	376.0
140.00	604.0
141.00	5684.0
142.00	1302.0
143.00	1148.0
145.00	354.0
145.90	936.0
146.90	2734.0
148.00	9143.0
148.90	1951.0
151.10	730.0
151.90	382.0
152.90	1623.0
154.00	1342.0
155.00	3187.0
156.10	3395.0
156.90	679.0
158.00	813.0
158.90	934.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

160.00	1694.0
161.00	2171.0
162.00	688.0
165.00	2333.0
166.00	1521.0
167.00	13155.0
168.00	7553.0
169.00	1110.0
169.90	477.0
171.00	457.0
172.00	1098.0
172.90	1005.0
174.00	2267.0
175.10	4028.0
176.00	1154.0
176.90	1859.0
177.90	808.0
179.00	7332.0
180.00	4853.0
181.00	2437.0
181.90	439.0
183.80	1003.0
184.90	3961.0
186.00	25080.0
187.00	6568.0
188.00	531.0
188.90	1841.0
190.00	591.0
190.90	1119.0
192.00	2534.0
193.00	2634.0
193.90	569.0
195.00	444.0
196.00	3350.0
197.20	1335.0
198.00	148544.0
198.90	10068.0
200.00	1091.0
201.70	514.0
203.00	1662.0
204.00	6840.0
205.10	11239.0
206.00	42744.0
207.10	6428.0
208.00	1714.0
209.10	967.0
210.10	848.0
211.00	2284.0
215.10	645.0
215.90	1238.0
217.00	14101.0
218.00	2090.0
221.00	5251.0
222.10	1119.0
223.00	3168.0
224.00	24824.0
225.00	5646.0
226.10	833.0
227.00	14799.0
228.00	1992.0
229.10	2585.0
230.10	503.0
231.10	1127.0
234.00	683.0
235.00	868.0
236.10	568.0
237.00	1377.0
238.90	412.0
240.00	441.0
241.00	751.0
242.00	1330.0
243.10	1318.0
244.00	18840.0
245.00	2721.0
246.00	4221.0
247.00	1357.0
248.90	634.0
253.00	641.0
253.90	718.0
255.00	94320.0
256.00	13674.0
257.00	1484.0
258.00	8038.0
258.90	1299.0
265.00	3340.0
270.00	385.0
270.90	333.0
273.00	3060.0
274.00	7503.0
275.00	40304.0
276.00	5395.0
277.00	4943.0
278.00	721.0
282.90	661.0
283.90	352.0
285.10	473.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

292.90	887.0
294.10	481.0
296.00	17776.0
297.00	2444.0
303.10	1446.0
304.00	436.0
314.00	492.0
315.00	2420.0
316.00	804.0
321.00	539.0
322.00	323.0
323.00	3970.0
324.00	563.0
327.00	999.0
328.00	334.0
332.00	366.0
333.00	345.0
334.00	2405.0
334.90	625.0
346.00	1128.0
351.90	763.0
353.00	606.0
353.90	1024.0
364.90	6734.0
365.90	1270.0
372.00	1880.0
373.00	1043.0
383.10	340.0
402.00	549.0
403.00	801.0
403.90	324.0
421.00	1126.0
422.00	916.0
423.00	5366.0
424.00	1448.0
441.10	16464.0
442.00	85784.0
443.00	20320.0
444.10	1855.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

m/z	Abundance
36.90	534.0
38.00	1233.0
39.10	5451.0
40.00	528.0
41.10	359.0
44.00	911.0
49.00	764.0
50.00	20720.0
51.10	67600.0
52.10	3806.0
56.00	1815.0
57.00	4276.0
61.00	463.0
62.00	1144.0
63.10	3292.0
63.90	441.0
65.00	1169.0
68.10	1388.0
69.00	61456.0
69.90	455.0
73.00	1110.0
74.00	8872.0
75.00	10863.0
76.00	3968.0
77.10	71448.0
78.00	4711.0
79.00	6341.0
80.00	4801.0
81.00	4925.0
82.00	1259.0
83.00	1002.0
85.10	706.0
86.00	1785.0
87.00	989.0
91.00	1176.0
92.00	1521.0
93.00	12657.0
94.10	530.0
96.10	390.0
98.00	9113.0
99.00	6259.0
99.90	554.0
101.00	2307.0
103.00	838.0
103.90	2106.0
105.00	1753.0
106.00	686.0
107.00	20752.0
108.00	3172.0
109.00	647.0
110.00	31832.0
111.00	5029.0
112.10	701.0
116.10	1147.0
117.00	32928.0
118.00	1885.0
119.00	457.0
119.90	400.0
122.00	1711.0
123.00	2593.0
123.90	889.0
125.00	754.0
127.00	61744.0
128.00	4789.0
129.00	31944.0
130.00	2272.0
130.90	680.0
131.90	418.0
133.90	605.0
135.00	2653.0
136.00	1216.0
137.00	1527.0
140.00	328.0
141.00	4556.0
142.00	1898.0
143.00	1090.0
145.90	1175.0
147.00	2678.0
148.00	6619.0
149.00	1759.0
149.90	360.0
151.10	1006.0
151.60	421.0
153.00	1650.0
153.90	965.0
155.00	2255.0
156.00	3520.0
156.90	895.0
158.00	745.0
158.90	628.0
160.00	1119.0
161.00	2142.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

162.00	613.0
165.00	1660.0
165.90	1427.0
167.00	11937.0
168.00	7069.0
169.00	600.0
170.00	432.0
171.10	635.0
172.00	739.0
173.00	1298.0
174.00	1745.0
175.00	3757.0
176.00	846.0
177.00	1779.0
177.90	763.0
179.00	6944.0
180.00	3598.0
181.00	2138.0
183.80	376.0
185.00	3683.0
186.00	21904.0
187.00	6356.0
188.00	487.0
189.00	1763.0
190.90	1047.0
192.00	2549.0
193.00	2803.0
193.90	433.0
195.10	329.0
196.00	3931.0
198.00	125112.0
199.00	8332.0
200.00	799.0
201.60	680.0
202.90	1194.0
204.00	6733.0
205.00	10518.0
206.10	41632.0
207.00	5857.0
207.90	2233.0
209.00	646.0
209.80	667.0
210.90	1848.0
214.90	979.0
216.00	1201.0
217.00	11218.0
217.90	1478.0
221.00	3227.0
221.90	1750.0
222.90	2539.0
224.10	20744.0
225.00	5410.0
226.10	836.0
227.00	11974.0
228.00	1368.0
229.00	2078.0
231.00	541.0
233.90	1129.0
234.90	774.0
236.10	442.0
237.00	699.0
238.90	388.0
240.10	324.0
241.00	525.0
242.00	1250.0
243.10	1492.0
244.00	16720.0
245.00	2442.0
245.90	4031.0
246.90	971.0
249.00	581.0
253.00	716.0
254.00	665.0
255.00	90400.0
256.00	12947.0
257.10	1019.0
258.00	6882.0
259.00	932.0
265.00	2497.0
266.00	406.0
267.00	370.0
270.80	313.0
273.10	2236.0
274.00	8325.0
275.00	39984.0
276.00	4811.0
277.00	4598.0
278.00	668.0
283.00	723.0
285.10	871.0
293.00	864.0
294.90	352.0
296.00	16036.0
296.90	2456.0
303.00	1454.0
304.00	313.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31

314.10	716.0
314.90	1337.0
316.00	632.0
321.00	523.0
323.10	2796.0
323.90	701.0
327.00	873.0
328.10	375.0
333.10	559.0
334.00	2972.0
334.90	710.0
341.10	525.0
345.90	1048.0
352.00	1119.0
352.90	853.0
354.00	1283.0
355.00	404.0
365.00	7292.0
366.10	818.0
370.90	315.0
372.00	1712.0
372.90	432.0
383.00	378.0
401.90	628.0
402.90	904.0
421.00	781.0
421.90	810.0
423.00	6761.0
424.10	1727.0
425.10	316.0
441.00	20808.0
442.00	133952.0
443.00	22624.0
444.00	2261.0

Instrument :
BNA_N
ClientSampleId :
DFTPP31