

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN122018\
 Data File : BN004154.D
 Acq On : 20 Dec 2018 17:27
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN122018

Quant Time: Dec 20 18:01:14 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN122018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 20 17:31:49 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.352	0.320	9.1	106	0.00
3	n-Nitrosodimethylamine	0.187	0.175	6.4	113	0.00
4 S	2-Fluorophenol	1.251	1.246	0.4	100	0.00
5 S	Phenol-d6	1.655	1.599	3.4	99	0.00
6	bis(2-Chloroethyl)ether	1.194	1.152	3.5	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	-0.01
8 S	Nitrobenzene-d5	0.257	0.235	8.6	99	0.00
9	Naphthalene	0.993	0.938	5.5	99	0.00
10	Hexachlorobutadiene	0.169	0.158	6.5	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.673	0.638	5.2	100	0.00
12	2-Methylnaphthalene	0.719	0.684	4.9	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.240	0.220	8.3	101	0.00
15 S	2-Fluorobiphenyl	1.596	1.444	9.5	99	0.00
16	Acenaphthylene	1.881	1.724	8.3	98	0.00
17	Acenaphthene	1.233	1.158	6.1	98	0.00
18	Fluorene	1.557	1.456	6.5	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4-Bromophenyl-phenylether	0.222	0.206	7.2	99	0.00
21	Hexachlorobenzene	0.252	0.238	5.6	100	0.00
22	Pentachlorophenol	0.046	0.047	-2.2	116	0.00
23	Phenanthrene	1.081	1.001	7.4	99	0.00
24	Anthracene	0.996	0.912	8.4	99	0.00
25 SURR	Fluoranthene-d10	1.179	1.089	7.6	99	0.00
26	Fluoranthene	1.284	1.189	7.4	100	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
28	Pvrene	1.253	1.179	5.9	100	0.00
29 S	Terphenyl-d14	0.990	0.909	8.2	99	0.00
30	Benzo(a)anthracene	1.111	1.027	7.6	98	0.00
31	Chrysene	1.136	1.074	5.5	103	0.00
32	Bis(2-ethylhexyl)phthalate	0.680	0.563	17.2	96	0.01
33	Indeno(1,2,3-cd)pyrene	1.211	1.154	4.7	101	0.00
34 I	Perylene-d12	1.000	1.000	0.0	101	0.00
35	Benzo(b)fluoranthene	1.151	1.063	7.6	102	0.00
36	Benzo(k)fluoranthene	1.211	1.166	3.7	104	0.00
37 C	Benzo(a)pyrene	1.089	1.022	6.2	100	0.00
38	Dibenzo(a,h)anthracene	1.090	1.040	4.6	101	0.00
39	Benzo(a,h,i)perylene	1.080	1.056	2.2	105	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN122018\
Data File : BN004154.D
Acq On : 20 Dec 2018 17:27
Operator : JU/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN122018

Quant Time: Dec 20 18:01:14 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN122018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 20 17:31:49 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			