

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122121\
 Data File : BN018137.D
 Acq On : 21 Dec 2021 21:47
 Operator : CG/JU
 Sample : SST0.853
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST0.8204

Quant Time: Dec 22 03:06:24 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 22 03:03:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.615	152	3893	0.400	ng/ul	0.00
4) Naphthalene-d8	10.392	136	11781	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.253	164	7065	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.994	188	15516	0.400	ng/ul	0.00
17) Chrysene-d12	21.187	240	17273	0.400	ng/ul	0.00
23) Perylene-d12	23.417	264	16690	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.080	96	3241	0.802	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.993	152	12938	0.942	ng/ul	0.00
18) Fluoranthene-d10	19.030	212	35038	0.712	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.113	88	3811	0.803	ng/ul	97
5) Naphthalene	10.442	128	26078	0.808	ng/ul	99
7) 2-Methylnaphthalene	12.064	142	17161	0.946	ng/ul	100
8) 1-Methylnaphthalene	12.284	142	17596	0.835	ng/ul	100
10) Acenaphthylene	13.970	152	21943	0.767	ng/ul	99
11) Acenaphthene	14.313	153	19330	0.760	ng/ul	100
12) Fluorene	15.303	166	22415	0.799	ng/ul	99
14) Pentachlorophenol	16.660	266	2018	0.743	ng/ul	98
15) Phenanthrene	17.036	178	39869	0.741	ng/ul	100
16) Anthracene	17.129	178	32695	0.772	ng/ul	99
19) Fluoranthene	19.058	202	50679	0.728	ng/ul	100
20) Pyrene	19.421	202	51488	0.661	ng/ul	99
21) Benzo(a)anthracene	21.172	228	44028	0.857	ng/ul	99
22) Chrysene	21.225	228	55029	0.746	ng/ul	100
24) Benzo(b)fluoranthene	22.745	252	51080	0.766	ng/ul	99
25) Benzo(k)fluoranthene	22.788	252	63703	0.716	ng/ul	99
26) Benzo(a)pyrene	23.318	252	47022	0.730	ng/ul	99
27) Indeno(1,2,3-cd)pyrene	25.675	276	67883	0.936	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.689	278	51889	0.978	ng/ul	97
29) Benzo(g,h,i)perylene	26.363	276	55262	0.889	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122121\
Data File : BN018137.D
Acq On : 21 Dec 2021 21:47
Operator : CG/JU
Sample : SST0.853
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.8204

Quant Time: Dec 22 03:06:24 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 22 03:03:54 2021
Response via : Initial Calibration

