

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122121\
 Data File : BN018139.D
 Acq On : 21 Dec 2021 22:59
 Operator : CG/JU
 Sample : SSTD3.255
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.2206

Quant Time: Dec 22 03:06:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 22 03:03:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.611	152	4486	0.400	ng/ul	0.00
4) Naphthalene-d8	10.387	136	14209	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.253	164	9038	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.994	188	19931	0.400	ng/ul	0.00
17) Chrysene-d12	21.187	240	21694	0.400	ng/ul	0.00
23) Perylene-d12	23.417	264	19101	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.076	96	13997	3.006	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.109	88	16139	2.951	ng/ul#	86
14) Pentachlorophenol	16.656	266	15545	4.454	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122121\
Data File : BN018139.D
Acq On : 21 Dec 2021 22:59
Operator : CG/JU
Sample : SST3.255
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.2206

Quant Time: Dec 22 03:06:40 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 22 03:03:54 2021
Response via : Initial Calibration

