

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122221\
 Data File : BN018156.D
 Acq On : 22 Dec 2021 18:23
 Operator : CG/JU
 Sample : M5091-01
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BG3C0

Quant Time: Dec 23 00:08:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 22 14:08:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.611	152	3946	0.400	ng/ul	0.00
4) Naphthalene-d8	10.388	136	13061	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.249	164	8735	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.991	188	20459	0.400	ng/ul	0.00
17) Chrysene-d12	21.192	240	18946	0.400	ng/ul	0.00
23) Perylene-d12	23.416	264	20688	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.076	96	19228	4.715	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.994	152	5661	0.315	ng/ul	0.00
18) Fluoranthene-d10	19.027	212	18961	0.369	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122221\
Data File : BN018156.D
Acq On : 22 Dec 2021 18:23
Operator : CG/JU
Sample : M5091-01
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BG3C0

Quant Time: Dec 23 00:08:14 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 22 14:08:04 2021
Response via : Initial Calibration

