

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122221\
 Data File : BN018162.D
 Acq On : 22 Dec 2021 21:59
 Operator : CG/JU
 Sample : M5091-08
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BG3C9

Quant Time: Dec 23 00:09:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 22 14:08:04 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.611	152	3268	0.400	ng/ul	0.00
4) Naphthalene-d8	10.387	136	11192	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.248	164	7782	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.994	188	18366	0.400	ng/ul	0.00
17) Chrysene-d12	21.190	240	17625	0.400	ng/ul	0.00
23) Perylene-d12	23.414	264	20963	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.075	96	17280	5.116	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.992	152	5745	0.373	ng/ul	0.00
18) Fluoranthene-d10	19.026	212	22413	0.469	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.113	88	264	0.065	ng/ul#	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Dec 23 00:09:54 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 22 14:08:04 2021
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.065 ng/ul
 RT: 3.113 min Scan# 31
 Delta R.T. -0.000 min
 Lab File: BN018162.D
 Acq: 22 Dec 2021 21:59

Instrument :
 BNA_N
 ClientSampleId :
 BG3C9



Tgt Ion: 88 Resp: 264

Ion	Ratio	Lower	Upper
88	100		
43	37.3	24.6	36.8#
58	23.8	45.2	67.8#

