

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122321\
 Data File : BN018183.D
 Acq On : 23 Dec 2021 16:01
 Operator : CG/JU
 Sample : M5179-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 POST-EXCAVATION-BOTTOM-WEST

Quant Time: Dec 24 23:31:51 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN122321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 23 14:22:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.624	152	21728	0.400	ng	0.00
7) Naphthalene-d8	10.394	136	95867	0.400	ng	# 0.00
13) Acenaphthene-d10	14.256	164	60351	0.400	ng	0.00
19) Phenanthrene-d10	16.993	188	114404	0.400	ng	0.00
29) Chrysene-d12	21.196	240	105858	0.400	ng	0.00
35) Perylene-d12	23.420	264	102018	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.254	112	13335	0.281	ng	0.03
5) Phenol-d6	6.822	99	15289	0.301	ng	0.00
8) Nitrobenzene-d5	8.772	82	18069	0.280	ng	0.00
11) 2-Methylnaphthalene-d10	11.994	152	42155	0.256	ng	0.00
14) 2,4,6-Tribromophenol	15.753	330	5581	0.322	ng	0.00
15) 2-Fluorobiphenyl	12.886	172	64439	0.286	ng	0.00
27) Fluoranthene-d10	19.033	212	92326	0.255	ng	0.00
31) Terphenyl-d14	19.639	244	71925	0.309	ng	0.00
Target Compounds						
						Qvalue
3) n-Nitrosodimethylamine	3.438	42	943	0.050	ng	# 1
9) Naphthalene	10.445	128	28119	0.119	ng	98
12) 2-Methylnaphthalene	12.070	142	12401	0.079	ng	99
16) Acenaphthylene	13.964	152	14902	0.066	ng	# 86
17) Acenaphthene	14.320	154	10087	0.065	ng	99
18) Fluorene	15.309	166	13889	0.072	ng	# 95
20) 4,6-Dinitro-2-methylph...	15.360	198	115	0.172	ng	# 1
23) Atrazine	16.397	200	458	0.110	ng	# 1
24) Pentachlorophenol	16.665	266	383	0.303	ng	# 66
25) Phenanthrene	17.127	178	17853	0.059	ng	98
26) Anthracene	17.127	178	17844	0.071	ng	97
28) Fluoranthene	19.063	202	37997	0.107	ng	100
30) Pyrene	19.425	202	37258	0.110	ng	# 95
32) Benzo(a)anthracene	21.174	228	28194	0.092	ng	# 91
33) Chrysene	21.228	228	30097	0.090	ng	# 94
34) Bis(2-ethylhexyl)phtha...	21.121	149	589020	3.692	ng	99
36) Indeno(1,2,3-cd)pyrene	25.685	276	20182	0.068	ng	98
37) Benzo(b)fluoranthene	22.752	252	29572	0.094	ng	# 65
38) Benzo(k)fluoranthene	22.793	252	21979	0.069	ng	# 59
39) Benzo(a)pyrene	23.324	252	21504	0.076	ng	# 63
40) Dibenzo(a,h)anthracene	25.703	278	13106	0.082	ng	# 74
41) Benzo(g,h,i)perylene	26.377	276	21456	0.079	ng	# 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122321\
Data File : BN018183.D
Acq On : 23 Dec 2021 16:01
Operator : CG/JU
Sample : M5179-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
POST-EXCAVATION-BOTTOM-WEST

Quant Time: Dec 24 23:31:51 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN122321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 23 14:22:48 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.624 min Scan# 49
Delta R.T. 0.000 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01




Instrument :
BNA_N
ClientSampleId :
POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:152 Resp: 21728
Ion Ratio Lower Upper
152 100
150 155.0 121.7 182.5
115 56.1 43.8 65.8



#3
n-Nitrosodimethylamine
Concen: 0.050 ng
RT: 3.438 min Scan# 39
Delta R.T. -0.055 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01

Tgt Ion: 42 Resp: 943
Ion Ratio Lower Upper
42 100
74 1.1 108.6 163.0#
44 118.0 5.8 8.8#









#4
2-Fluorophenol
Concen: 0.281 ng
RT: 5.254 min Scan# 21
Delta R.T. 0.028 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01




Instrument :
BNA_N
ClientSampleId :
POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:112 Resp: 13335
Ion Ratio Lower Upper
112 100
64 60.2 48.0 72.0
63 33.5 25.3 37.9



#5
Phenol-d6
Concen: 0.301 ng
RT: 6.822 min Scan# 406
Delta R.T. 0.010 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01

Tgt Ion: 99 Resp: 15289
Ion Ratio Lower Upper
99 100
42 67.1 16.3 24.5#
71 208.4 26.8 40.2#









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.394 min Scan# 71
Delta R.T. 0.000 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01




Instrument :
BNA_N
ClientSampleId :
POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:	136	Resp:	95867
Ion Ratio	Lower	Upper	
136	100		
137	11.3	8.8	13.2
54	10.2	5.0	7.6#
68	8.9	4.4	6.6#



#8
Nitrobenzene-d5
Concen: 0.280 ng
RT: 8.772 min Scan# 604
Delta R.T. -0.000 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01

Tgt Ion:	82	Resp:	18069
Ion Ratio	Lower	Upper	
82	100		
128	36.3	33.7	50.5
54	45.1	36.6	55.0









#9
Naphthalene
Concen: 0.119 ng
RT: 10.445 min Scan# 736
Delta R.T. 0.001 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01




Instrument :
BNA_N
ClientSampleId :
POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:128 Resp: 28119
Ion Ratio Lower Upper
128 100
129 11.5 8.7 13.1
127 13.8 10.4 15.6



#11
2-Methylnaphthalene-d10
Concen: 0.256 ng
RT: 11.994 min Scan# 929
Delta R.T. -0.000 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01

Tgt Ion:152 Resp: 42155
Ion Ratio Lower Upper
152 100
151 21.2 16.2 24.2





#12

2-Methylnaphthalene

Concen: 0.079 ng

RT: 12.070 min Scan# 946

Delta R.T. 0.001 min

Lab File: BN018183.D

Acq: 23 Dec 2021 16:01

Instrument :

BNA_N

ClientSampleId :

POST-EXCAVATION-BOTTOM-WEST





Tgt Ion:142 Resp: 12401

Ion Ratio Lower Upper

142 100

141 88.6 70.0 105.0

115 35.2 27.8 41.8



#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.256 min Scan# 1182

Delta R.T. 0.000 min

Lab File: BN018183.D

Acq: 23 Dec 2021 16:01




Tgt Ion:164 Resp: 60351


Ion Ratio Lower Upper

164 100

162 98.3 80.2 120.2

160 43.1 36.9 55.3









#14
2,4,6-Tribromophenol
Concen: 0.322 ng
RT: 15.753 min Scan# 1300
Delta R.T. 0.000 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01




Instrument :
BNA_N
ClientSampleId :
POST-EXCAVATION-BOTTOM-WEST


Tgt Ion: 330 Resp: 5581
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 37.3 21.3 31.9#



#15
2-Fluorobiphenyl
Concen: 0.286 ng
RT: 12.886 min Scan# 1074
Delta R.T. -0.000 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01

Tgt Ion: 172 Resp: 64439
Ion Ratio Lower Upper
172 100
171 33.2 27.4 41.0
170 21.1 18.4 27.6









#16
Acenaphthylene
Concen: 0.066 ng
RT: 13.964 min Scan# 1159
Delta R.T. -0.013 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01




Instrument :
BNA_N
ClientSampleId :
POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:152 Resp: 14902
Ion Ratio Lower Upper
152 100
151 27.9 15.8 23.6#
153 16.6 10.5 15.7#



#17
Acenaphthene
Concen: 0.065 ng
RT: 14.320 min Scan# 1187
Delta R.T. 0.001 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01

Tgt Ion:154 Resp: 10087
Ion Ratio Lower Upper
154 100
153 113.3 90.9 136.3
152 57.9 45.0 67.4









#18
Fluorene
Concen: 0.072 ng
RT: 15.309 min Scan# 1265
Delta R.T. 0.000 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01




Instrument :
BNA_N
ClientSampleId :
POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:166 Resp: 13889
Ion Ratio Lower Upper
166 100
165 101.2 77.8 116.6
167 19.0 10.9 16.3#



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.993 min Scan# 1401
Delta R.T. 0.000 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01

Tgt Ion:188 Resp: 114404
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.0 7.0 10.4









#20
4,6-Dinitro-2-methylphenol
Concen: 0.172 ng
RT: 15.360 min Scan# 11
Delta R.T. -0.038 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01




Instrument :
BNA_N
ClientSampleId :
POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:198 Resp: 115
Ion Ratio Lower Upper
198 100
182 405.3 15.5 23.3#
77 0.0 0.0 0.0



#23
Atrazine
Concen: 0.110 ng
RT: 16.397 min Scan# 1352
Delta R.T. -0.073 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01

Tgt Ion:200 Resp: 458
Ion Ratio Lower Upper
200 100
173 401.3 20.3 30.5#
215 225.8 40.4 60.6#









#24
 Pentachlorophenol
 Concen: 0.303 ng
 RT: 16.665 min Scan# 111
 Delta R.T. 0.001 min
 Lab File: BN018183.D
 Acq: 23 Dec 2021 16:01




Instrument :
 BNA_N
 ClientSampleId :
 POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:266 Resp: 383
 Ion Ratio Lower Upper
 266 100
 264 99.2 49.5 74.3#
 268 47.5 50.6 76.0#



#25
 Phenanthrene
 Concen: 0.059 ng
 RT: 17.127 min Scan# 1412
 Delta R.T. 0.085 min
 Lab File: BN018183.D
 Acq: 23 Dec 2021 16:01

Tgt Ion:178 Resp: 17853
 Ion Ratio Lower Upper
 178 100
 176 19.2 15.3 22.9
 179 16.8 12.3 18.5









#26
 Anthracene
 Concen: 0.071 ng
 RT: 17.127 min Scan# 1412
 Delta R.T. 0.000 min
 Lab File: BN018183.D
 Acq: 23 Dec 2021 16:01




Instrument :
 BNA_N
 ClientSampleId :
 POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:178 Resp: 17844
 Ion Ratio Lower Upper
 178 100
 176 19.2 14.8 22.2
 179 16.8 12.2 18.4



#27
 Fluoranthene-d10
 Concen: 0.255 ng
 RT: 19.033 min Scan# 1690
 Delta R.T. 0.000 min
 Lab File: BN018183.D
 Acq: 23 Dec 2021 16:01

Tgt Ion:212 Resp: 92326
 Ion Ratio Lower Upper
 212 100
 106 10.4 8.4 12.6
 104 5.9 4.5 6.7









#28
Fluoranthene
Concen: 0.107 ng
RT: 19.063 min Scan# 1
Delta R.T. 0.000 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01




Instrument :
BNA_N
ClientSampleId :
POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:202 Resp: 37997
Ion Ratio Lower Upper
202 100
101 9.0 7.2 10.8
203 17.4 13.7 20.5



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.196 min Scan# 1998
Delta R.T. 0.001 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01

Tgt Ion:240 Resp: 105858
Ion Ratio Lower Upper
240 100
120 6.8 6.0 9.0
236 26.5 21.4 32.2









#30
Pyrene
Concen: 0.110 ng
RT: 19.425 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01




Instrument :
BNA_N
ClientSampleId :
POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:202 Resp: 37258
Ion Ratio Lower Upper
202 100
200 21.5 16.6 25.0
203 21.9 14.1 21.1#



#31
Terphenyl-d14
Concen: 0.309 ng
RT: 19.639 min Scan# 1812
Delta R.T. -0.005 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01

Tgt Ion:244 Resp: 71925
Ion Ratio Lower Upper
244 100
212 7.9 6.2 9.2
122 10.9 7.9 11.9









#32
Benzo(a)anthracene
Concen: 0.092 ng
RT: 21.174 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01




Instrument : BNA_N
ClientSampleId : POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:228 Resp: 28194
Ion Ratio Lower Upper
228 100
226 29.6 21.3 31.9
229 26.1 15.9 23.9#



#33
Chrysene
Concen: 0.090 ng
RT: 21.228 min Scan# 2001
Delta R.T. 0.001 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01

Tgt Ion:228 Resp: 30097
Ion Ratio Lower Upper
228 100
226 31.5 24.2 36.4
229 25.4 15.5 23.3#









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 3.692 ng
 RT: 21.121 min Scan# 1991
 Delta R.T. 0.000 min
 Lab File: BN018183.D
 Acq: 23 Dec 2021 16:01




Instrument : BNA_N
 ClientSampleId : POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:149 Resp: 589020
 Ion Ratio Lower Upper
 149 100
 167 30.4 23.7 35.5
 279 5.0 3.6 5.4



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.420 min Scan# 2408
 Delta R.T. 0.001 min
 Lab File: BN018183.D
 Acq: 23 Dec 2021 16:01

Tgt Ion:264 Resp: 102018
 Ion Ratio Lower Upper
 264 100
 260 25.5 20.2 30.4
 265 22.8 18.2 27.2









#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.068 ng
 RT: 25.685 min Scan# 3070
 Delta R.T. 0.000 min
 Lab File: BN018183.D
 Acq: 23 Dec 2021 16:01




Instrument :
 BNA_N
 ClientSampleId :
 POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:	276	Resp:	20182
Ion Ratio	Lower	Upper	
276	100		
138	18.1	14.9	22.3
277	23.5	19.9	29.9



#37
 Benzo(b)fluoranthene
 Concen: 0.094 ng
 RT: 22.752 min Scan# 2213
 Delta R.T. 0.000 min
 Lab File: BN018183.D
 Acq: 23 Dec 2021 16:01

Tgt Ion:	252	Resp:	29572
Ion Ratio	Lower	Upper	
252	100		
253	30.7	17.8	26.6#
125	37.4	6.0	9.0#









#38
Benzo(k)fluoranthene
Concen: 0.069 ng
RT: 22.793 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01




Instrument :
BNA_N
ClientSampleId :
POST-EXCAVATION-BOTTOM-WEST


Tgt Ion:252 Resp: 21979
Ion Ratio Lower Upper
252 100
253 30.2 17.7 26.5#
125 46.6 6.2 9.2#



#39
Benzo(a)pyrene
Concen: 0.076 ng
RT: 23.324 min Scan# 2380
Delta R.T. 0.001 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01

Tgt Ion:252 Resp: 21504
Ion Ratio Lower Upper
252 100
253 29.2 17.8 26.6#
125 43.2 7.0 10.4#









#40
Dibenzo(a,h)anthracene
Concen: 0.082 ng
RT: 25.703 min Scan# 3075
Delta R.T. 0.004 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01

Instrument :
BNA_N
ClientSampleId :
POST-EXCAVATION-BOTTOM-WEST

Tgt Ion:278 Resp: 13106
Ion Ratio Lower Upper
278 100
139 34.5 10.8 16.2#
279 28.8 19.0 28.4#



#41
Benzo(g,h,i)perylene
Concen: 0.079 ng
RT: 26.377 min Scan# 3272
Delta R.T. 0.007 min
Lab File: BN018183.D
Acq: 23 Dec 2021 16:01

Tgt Ion:276 Resp: 21456
Ion Ratio Lower Upper
276 100
277 26.7 19.0 28.6
138 22.5 12.6 19.0#

