


2,4,6-Tribromophenol

Response Ratio




$$\text{Response} = 2.613\text{e-}001 * \text{Amt} - 1.177\text{e-}001$$

Coef of Det (r^2) = 0.999517 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN122321.M

Calibration Table Last Updated: Thu Dec 23 14:22:48 2021


4,6-Dinitro-2-methylphenol



$\text{Response} = 2.808\text{e-}002 * \text{Amt} - 1.106\text{e-}002$
 Coef of Det (r^2) = 0.990226 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN122321.M
 Calibration Table Last Updated: Thu Dec 23 14:22:48 2021

Atrazine

Response Ratio



$$\text{Response} = 2.004\text{e-}001 * \text{Amt} - 5.095\text{e-}002$$


Coef of Det (r^2) = 0.999041 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN122321.M

Calibration Table Last Updated: Thu Dec 23 14:22:48 2021

Pentachlorophenol

Response Ratio



$$\text{Response} = 1.306\text{e-}001 * \text{Amt} - 9.560\text{e-}002$$


Coef of Det (r^2) = 0.996821 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN122321.M

Calibration Table Last Updated: Thu Dec 23 14:22:48 2021

Bis(2-ethylhexyl)phthalate

Response Ratio



$$\text{Response} = 6.494\text{e-}001 * \text{Amt} - 4.299\text{e-}001$$


Coef of Det (r^2) = 0.996861 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN122321.M

Calibration Table Last Updated: Thu Dec 23 14:22:48 2021

Dibenzo (a,h) anthracene

Response Ratio



$$\text{Response} = 9.894\text{e-}001 * \text{Amt} - 7.385\text{e-}002$$

Coef of Det (r^2) = 0.997676 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN122321.M

Calibration Table Last Updated: Thu Dec 23 14:22:48 2021