

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN122419\
 Data File : BN009153.D
 Acq On : 25 Dec 2019 08:18
 Operator : JU
 Sample : PB125493BL
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK93

Quant Time: Dec 25 12:28:32 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN122419MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 25 01:06:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	254789	20.00	ng/ul	0.00
18) Naphthalene-d8	10.28	136	1197871	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.16	164	826239	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.90	188	1799649	20.00	ng/ul	0.00
77) Chrysene-d12	21.11	240	1728701	20.00	ng/ul	0.00
85) Perylene-d12	23.25	264	1709377	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.13	96	48456	8.69	ng/uL	0.00
5) Phenol-d5	6.72	99	1017503	42.01	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.88	67	701426	45.20	ng/ul	0.00
9) 2-Chlorophenol-d4	7.06	132	768554	44.30	ng/ul	0.00
13) 4-Methylphenol-d8	8.23	113	859658	43.96	ng/ul	0.00
19) Nitrobenzene-d5	8.66	128	425785	47.20	ng/ul	0.00
22) 2-Nitrophenol-d4	9.38	143	481426	47.81	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.91	165	816270	42.93	ng/ul	0.00
29) 4-Chloroaniline-d4	10.42	131	1222971	53.38	ng/ul	0.00
43) Dimethylphthalate-d6	13.58	166	2841484	45.79	ng/ul	0.00
46) Acenaphthylene-d8	13.85	160	3382929	46.50	ng/ul	0.00
51) 4-Nitrophenol-d4	14.37	143	532655	43.17	ng/ul	0.00
57) Fluorene-d10	15.16	176	2442003	45.18	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.29	200	434531	38.70	ng/ul	0.00
70) Anthracene-d10	17.00	188	3850452	46.00	ng/ul	0.00
78) Pyrene-d10	19.30	212	4221279	46.80	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.12	264	4151984	48.21	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN122419\
Data File : BN009153.D
Acq On : 25 Dec 2019 08:18
Operator : JU
Sample : PB125493BL
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK93

Quant Time: Dec 25 12:28:32 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN122419MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 25 01:06:45 2019
Response via : Initial Calibration

