

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122422\
 Data File : BN023463.D
 Acq On : 26 Dec 2022 00:31
 Operator : CG/JU
 Sample : N6032-07
 Misc :
 ALS Vial : 99 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0L98

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/26/2022
 Supervised By :Sohil Jodhani 12/26/2022

Quant Time: Dec 26 02:58:30 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Dec 25 05:17:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.985	152	5958	0.400	ng/ul	0.00
4) Naphthalene-d8	10.803	136	19859	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.635	164	11587	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.385	188	26202	0.400	ng/ul	0.00
17) Chrysene-d12	21.569	240	16724	0.400	ng/ul	0.00
23) Perylene-d12	24.013	264	13416	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.307	96	734412m	113.447	ng/ul	-0.03
6) 2-Methylnaphthalene-d10	12.392	152	8648	0.282	ng/ul	0.00
18) Fluoranthene-d10	19.409	212	20747	0.343	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.378	88	853	0.129	ng/ul#	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122422\
Data File : BN023463.D
Acq On : 26 Dec 2022 00:31
Operator : CG/JU
Sample : N6032-07
Misc :
ALS Vial : 99 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
COL98

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/26/2022
Supervised By :Sohil Jodhani 12/26/2022

Quant Time: Dec 26 02:58:30 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN122122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Dec 25 05:17:59 2022
Response via : Initial Calibration

